

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Institute of Calibration Technology Co.,Ltd

72, Dongtancheomdansaneop 2-ro, Hwaseong-si, Gyeonggi-do, Republic of Korea
Phone : 82-31-379-5114, Fax : 82-31-379-5115 e-mail : sictadmin@sict.co.kr

CALIBRATION

Valid To : Oct. 29. 2025

Accreditation No : KC01-018

In recognition of the successful completion of the KOLAS evaluation process,
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
102. Linear dimension			10235	Ultrasonic/coating thickness specimens	N	104. Form		
10201	Balls	N				10401	Form testers	Y
10203	Electrical/Mechanical comparators	Y	10236	Coating thickness testers	Y	10404	Optical flats	N
			10237	Torque arms	N	10405	Optical parallels	N
10204	Gauge block comparators	Y	10238	Width measuring specimens	N	10406	Parallel blocks	Y
10206	Dial/cylinder gauge testers	N	103. Angle			10407	Precision surface plates	Y
10207	Doctor blades	N	10302	Angle gauge blocks	N	10408	Profile gauges	N
10208	Distance meters: electrooptic /laser/ultrasonic	N	10303	Autocollimators	N	10409	Roundness measurement instruments	Y
			10304	Bevel protractors	Y			
10209	End bars	N	10306	Clinometers	N	10410	Form standard specimens	N
10210	Extensometers, linear displacement transducers	Y	10307	Collimators	Y	10411	Roundness standard/roundness magnification standard specimens	N
			10308	Fine angle generators, level comparators	N			
10211	Filler gauges	Y						
10212	Film applicators	N	10310	Indexing tables	N	10412	Straight edges	Y
10213	Gap gauges	N	10311	Plate/square/electric levels	N	10413	Straight rules	N
10214	Gauge blocks, by comparison	N	10312	Auto levels	N	10415	Test bars	N
10216	Height gauges/measuring machines	Y	10314	Penta-prisms	N	10416	Spherometers	N
			10315	Polygons	N	105. Complex geometry		
10219	Linear scales	N	10316	Rotary tables	Y	10501	Base gauges for electric bulb	N
10220	Standard measuring machines	Y	10317	Sine bars, plates, tables, centers	N	10502	Bench centers	Y
10221	Micro scales/Standard scales	N						
10223	Electronic micrometers	Y	10318	Squareness testers, right angle testers	Y	10503	Contact coordinate measuring machines	Y
10224	Height micrometers, riser blocks	N	10319	Cylindrical squares	N	10504	Non-contact coordinate measuring machines	Y
10225	Laser scan micrometers	Y	10320	Precision squares	N			
10227	Standard tape rules, peripheral gauges	N	10321	Theodolites, transits	N	10505	Gauge block accessories	N
			10322	Angular displacement transducers	Y	10508	Hardness indenters	N
10228	Cylindrical plug/pin gauges, Thread measuring wire gauges	Y	10323	Alignment telescopes, line of sight collimators	N	10511	Measuring microscopes, Profile projectors	Y
10229	Radius gauges	N				10512	Micro measuring microscopes	Y
10230	Cylindrical ring gauges	N	10324	Calibration system for survey instruments	Y	10513	Orifice plates	N
10231	Step blocks	N				10514	Taper plug gauges	N
10232	Step gauges	N	10325	Jig transits	N	10515	Taper ring gauges	N
10233	Taper thickness gauges	N	10326	Laser levels	N	10517	Stylus type roughness testers	Y
10234	Ultrasonic thickness gauges	Y	10327	Optical wedges	N			

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site		
10518	Socket gauges for electric bulb	N	202. Force			20706	Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	N		
10519	Roughness standard /comparison specimens		20202	Force measuring devices	N					
10525	Thread plug gauges	N	20204	push-pull gauge	Y	20707	Chloride meters	N		
10526	Taper thread plug gauges	N	203. Torque			20799	Others; Solid density	N		
10527	Thread ring gauges	N	20302	Torque measuring devices	Y	208. Viscosity				
10528	Taper thread ring gauges	N	20303	Torque wrenches/drivers	Y	20801	Kinematic viscometers; capillary, etc.	N		
10529	V-blocks, box blocks	N	204. Pressure							
10530	Position gauges	N	20401	Altimeters	Y	20802	Dynamic viscometers; rotaional, etc.	Y		
10531	SEM/TEM/SPM/AFM microscopes	Y	20402	Manometers	Y					
106. Various dimensional			20403	Pneumatic pressure ballances	N	209. Fluid flow				
10601	Inside/outside/gear tooth calipers, caliper gauges	Y	20404	Hydraulic pressure ballances	N	20901	Anemometers; hot-wire	N		
			20405	Air data test systems	Y	20902	Anemometers; pitot tube, etc.	N		
10603	Cylinder/bore gauges	Y	20406	Absolute pressure gauges	Y	20908	Gas flowmeters; differential pressure	N		
10604	Depth gauges, depth micrometers	Y	20407	Blood pressure gauges	Y					
			20408	Compound pressure gauges	Y	20909	Liquid flowmeters; differential pressure	N		
10605	Dial/digital gauges	Y	20409	Differential pressure gauges	Y					
10606	Geodesic baselines	Y	20411	Gauge pressure gauges	Y	20910	Liquid flowmeters; electromagnetic	N		
10608	Grind gauges	N	20412	Pressure transducers/transmitters	Y					
10609	Micro indicators, test indicators	Y			20911	Gas flowmeters; thermal mass, etc.	N			
10610	Micrometer heads	Y	20414	Water depth meters	Y	20912	Liquid flowmeters; Coriolis, etc.	N		
10611	3-points, micrometers	Y	205. Vacuum							
10612	Inside micrometers	Y	20501	Capacitance diaphragm gauges	N	20914	Gas flowmeters; positive displacement	N		
10613	Outside micrometers	Y	20502	Spinning rotor gauges	N					
10615	Particle counters	N	20503	Ionization gauges	N	20915	Liquid flowmeters; positive displacement	N		
10617	Standard sieves	N	20504	Thermal conductivity gauge; pirani, thermocouple, convectron, etc.	N					
10619	Water level meters	N			20916	Gas flowmeters; turbine	N			
10620	Welding gauges	N	20505	Standard leaks, Helium leak detectors	Y	20917	Liquid flowmeters; turbine	N		
10621	Optical micrometers	N				20918	Gas flowmeters; ultrasonic	N		
201. Mass						20919	Liquid flowmeters; ultrasonic	N		
20102	Auto-hopper scale balances	Y	206. Volume							
20103	Auto-packer scale balances	Y	20601	Volumetric glasswares	N	20920	Gas flowmeters; variable area	N		
20104	Axle weigher balances	N	20602	Pycnometers	N					
20106	Dial platform scale balances	Y	20603	Rain gauges	Y	20921	Liquid flowmeters; variable area	N		
20107	Dial swing scale balances	Y	20604	Standard volume vessels	Y					
20109	Electric balances	Y	20605	Concrete air content meters	N	20922	Gas flowmeters; vortex	N		
20111	Manual swing scale balances	Y	20606	Piston type volume meters	N					
20112	Platform scale balances	Y	207. Density			20923	Liquid flowmeters; vortex	N		
20113	Spring scale balances	Y	20702	Liquid density meters	N	20925	Anemometers; vane, etc	N		
20116	Weights	Y	20704	Salinity meters	N	20999	Others; Anemometers; ultrasonic waves	N		
			20705	Sucrose meters	N					

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
210. Hardness			402. Resistance, Capacitance and Inductance			404. Other DC & LF measurements		
21001	Brinell hardness testers	Y	40201	Capacitance bridges /indicators	Y	40401	LF amplifiers	Y
21002	Rockwell hardness testers	Y				40402	DC/LF attenuators	Y
21003	Shore hardness testers	Y				40403	Multimeter calibrators	N
21004	Vickers hardness testers	Y	40202	Decade capacitors	Y	40404	Oscilloscope calibrators	N
21005	Durometer hardness testers	N	40204	Standard capacitors	Y	40405	CD/DVD meters/analyzers	Y
21006	Leeb hardness testers	Y	40205	Earth testers	Y	40406	Video signal generators	Y
211. Impact			40206	Inductance bridges /indicators	Y	40407	Audio distortion analyzers /meters	Y
21102	Charpy impact testers	Y	40208	Inductors	Y	40408	LF filters	Y
301. Time/frequency			40210	Insulation testers	Y	40409	LF/audio signal analyzers	Y
30102	Frequency standards	N	40211	Q-meters	Y	40410	Line frequency meters	Y
30103	General frequency sources	Y	40213	Resistance bridges & similar instruments	Y	40411	Function generators	Y
30104	Frequency meters/counters	Y	40214		Y	40412	Genescopes	Y
30105	Time interval sources	Y	40215	Resistors	Y	40413	AC/DC high voltage voltmeters	Y
30106	Time interval meters /stop watches/timers	Y	40217	Impedance bridges/LCR meters	Y	40415	Jitter meters	Y
302. Velocity & revolution			403. AC voltage, current & power			40416	Leakage current testers	Y
30201	Standard RPM generators	Y	40301	AC ammeters	Y	40417	Electronic AC/DC loads	Y
30202	Contact type tachometers	Y	40302	Clamp ammeters/voltmeters	Y	40418	Modulation meters	Y
30203	Photo tachometers /stroboscopes	Y	40303	AC voltage/current calibrators	Y	40419	Analogue/digital multimeters	Y
30204	Speed meters	Y	40304		Y	40420	Noise meters	Y
30205	Wow-flutter generators	N	40305	AC current shunts	Y	40422	LF phase meters	Y
30206	Wow-flutter meters	Y	40306	Phase angle generators, synchro resolve generators	N	40424	Voltage/current recorders	Y
401. DC Voltage & current			40307		Y	40425	Relay test sets	Y
40101	DC ammeters	Y	40308	Voltage/current phase angle meters/synchro resolve meters	N	40426	LF signal generators	Y
40102	Transconductance amplifiers	Y			Y	40427	LF spectrum analyzers	Y
40103	DC voltage/current calibrators	Y			Y	40428	Spot generators	Y
40104	Electrical temperature calibrators	Y	40309	Potential transformer	N	40429	Sweep generators	Y
40105	DC current shunts	Y	40310	Power factor meters	Y	40430	Signal transducers	Y
40106	Galvanometers /null detectors	Y	40311	AC power meters	Y	40432	Transistor curve tracers	Y
40107	Potentiometers	Y	40312	AC power supplies	Y	40434	AC/DC high voltage generators	Y
40108	DC power supplies	Y	40313	Puncture/safety testers	Y	40435	AC/DC high voltage probes	Y
40110	DC voltage dividers	N	40314	Power recorders	Y	40436	Logic analyzers	Y
40111	DC voltage standards	N	40315	Current transformer test sets	Y	40437	Telephone testers	Y
40112	DC voltmeters	Y	40316		N	40438	Video signal analyzers	Y
40113	Static/ionic voltmeters	N	40318	AC voltmeters	Y	40504	Flux sources	N
			40319	Watt hour meters	Y	40508	Magnetometers	Y
			40321	Ratio transformers	N	40510	Reference/standard magnets	N
405. Low frequency electric & magnetic field								

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
406.	Radio frequency measurement		40652	Field strength meters	Y	50302	Relative humidity hygrometers; polymer thin film, hair, etc.	
40601	RF amplifiers	Y	40653	AM/FM test sources	Y			Y
40602	Coaxial attenuators	Y	40654	Dip simulators	Y			
40605	Burst pulse generators	Y	407. Field strength & antenna			50303	Psychrometers; Assmann ventilated, PRT type, etc.	N
40606	Attenuator calibrators	N	40702	Probes	N			
40607	RF power meter calibrators	Y	40703	Dipole antennas	N			
40608	EMC transducers; current probes, absorbing clamps, etc.	Y	40704	Loop antennas	N	50304	Temperature humidity recorders; hygrothermograph, etc	Y
			40705	Monopole antennas	N			
			40707	Horn antennas	N			
40610	Coaxial directional couplers /splitters	Y	501. Contact thermometry			50305	Transducers; dew-point /relative humidity	N
40613	Electrostatic discharge generators	Y	50101	Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	Y	50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	
40614	EMC receivers	Y	50102	Temperature indicators /recorders/controllers, temperature calibrators	Y			Y
40615	RF filters	Y	50103	Glass thermometers; liquid-in-glass, Beckmann	N	504. Moisture		
40616	RF impedance meters	N	50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.	Y	50401	Cereal moisture meters	Y
40617	RF impulse generators	Y	50105	Thermal expansion thermometers; bimetal, gas or liquid type	Y	50402	Wood moisture meters	N
40618	Line impedance stabilization networks; LISN, CDN, ISN, etc.	Y	50106	Thermocouples: noble metal, base metal, pure metal, special type, etc.	Y	50403	Paper moisture meters	N
40619	Coaxial standard mismatches	Y	50107	Temperature transducers	Y	601. Sound in air		
40621	Mobile communication test sets	Y	50108	Primary fixed-point cells and apparatus	N	60102	Sound calibrators	N
40622	Modulation meters	Y				60104	Microphones	N
40623	Network analyzers	Y				60106	Sound level meters	Y
40624	Noise figure meters	Y				603. Vibration		
40625	Noise generators	N				60301	Vibration calibrators	N
40626	Noise impulse simulators	Y				60302	Vibration transducers	N
40627	RF phase noise meters	N				60303	Vibration measuring instruments	N
40628	Coaxial noise sources	N				701. Photometry		
40635	RF power meters	Y				70101	Iluminance meters	N
40636	Diode power sensors	Y				702. Property of detectos & sources		
40637	Thermocouple power sensors	Y	502. Non contact thermometry			70102	Luminance meters	N
40638	Pulse generators	Y	50203	Optical pyrometers	N	70103	Total luminous flux meters	Y
40639	Radar test sets	Y	50204	Standard radiation thermometers	N	70104	Luminous intensity meters	Y
40640	RF signal generators	Y	50205	Thermal image apparatus	N	702. Property of detectos & sources		
40641	RF spectrum analyzers	Y	50206	Blackbody furnaces	Y	70202	Color temperature meters	Y
40642	RF speed guns	Y	50207	Others; ear thermometers, etc.	N	70203	Color temperature standard lamps	N
40643	Surge generators	Y				70204	Colorimeters; source color	Y
40644	SWR meters	N				70207	Laser power meters	N
40645	RF terminations	Y	503. Humidity			70208	Standard LED light sources	N
40646	Coaxial thermistor mounts	Y	50301	Dew-point hygrometers; chilled mirror, alumina thin film, etc.	N			
40650	RF voltmeters	Y						
40651	Vector voltmeters	Y						

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
70209	Total luminous flux standard lamps	N	703. Property of materials			70415	Optical multimeters	Y
			70301	Colorimeters; material color	Y	70416	Optical network analyzers	Y
70210	Optical detectors	N	70304	Color standard tiles	N	70417	Optical spectrum analyzers	Y
70211	Pyranometers and pyrheliometers	N	70306	Gloss meters	Y	70418	Optical time domain reflectometers; OTDR	Y
			70307	Gloss standard plates	Y			
70213	Display color analyzers; luminance, chromaticity, white balance, etc.	Y	70308	Haze meters	Y	70423	Return loss meters	Y
			70312	Lens meters	Y	70429	Frequency stabilized lasers and LDs	Y
			70315	Optical densitometers	Y			
70214	Luminous intensity standard lamps	N	70319	Reflectance meters	Y	70430	ASE light sources	Y
			70321	Refractometers	Y	70433	Optical power stabilized lasers and LDs	Y
70215	Spectral irradiance standard lamps	N	70323	Transmittance meters	Y	901. Chemical analysis		
			70325	Spectrophotometers including FT-IR spectrophotometers	Y	90101	Breath alcohol analyzers	N
70216	Total spectral radiant flux standard lamps	N	70326	Wavelength reference materials; absorption cell, bandpass filter, etc.	N	90102	Environmental air quality monitoring instruments	Y
						90103	Gas analyzers	Y
70217	Luminance standard sources	N	70326	Wavelength reference materials; absorption cell, bandpass filter, etc.	N	90104	Exhaust gas test instruments	Y
						90199		
70219	UV irradiance meters	N	704. Fiber optics			90199	Others; pH meter, Electrical conductivity meter	Y
70220	Spectral irradiance meters	Y	70402	Broadband light sources	Y			
70221	Total spectral radiant flux meters	Y	70410	Optical attenuators	Y			
			70412	Fiber-optic power meters	Y			
70222	Spectral radiance meters	Y	70413	Optical loss testers	Y			

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k=2. It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Balls	10201	(0.3 ~ 100) mm	$\sqrt{0.38^2 + (0.0046 \times l_o)^2} \mu\text{m}$	Measuring Machine, Standard/ SICT-CP-10201
Electrical/Mechanical comparators	10203	(0 ~ 5) mm	0.14 μm	Gauge Block/ SICT-CP-10203
Gauge block comparators (Differences of central length)	10204			Gauge Block/ SICT-CP-10204
Comparison		(0 ~ 10) μm	$\sqrt{24^2 + (0.33 \times l_o)^2} \text{ nm}$	
Direct		(0 ~ 10) mm	0.043 μm	
Dial/cylinder gauge testers	10206	(0 ~ 100) mm	$\sqrt{0.12^2 + (0.0030 \times l_o)^2} \mu\text{m}$	Laser Measurement Machine/ SICT-CP-10206
Doctor blades	10207	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10207
Distance meters; electrooptic/laser/ultrasonic	10208	(0 ~ 40) m	$\sqrt{0.28^2 + (0.005 \times l_o)^2} \text{ mm}$	Laser interferometer/ SICT-CP-10208
End bars	10209	(25 ~ 1 500) mm	$\sqrt{0.6^2 + (0.0016 \times l_o)^2} \mu\text{m}$	Linear measuring system/ SICT-CP-10209
Extensometers, linear displacement transducers	10210	(0 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{1.9^2 + (0.0042 \times l_o)^2} \mu\text{m}$ $\sqrt{7.8^2 + (0.058 \times l_o)^2} \mu\text{m}$	Gauge Block, Laser Measurement Machine/ SICT-CP-10210
Filler gauges	10211	(0 ~ 10) mm	1.2 μm	Measuring Machine, Standard/ SICT-CP-10211
Film applicators	10212	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10212
Gap gauges	10213	(1 ~ 500) mm	$\sqrt{0.72^2 + (0.0048 \times l_o)^2} \mu\text{m}$	Gauge Block, contact coordinate measuring machines/ SICT-CP-10213
Gauge blocks, by comparison	10214	(0.1 ~ 100) mm (100 ~ 250) mm (250 ~ 500) mm	$\sqrt{70^2 + (1.3 \times l_o)^2} \text{ nm}$ $\sqrt{80^2 + (0.71 \times l_o)^2} \text{ nm}$ $\sqrt{152^2 + (0.71 \times l_o)^2} \text{ nm}$	Gauge Block Comparator/ SICT-CP-10214
Height gauges/measuring machines	10216	(0 ~ 1 500) mm	$\sqrt{0.68^2 + (0.0035 \times l_o)^2} \mu\text{m}$	Gauge Block, Step gauge/ SICT-CP-10216
Linear scales	10219	(0 ~ 40) m	$\sqrt{0.03^2 + (0.0027 \times l_o)^2} \text{ mm}$	Laser interferometer/ SICT-CP-10219
Standard measuring machines	10220	(0 ~ 500) mm	$\sqrt{0.38^2 + (0.002 \times l_o)^2} \mu\text{m}$	Gauge Block/ SICT-CP-10220
Micro scales/Standard scales	10221	(0.01 ~ 1 500) mm	$\sqrt{0.17^2 + (0.0003 \times l_o)^2} \mu\text{m}$	Linear measuring system/ SICT-CP-10221
Electronic micrometers	10223	(0 ~ 5) mm	0.14 μm	Gage Block/ SICT-CP-10223
Height micrometers, riser blocks Block	10224	(0 ~ 600) mm	$\sqrt{1.1^2 + (0.0019 \times l_o)^2} \mu\text{m}$	Gauge Block Electronic Micrometer/ SICT-CP-10224
Head		(0 ~ 25) mm	1.2 μm	
Laser scan micrometers	10225	(0.5 ~ 85) mm	$\sqrt{0.46^2 + (0.003 \times l_o)^2} \mu\text{m}$	Cylindrical plug/pin gauge/ SICT-CP-10225
Standard tape rules, peripheral gauges	10227	(0 ~ 40) m (40 ~ 80) m (80 ~ 100) m	$\sqrt{0.22^2 + (0.0046 \times l_o)^2} \text{ mm}$ $\sqrt{0.25^2 + (0.0046 \times l_o)^2} \text{ mm}$ $\sqrt{0.34^2 + (0.0046 \times l_o)^2} \text{ mm}$	Laser Measurement Machine/ SICT-CP-10227

Note 1. l_o unit : mm (10208,10227 unit : m)

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cylindrical plug/pin gauges, Thread measuring wire gauges	10228			Mesuring Machine, Standard/ SICT-CP-10228
Cylindrical plug/pin gauges		(0.01 ~ 200) mm	$\sqrt{0.42^2 + (0.003 \times l_0)^2} \mu\text{m}$	
Thread measuring wire gauges		(ø0.1 ~ ø10) mm	0.41 μm	
Radius gauges	10229	(0.1 ~ 100) mm	2.8 μm	Non-contact coordinate measuring machines,
Cylindrical ring gauges	10230	(1.0 ~ 100) mm	$\sqrt{0.55^2 + (0.003 0 \times l_0)^2} \mu\text{m}$	Mesuring Machine, Standard/ SICT-CP-10230
		(100 ~ 300) mm	$\sqrt{1.0^2 + (0.003 0 \times l_0)^2} \mu\text{m}$	
Step blocks	10231	(0 ~ 300) μm	0.23 μm	Gauge Block/ SICT-CP-10231
Step gauges	10232	(0 ~ 1 510) mm	$\sqrt{0.28^2 + (0.000 95 \times l_0)^2} \mu\text{m}$	Linear measuring system/ SICT-CP-10232
Taper thickness gauges	10233	(0 ~ 60) mm	4.3 μm	Profile Projector/ SICT-CP-10223
Ultrasonic thickness gauges	10234	(2.5 ~ 100) mm	3 μm	Ultrasonic Tester Blocks/ SICT-CP-10234
Ultrasonic/coating thickness specimens	10235			Gauge Block,Mesuring Machine, Standard/
coating thickness specimens		(0.01 ~ 25) mm	1.9 μm	SICT-CP-10235
Plateness			0.8 μm	
Ultrasonic thickness specimens		(0.5 ~ 500) mm	$\sqrt{0.64^2 + (0.006 \times l_0)^2} \mu\text{m}$	
Coating thickness testers	10236	(0 ~ 25) mm	1.2 μm	Thickness specimens/ SICT-CP-10236
Torque arms	10237			Gauge Block, contact coordinate measuring machines/
Torque arms		(1 ~ 1 500) mm	$\sqrt{0.60^2 + (0.006 1 \times l_0)^2} \mu\text{m}$	SICT-CP-10237
Wires		(0 ~ 5) mm	1.2 μm	
Width measuring specimens	10238			Mesuring Machine, contact coordinate measuring machines/ SICT-CP-10237
		(0 ~ 200) mm	$\sqrt{1.3^2 + (0.003 4 \times l_0)^2} \mu\text{m}$	
		(200 ~ 1 000) mm	$\sqrt{1.2^2 + (0.005 4 \times l_0)^2} \mu\text{m}$	

Note 1. l_0 unit : mm (10208,10227 unit : m)

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Angle gauge blocks	10302			Indexing tables/ SICT-CP-10302
Angle		(0 ~ 360)°	0.6"	
Autocollimators	10303			Fine angle generators/ SICT-CP-10303
Angle		(-1 000 ~ 1 000) "	0.5"	
Bevel protractors	10304			Angle Gauge Block, Coordinate Measur Machine/ SICT-CP-10304
Angle Accuracy		(0 ~ 360)°	0.9'	
Accessory Angle		(0 ~ 90)°	0.7'	
Clinometers	10306			Rotary tables/ SICT-CP-10306
Angle		(0 ~ 360)°	3.3"	
Collimators	10307			Total station/ SICT-CP-10307
Angle scale on the reticle		(-30 ~ 30) '	5"	
Fine angle generators, level comparators	10308			Autocollimators/ SICT-CP-10308
Angle		±1 000"	0.4"	
Indexing tables	10310			Polygons/ SICT-CP-10310
Angle		(0 ~ 360)°	0.5"	
Plate/Square/Electric levels	10311			Fine angle generators, Rotary tables/ SICT-CP-10311
Angle		(0 ~ 516) "	0.4"	
		(516 ~ 1 000) "	1.2"	
Inclino meter		(0 ~ 90)°	0.05'	
Squareness		(0 ~ 400) mm	1.8 μm	
Auto levels	10312			Calibration system for survey instrument, Indexing table, Rotary table/ SICT-CP-10312
Straightness of Line of Sight		(0.6 ~ ∞) m	0.03 mm	
Horizontal Angle		(0 ~ 360)°	1.3'	
Automatic level compensation range		(-30 ~ 30) °	5"	
Penta-prisms	10314			Autocollimators/ SICT-CP-10314
Angle		90°	0.6"	
Polygons	10315			Indexing tables/ SICT-CP-10315
Angle		(0 ~ 360)°	0.4"	
Rotary tables	10316			Polygons/ SICT-CP-10316
Angle		(0 ~ 360)°	0.5"	
Sine bars, plates, tables, centers	10317			Mesuring Machine, Standard/ SICT-CP-10317
(Sinebars)				
distance,between two roller center		(100 ~ 300) mm	$\sqrt{0.36^2 + (0.002 \times l_0)^2} \mu\text{m}$	
parallelism,between two roller		(100 ~ 300) mm	0.5 μm	
parallelism,between flat-two roller		(100 ~ 300) mm	0.6 μm	
(Plates)				
Center length		(100 ~ 300) mm	$\sqrt{0.12^2 + (0.028 \times l_0)^2} \mu\text{m}$	
Flatness		(100 ~ 300) mm	1.0 μm	
Parallelism		(100 ~ 300) mm	1.2 μm	
Squareness testers, right angle testers	10318	(0 ~ 600) mm	2.0 μm	Cylindrical Square, Precision Square/ SICT-CP-10318
Cylindrical squares	10319	(0 ~ 300) mm	1.6 μm	Cylindrical Square/ SICT-CP-10319
		(300 ~ 600) mm	1.9 μm	
Precision squares	10320			contact coordinate measuring machines/ SICT-CP-10320
Squareness		(0 ~ 600) mm	2.9 μm	
Parallelism		(0 ~ 600) mm	2.0 μm	
Theodolites, transits	10321			Calibration system for survey instrument, Indexing table/ SICT-CP-10321
Straightness of Line of Sight		(0.6 ~ ∞) m	0.09 mm	
Horizontal Angle		(0 ~ 360)°	1.3"	
Vertical Angle		(-45 ~ 45)°	1.3"	

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Angular displacement transducers Angle	10322	(0 ~ 360)°	2.9"	Rotary tables/ SICT-CP-10322
Alignment telescopes, line of sight collimators Straightness of Line of Sight Optical micrometer	10323	0 ~ ∞ (0 ~ 1.2) mm	0.022 mm 2 μm	Calibration system for survey instrument, Alignment telescope/ SICT-CP-10323
Calibration system for survey instruments Straightness of Line of Sight Horizontal Angle Vertical Angle	10324	(0.6 ~ ∞) m (0 ~ 360)° (-45 ~ 45)°	0.022 mm 2.1" 2.0"	Total station, Alignment telescope/ SICT-CP-10324
Jig transits Straightness of Line of Sight Horizontal Angle Vertical Angle	10325	(0.6 ~ ∞) m (0 ~ 360)° (-30 ~ 30)°	0.09 mm 1.3" 4.2"	Calibration system for survey instrument, Indexing table/ SICT-CP-10325
Laser levels Difference from absolute horizontal Difference to vertical about absolute horizontal Automatic level compensation range	10326	(0 ~ 4)' (0 ~ 4)' (-10 ~ 10) °	5" 8" 5"	Autocollimator, Rotary table/ SICT-CP-10326
Optical wedges Angular value on the wedge scale	10327	(-30 ~ 30) "	0.7"	Autocollimator, Rotary table/ SICT-CP-10327

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Form testers	10401	(0 ~ 50) mm	1.0 μm	Form Standard Specimens Gage Block, Angle Gage Block/ SICT-CP-10401
		Width (0 ~ 50) mm	1.4 μm	
		Angle 15° ~ 120°	2 °	
Optical flats	10404	ϕ (0 ~ 60) mm	0.05 μm	Optical Flat/ SICT-CP-10404
		ϕ (60 ~ 100) mm	0.08 μm	
Optical parallels	10405	ϕ (0 ~ 30) mm	0.05 μm	Optical Flat, Gauge block comparator/ SICT-CP-10405
		ϕ (0 ~ 30) mm	0.07 μm	
Parallel blocks	10406	(0 ~ 1 000) mm	1.5 μm	Electronic Micrometer/ SICT-CP-10406
		Flatness (0 ~ 1 000) mm	1.5 μm	
		Length Difference (0 ~ 1 000) mm	2.2 μm	
Precision surface plates	10407	(2 000 × 2 000) mm	2.0 μm	Electronic Level/ SICT-CP-10407
		(5 000 × 5 000) mm	4.8 μm	
Profile gauges	10408	(0 ~ 5) mm	0.7 μm	Gage Block/ SICT-CP-10408
Roundness measurement instruments	10409	Detector accuracy (0 ~ 20) μm	0.51 μm	Roundness Standard Ball/ SICT-CP-10409
		Rotational accuracy of spindle 360°	0.03 μm	
		Rotational accuracy of axis 360°	0.04 μm	
Form standard specimens	10410	Height (0 ~ 50) mm	0.9 μm	Standard measuring machine, Non-contact coordinate measuring machine/ SICT-CP-10410
		Width (0 ~ 100) mm	$\sqrt{0.59^2 + (0.0079 \times l_o)^2} \mu\text{m}$	
		Radius (2.5 ~ 10) mm	1.0 μm	
		Angle (10 ~ 50) mm (0 ~ 90) °	1.4 μm 0.007 °	
Roundness standard/roundness magnification standard specimens	10411	Standard specimens (0 ~ 300) μm	0.52 μm	Roundness Tester/ SICT-CP-10411
		Standard ball 360°	0.08 μm	
Straight edges	10412	Straightness (0 ~ 2 000) mm	1.8 μm	Electronic levels/ SICT-CP-10412
		Parallelism (0 ~ 2 000) mm	1.8 μm	
Straight rules	10413	Length (0 ~ 2 000) mm	0.10 mm	LASER INTERFEROMETER/ SICT-CP-10413
Test bars	10415	Roundness (0 ~ 400) mm	0.6 μm	Roundness Tester, Electronic Micrometer/ SICT-CP-10415
		Cylindricity (0 ~ 400) mm	0.6 μm	
		Run-out (0 ~ 400) mm	1.1 μm	
Spherometers	10416	(0 ~ 30) mm	0.3 μm	Gauge Block/ SICT-CP-10416

Note 1. l_o unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Base gauges for electric bulb Inside diameter of pass/stop and screw Pitch	10501	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.47^2 + (0.0028 \times l_0)^2} \mu\text{m}$ 1.6 μm	Mesuring Machine, Standard/ SICT-CP-10501
Bench centers Difference of both center Flatness of both bed Center height difference	10502	(0 ~ 200) mm (200 ~ 500) mm (0 ~ 500) mm (0 ~ 200) mm (200 ~ 500) mm	1.8 μm 3.4 μm 1.5 μm 1.8 μm 3.4 μm	Test Bar/ SICT-CP-10502
Contact coordinate measuring machines	10503	(0 ~ 1 500) mm (0 ~ 600) mm (0 ~ 600) mm	$\sqrt{0.56^2 + (0.0044 \times l_0)^2} \mu\text{m}$ 3.2 μm 1.2 μm	Step Gauge/ SICT-CP-10503
Non-contact coordinate measuring machines	10504	(0 ~ 1 000) mm	$\sqrt{0.43^2 + (0.0034 \times l_0)^2} \mu\text{m}$	Standard Scale/ SICT-CP-10504
Gauge block accessories Round the ministry of Justice A he ministry of Justice,Base block Center point Plane figure Parallelism(triangle edge) Parallelism(equilibrium tide)	10505	(0 ~ 50) mm (0 ~ 50) mm (0 ~ 20) mm (0 ~ 50) mm (0 ~ 300) mm (0 ~ 150) mm	$\sqrt{0.32^2 + (0.0044 \times l_0)^2} \mu\text{m}$ $\sqrt{0.26^2 + (0.0044 \times l_0)^2} \mu\text{m}$ 1.1 μm 0.04 μm 0.4 μm 0.4 μm	Gauge Block/ SICT-CP-10505
Hardness indenters Diameter Radius Angle	10508	(1 ~ 13) mm (0.2 ~ 6) mm (0 ~ 173)°	0.5 μm 1.0 μm 0.012 °	Standard measuring machine, Non-contact coordinate measuring machine/ SICT-CP-10410
Measuring microscopes, Profile projectors Length Magnification Angle	10511	(0 ~ 500) mm (5 ~ 100) 배 (0 ~ 360) °	$\sqrt{0.43^2 + (0.0034 \times l_0)^2} \mu\text{m}$ 0.04 % 0.9'	Standard Scale/ SICT-CP-10511
Micro measuring microscopes	10512	(0 ~ 1) mm (1 ~ 50) mm	0.7 μm 3.0 μm	Standard Scale/ SICT-CP-10512
Orifice plates Inside diameter Thickness	10513	(12.7 ~ 100) mm (100 ~ 300) mm (0 ~ 15) mm	$\sqrt{0.55^2 + (0.0030 \times l_0)^2} \mu\text{m}$ $\sqrt{1.0^2 + (0.0030 \times l_0)^2} \mu\text{m}$ 0.6 μm	Standard measuring machine/ SICT-CP-10513
Taper plug gauges Small end diameter Big end diameter Plane angle Gage height	10514	(2 ~ 200) mm (2 ~ 200) mm (0 ~ 90) ° (2 ~ 200) mm	$\sqrt{1.3^2 + (0.0041 \times l_0)^2} \mu\text{m}$ $\sqrt{1.4^2 + (0.0041 \times l_0)^2} \mu\text{m}$ 5.9" $\sqrt{1.2^2 + (0.0044 \times l_0)^2} \mu\text{m}$	Measuring Machine, Standard/ SICT-CP-10514
Taper ring gauges Small end diameter Big end diameter Plane angle	10515	(5 ~ 200) mm (5 ~ 200) mm (0 ~ 90) °	2.5 μm 2.5 μm 0.006"	contact coordinate measuring machines/ SICT-CP-10515

Note 1. l_0 unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Stylus type roughness testers Roughness parameter(Ra) Roughness parameter(Rz) Mean width(RSm) H,D	10517	(0 ~ 2) μm (2 ~ 10) μm (0 ~ 7) μm (7 ~ 30) μm (0 ~ 300) μm (0 ~ 6) μm (6 ~ 20) μm	9 nm 24 nm 77 nm 0.27 μm 1.3 μm 63 nm 97 nm	Roughness Specimen/ SICT-CP-10517
Socket gauges for electric bulb Outside diameter of pass/stop and screw Pitch	10518	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.44^2 + (0.0028 \times l_0)^2} \mu\text{m}$ 1.6 μm	Measuring Machine, Standard/ SICT-CP-10518
Roughness standard/comparison specimens Depth(H) Mean width(RSm) Roughness parameter(Ra) Roughness parameter(Rz)	10519	(0 ~ 6) μm (6 ~ 20) μm (0 ~ 300) μm (0 ~ 2) μm (2 ~ 10) μm (0 ~ 7) μm (7 ~ 30) μm	$\sqrt{(9.6 \times R)^2 + 12^2} \text{ nm}$ $\sqrt{(8.94 \times R)^2 + 15^2} \text{ nm}$ $\sqrt{(0.01 \times R)^2 + 0.51^2} \mu\text{m}$ $\sqrt{(10 \times R)^2 + 2.6^2} \text{ nm}$ $\sqrt{(9.2 \times R)^2 + 7.4^2} \text{ nm}$ $\sqrt{(29.2 \times R)^2 + 7.4^2} \text{ nm}$ $\sqrt{(0.025 \times R)^2 + 0.096^2} \mu\text{m}$	Roughness Tester/ SICT-CP-10519
Thread plug gauges Outside diameter Effective diameter Pitch Half angle	10525	(1 ~ 205) mm (1 ~ 210) mm (0.3 ~ 10) mm (0.5 ~ 45) °	1.7 μm 1.1 μm 1.2 μm 2'	Measuring Machine, Standard/ SICT-CP-10525
Taper thread plug gauges Half angle Pitch Gage length Notch length Small outside diameter Big outside diameter Small effective diameter Big effective diameter	10526	(0 ~ 45) ° (0.3 ~ 6) mm (2 ~ 50) mm (2 ~ 50) mm (2 ~ 200) mm (2 ~ 200) mm (2 ~ 200) mm (2 ~ 200) mm	2' 1.3 μm 2.6 μm 3.6 μm 2.3 μm 4.8 μm 2.9 μm 5.1 μm	Measuring Machine, Standard/ SICT-CP-10526
Thread ring gauges Minor diameter Effective diameter Pitch	10527	(3 ~ 200) mm (3 ~ 200) mm (0.3 ~ 10) mm	1.5 μm 2.3 μm 1.6 μm	Measuring Machine, Standard/ SICT-CP-10527
Taper thread ring gauges Alternation Thickness Notch length	10528	±3 mm (0 ~ 100) mm (0 ~ 100) mm	2.7 μm 2.3 μm 3.2 μm	Measuring Machine, Standard/ SICT-CP-10528
V-blocks, box blocks Plane figure Parallelism Difference of both part	10529	(5 ~ 300) mm (5 ~ 300) mm (5 ~ 300) mm	1.7 μm 2.0 μm 2.8 μm	contact coordinate measuring machines/ SICT-CP-10529
Position gauges Distance Diameter Angle	10530	(0 ~ 1 000) mm (2 ~ 12) mm (12 ~ 200) mm (0 ~ 360) °	$\sqrt{4.9^2 + (0.0054 \times l)^2} \mu\text{m}$ $\sqrt{2.4^2 + (0.0028 \times l)^2} \mu\text{m}$ $\sqrt{3.3^2 + (0.0028 \times l)^2} \mu\text{m}$ 5.2"	Contact coordinate measuring machine/ SICT-CP-10530
SEM/TEM/SPM/AFM microscopes Magnification	10531	(5 ~ 100) × (100 ~ 500 000) ×	0.003 5 0.003 0	Magnification reference specimen/ SICT-CP-10531

Note 1. l_0 unit : mm, R unit : μm

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/outside/gear tooth calipers, caliper gauges	10601			Gauge Block/ SICT-CP-10601
Inside/Outside calipers		(0 ~ 2 000) mm	$\sqrt{8.2^2 + (0.007 \times l_0)^2} \mu\text{m}$	
Caliper gauges		(0 ~ 300) mm	$\sqrt{3.7^2 + (0.0032 \times l_0)^2} \mu\text{m}$	
Cylinder/bore gauges	10603			Dial Gauge Tester/ SICT-CP-10603
Cylinder gauges		(0 ~ 1 000) mm	1.0 μm	
Hole gauges		(0.1 ~ 25) mm	3.3 μm	
Depth gauges, depth micrometers	10604			Gauge Block/ SICT-CP-10604
Depth micrometers		(0 ~ 300) mm	$\sqrt{0.86^2 + (0.0034 \times l_0)^2} \mu\text{m}$	
Depth gauges		(0 ~ 1 000) mm	$\sqrt{5.9^2 + (0.0048 \times l_0)^2} \mu\text{m}$	
Dial/digital gauges	10605			Dial Gauge Tester/
		(0 ~ 100) mm	$\sqrt{0.21^2 + (0.0082 \times l_0)^2} \mu\text{m}$	
		(0 ~ 25) mm	$\sqrt{0.59^2 + (0.004 \times l_0)^2} \mu\text{m}$	
Geodesic baselines	10606		$\sqrt{1.7^2 + 0.0033^2 \times l^2} \text{ mm}$	Total station/ SICT-CP-10606
Grind gauges	10608			Electronic micrometer/ SICT-CP-10608
Depth		(0 ~ 1) mm	1.8 μm	
Straightness		(0 ~ 150) mm	2.5 μm	
Micro indicators, test indicators	10609			Dial Gauge Tester/ SICT-CP-10609
		(0 ~ 5) mm	0.22 μm	
Micrometer heads	10610	(0 ~ 50) mm	0.8 μm	Gauge Block/ SICT-CP-10610
3-points, Micrometers	10611	(2 ~ 200) mm	$\sqrt{1.3^2 + (0.0034 \times l_0)^2} \mu\text{m}$	Ring Gauge/ SICT-CP-10611
		(200 ~ 300) mm	3 μm	
Inside micrometers	10612			Gauge Block/ SICT-CP-10612
Length		(5 ~ 300) mm	$\sqrt{1.1^2 + (0.0042 \times l_0)^2} \mu\text{m}$	
Accuracy of scale		(25 ~ 500) mm	$\sqrt{1.1^2 + (0.0042 \times l_0)^2} \mu\text{m}$	
Extension rod		(13 ~ 500) mm	$\sqrt{1.2^2 + (0.0048 \times l_0)^2} \mu\text{m}$	
Outside micrometers	10613			Gauge Block, cylindrical plug gauges/ SICT-CP-10613
Outside micrometers		(0 ~ 25) mm	$\sqrt{0.2^2 + (0.003 \times l_0)^2} \mu\text{m}$	
V-anvil micrometers		(25 ~ 1 000) mm	$\sqrt{0.83^2 + (0.003 \times l_0)^2} \mu\text{m}$	
		(1 ~ 85) mm	0.8 μm	
Particle counters	10615			Particle calibration system/ SICT-CP-10615
(Air)		(0.1 ~ 1) μm		
Flow		(0 ~ 100) L/min	0.09 L/min	
Threshold voltage		(0 ~ 10) V	0.42 mV	
Counting efficiency		(0 ~ 110) %	4.1 %	
(Liquid)		(0.05 ~ 25) μm		
Flow		(0 ~ 100) mL/min	1.4 mL/min	
Threshold voltage		(0 ~ 10) V	0.42 mV	

Note 1. l_0 unit : mm (10606 unit : m)

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard sieves	10617			Non-contact coordinate measuring machines/ SICT-CP-10617
Sieve opening Wire rod diameter		(0.004 ~ 10) mm (0.004 ~ 130) mm	1.5 μm 2.4 μm	
Water level meters	10619	(0.05 ~ 6.5) m	1.6 mm	Laser distance meter/ SICT-CP-10619
Welding gauges	10620			Non-contact coordinate measuring machine,Gauge Block/ SICT-CP-10620
Height or depth Rule Angle		(0 ~ 100) mm (0 ~ 100) mm (0 ~ 90)°	8.2 μm 6.0 μm 0.7'	
Optical micrometers	10621			Standard Scale/ SICT-CP-10621
Optical axis shift		(0 ~ 1.2) mm (1.2 ~ 5) mm (5 ~ 10) mm	2 μm 3 μm 0.03 mm	

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Auto-hopper scale balances	20102	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	1.2 g 14 g 21 g 49 g 0.10 kg	Hopper Scale Weight/ SICT-CP-20102
Auto-packer scale balances	20103	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg	0.8 g 7.7 g 16 g	Weight/ SICT-CP-20103
Axle weigher balances Portable	20104	(100 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg (5 000 ~ 10 000) kg (10 000 ~ 30 000) kg	0.2 kg 0.4 kg 1 kg 5 kg 10 kg 24 kg	Force Calibration Machine/ SICT-CP-20104
Dial platform scale balances	20106	(0 ~ 30) kg (30 ~ 60) kg (60 ~ 100) kg	42 g 0.08 kg 0.21 kg	Weight/ SICT-CP-20106
Dial swing scale balances	20107	(0 ~ 1) kg (1 ~ 10) kg (10 ~ 20) kg (20 ~ 50) kg (50 ~ 100) kg (100 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	0.96 g 9.6 g 20 g 48 g 96 g 0.23 kg 0.48 kg 0.96 kg 1.9 kg 4.6 kg	Weight/ SICT-CP-20107
Electric balances	20109	(0 ~ 2) g (2 ~ 6) g (6 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 200) g (200 ~ 500) g (500 ~ 1 000) g (1 ~ 2) kg (2 ~ 5) kg (5 ~ 10) kg (10 ~ 25) kg (25 ~ 40) kg (40 ~ 60) kg (60 ~ 150) kg (150 ~ 600) kg (600 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	7.0 µg 9.3 µg 14 µg 19 µg 29 µg 0.05 mg 0.10 mg 0.20 mg 0.5 mg 1.0 mg 3 mg 6 mg 16 mg 24 mg 0.30 g 1.2 g 2.0 g 38 g 0.10 kg	Weight/ SICT-CP-20109

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Manual swing scale balances	20111	(0 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	38 g 0.19 kg 0.38 kg 0.94 kg	Weight/ SICT-CP-20111
Platform scale balances	20112	(0 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	46 g 0.09 kg 0.46 kg	Weight/ SICT-CP-20112
Spring scale balances	20113	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 100) kg	21 g 0.08 kg 0.21 kg	Weight/ SICT-CP-20113
Weights	20116	(1 mg ~ 20 kg) 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg (20 ~ 100) kg 50 kg 100 kg (100 ~ 200) kg 200 kg (200 ~ 1 000) kg 500 kg 1 000 kg	(Class E2) 1.0 µg 1.0 µg 1.0 µg 1.0 µg 1.2 µg 1.4 µg 1.7 µg 2.3 µg 2.9 µg 3.5 µg 4.6 µg 5.8 µg 7.0 µg 9.3 µg 12 µg 18 µg 36 µg 95 µg 0.12 mg 0.36 mg 0.94 mg 1.8 mg 3.7 mg (Class F2) 0.12 g 0.21 g (Class M1) 1.0 g (Class F2) 1.3 g 2.1 g	Weights, Mass Comparator/ SICT-CP-20116

202. Force

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Force measuring devices	20202	(0.4 ~ 20) N (20 ~ 50) N (50 ~ 100) N (100 ~ 200) N (200 ~ 500) N (0.5 ~ 1) kN (1 ~ 2) kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 200) kN (200 ~ 500) kN (500 ~ 1 000) kN	6.0×10^{-4} 8.4×10^{-5} 8.5×10^{-5} 7.3×10^{-5} 6.9×10^{-5} 9.2×10^{-5} 9.0×10^{-5} 8.4×10^{-5} 8.7×10^{-5} 3.6×10^{-4} 4.1×10^{-4} 4.8×10^{-4} 4.5×10^{-4} 4.3×10^{-4} 4.5×10^{-4}	Force Calibration Machine/ SICT-CP-20202
Tension/compression testing machines	20203			Weights, Force Measuring Device/ SICT-CP-20203
	tensile	0.1 N ~ 2 kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN	1.2×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3}	
	compression	(0.1 ~ 50) N (50 ~ 100) N (100 ~ 200) N (200 ~ 500) N (0.5 ~ 1) kN (1 ~ 2) kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN (300 ~ 500) kN (500 ~ 1 000) kN (1 000 ~ 3 000) kN	1.2×10^{-3} 1.5×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.6×10^{-3}	
push-pull gauge	20204	(0.02 ~ 0.2) N 0.2 N ~ 2 kN (2 ~ 5) kN	1.4×10^{-2} 1.3×10^{-3} 8.4×10^{-4}	Weights, Force Calibration Machine/ SICT-CP-20204

203. Torque

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Torque measuring devices	20302	(0.001 ~ 1) N·m (1 ~ 10) N·m (10 ~ 20) N·m (20 ~ 50) N·m (50 ~ 100) N·m (100 ~ 200) N·m (200 ~ 500) N·m (500 ~ 1 000) N·m (1 000 ~ 2 000) N·m	3.3×10^{-3} 4.1×10^{-4} 7.5×10^{-4} 2.3×10^{-4} 3.4×10^{-4} 2.1×10^{-4} 2.2×10^{-4} 1.5×10^{-4} 1.6×10^{-4}	Torque Calibration Machine/ SICT-CP-20302
Torque wrenches/drivers	20303	(0.02 ~ 0.1) N·m (0.1 ~ 0.5) N·m (0.5 ~ 1) N·m (1 ~ 2) N·m (2 ~ 5) N·m (5 ~ 10) N·m (10 ~ 20) N·m (20 ~ 50) N·m (50 ~ 100) N·m (100 ~ 200) N·m (200 ~ 500) N·m (500 ~ 1 000) N·m (1 000 ~ 2 000) N·m	1.4×10^{-2} 9.5×10^{-3} 7.8×10^{-3} 6.2×10^{-3} 4.6×10^{-3} 4.5×10^{-3} 4.7×10^{-3} 4.5×10^{-3} 4.9×10^{-3} 3.8×10^{-3} 3.7×10^{-3} 3.8×10^{-3} 2.8×10^{-3}	Torque Measuring Device/ SICT-CP-20303

204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Altimeters	20401	(-4 000 ~ 20 000) m (20 000 ~ 47 000) m	12 m 15 m	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20401
Manometers	20402	(0 ~ 200) kPa	2.9×10^{-3}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20402
Pneumatic pressure ballances	20403	(5 ~ 7 000) kPa	4.2×10^{-5}	Dead Weight Tester/ SICT-CP-20403
Hydraulic pressure ballances	20404	(0.2 ~ 20) MPa (20 ~ 200) MPa	6.4×10^{-5} 7.0×10^{-5}	Dead Weight Tester/ SICT-CP-20404
Air data test systems	20405	(-2 500 ~ 20 000) m (20 000 ~ 30 500) m	0.8 m 7 m	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20405
		Dynamic pressure	(0 ~ 342) km/hr (342 ~ 2 122) km/hr	0.1 km/hr 0.3 km/hr
Absolute pressure gauges	20406	(5 ~ 350) kPa abs (350 ~ 7 000) kPa abs (7 ~ 200) MPa abs	4.0×10^{-5} 4.3×10^{-5} 7.4×10^{-5}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20406
Blood pressure gauges	20407	(0 ~ 40) kPa	0.01 kPa	Digital Manometer/ SICT-CP-20407
Compound pressure gauges	20408	(-95 ~ 7 000) kPa	4.4×10^{-5}	Air Dead Weight Tester/ SICT-CP-20408
Differential pressure gauges	20409	(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 350) kPa (350 ~ 7 000) kPa	0.10 Pa 1.0 Pa 4.2×10^{-5} 4.6×10^{-5}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20409
Gauge pressure gauges	20411	(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 350) kPa (350 ~ 7 000) kPa (7 ~ 200) MPa (200 ~ 500) MPa	0.10 Pa 1.0 Pa 4.2×10^{-5} 4.6×10^{-5} 8.6×10^{-5} 2.6×10^{-4}	Digital Manometer, Air Dead Weight Tester, Oil Dead Weight Tester/ SICT-CP-20411
Pressure transducers/transmitters	20412	(5 ~ 5 000) kPa abs (5 ~ 200) MPa abs	2.2×10^{-4} 2.4×10^{-4}	Digital Manometer, Air Dead Weight Tester, Oil Dead Weight Tester/ SICT-CP-20412
		Gauge pressure	(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 5 000) kPa (5 ~ 500) MPa	0.10 Pa 1.0 Pa 2.2×10^{-4} 2.4×10^{-4}
Dial type vacuum gauges	20413	(-95 ~ 0) kPa	0.059 kPa	Air Dead Weight Tester, SICT-CP-20413
Water depth meters	20414	(0 ~ 198.12) m (198.12 ~ 350.52) m	0.062 m 0.46 m	Digital Manometer/ SICT-CP-20414

205. Vacuum

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance diaphragm gauges	20501	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 12 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20501
Spinning rotor gauges	20502	0.15 mPa abs ~ 0.01 Pa abs	3.4×10^{-2}	Baratron gauge, SRG / SICT-CP-20502
Ionization gauges	20503	0.093 mPa abs ~ 0.15 mPa abs 0.15 mPa abs ~ 0.01 Pa abs	6.0×10^{-2} 3.5×10^{-2}	Baratron gauge, SRG, Ion / SICT-CP-20503
Thermal conductivity gauges; pirani, thermocouple, convectron, etc.	20504	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 13 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20504
Standard leaks, Helium leak detectors	20505	22.0 $\mu\text{Pa m}^3/\text{s}$ 1.60 $\mu\text{Pa m}^3/\text{s}$ 0.51 $\mu\text{Pa m}^3/\text{s}$ 15.0 $\text{nPa m}^3/\text{s}$ 6.4 $\text{nPa m}^3/\text{s}$ 0.24 $\text{nPa m}^3/\text{s}$	4.8 $\mu\text{Pa m}^3/\text{s}$ 0.38 $\mu\text{Pa m}^3/\text{s}$ 0.098 $\mu\text{Pa m}^3/\text{s}$ 3.2 $\text{nPa m}^3/\text{s}$ 1.3 $\text{nPa m}^3/\text{s}$ 0.049 $\text{nPa m}^3/\text{s}$	Standard leaks, Helium leak detectors / SICT-CP-20505

206. Volume

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Volumetric glasswares	20601	(0 ~ 0.1) ml (0.1 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (2 ~ 10) ml (10 ~ 25) ml (25 ~ 50) ml (50 ~ 100) ml (100 ~ 200) ml (200 ~ 250) ml (250 ~ 500) ml (500 ~ 1 000) ml (1 000 ~ 2 000) ml (2 000 ~ 5 000) ml (5 000 ~ 10 000) ml	0.31 μ l 0.40 μ l 0.43 μ l 1.1 μ l 1.7 μ l 2.5 μ l 4.3 μ l 5.1 μ l 7.9 μ l 13 μ l 43 μ l 69 μ l 92 μ l 0.17 ml 0.49 ml 0.87 ml	Weight, balances / SICT-CP-20601
Pycnometers	20602	(0 ~ 50) ml (50 ~ 100) ml (100 ~ 250) ml (250 ~ 500) ml	2.4 μ l 4.4 μ l 10 μ l 21 μ l	Weight, balances / SICT-CP-20602
Rain gauges	20603	tipping bucket type : Rainfall intensity : (5 ~ 300) mm/h (0.1 ~ 1) mm standard type : (0.1 ~ 10) mm (10 ~ 50.8) mm (50.8 ~ 178) mm	1.2×10^{-3} 0.021 mm 0.054 mm 0.30 mm	Weight, balances / SICT-CP-20603
Standard volume vessels	20604	(0 ~ 20) L (20 ~ 200) L (200 ~ 10 000) L	9.0×10^{-5} 1.3×10^{-4} 1.1×10^{-3}	Balances, Master Meter, Standard volume vessel/ SICT-CP-20604
Concrete air content meters	20605	(0 ~ 10) %	0.032 %	Weight, balances / SICT-CP-20605
Piston type volume meters	20606	(0 ~ 1) μ l (1 ~ 2) μ l (2 ~ 5) μ l (5 ~ 10) μ l (0.01 ~ 0.02) ml (0.02 ~ 0.05) ml (0.05 ~ 0.1) ml (0.1 ~ 0.2) ml (0.2 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (5 ~ 10) ml (10 ~ 20) ml (20 ~ 50) ml (50 ~ 100) ml	0.004 μ l 0.005 μ l 0.006 μ l 0.008 μ l 0.018 μ l 0.035 μ l 0.047 μ l 0.14 μ l 0.35 μ l 0.65 μ l 1.6 μ l 1.9 μ l 2.4 μ l 5.0 μ l 14 μ l 64 μ l	Weight, balances / SICT-CP-20606

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid density meters	20702	(0.7 ~ 1.4) g/cm ³	0.000 078 g/cm ³	Density standard materials/ SICT-CP-20702
Salinity meters	20704	(0 ~ 10) % (10 ~ 26) %	0.004 % 0.007 %	NaCl/ SICT-CP-20704
Sucrose meters	20705	(0 ~ 20) % (20 ~ 60) % (60 ~ 80) %	0.014 % 0.016 % 0.025 %	Sucrose/ SICT-CP-20705
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706			
density		(0.600 ~ 0.700) g/cm ³	0.000 035 g/cm ³	Solid density standard material,
		(0.700 ~ 0.800) g/cm ³	0.000 038 g/cm ³	Hydrostatic weighing Apparatus/ SICT-CP-20706-1
		(0.800 ~ 0.900) g/cm ³	0.000 042 g/cm ³	
		(0.900 ~ 1.000) g/cm ³	0.000 046 g/cm ³	
		(1.000 ~ 1.100) g/cm ³	0.000 050 g/cm ³	
		(1.100 ~ 1.200) g/cm ³	0.000 055 g/cm ³	
		(1.200 ~ 1.300) g/cm ³	0.000 059 g/cm ³	
		(1.300 ~ 1.400) g/cm ³	0.000 066 g/cm ³	
		(1.400 ~ 1.500) g/cm ³	0.000 071 g/cm ³	
		(1.500 ~ 1.600) g/cm ³	0.000 075 g/cm ³	
		(1.600 ~ 1.700) g/cm ³	0.000 079 g/cm ³	
		(1.700 ~ 1.800) g/cm ³	0.000 084 g/cm ³	
		(1.800 ~ 1.900) g/cm ³	0.000 088 g/cm ³	
		(1.900 ~ 2.000) g/cm ³	0.000 093 g/cm ³	
		(2.000 ~ 2.200) g/cm ³	0.000 25 g/cm ³	
		(2.200 ~ 3.000) g/cm ³	0.000 28 g/cm ³	
		(3.000 ~ 3.600) g/cm ³	0.000 30 g/cm ³	
		(3.600 ~ 4.000) g/cm ³	0.000 32 g/cm ³	
specific gravity		0.590 ~ 0.700	0.000 068	Solid density standard material,
		0.700 ~ 0.800	0.000 069	Hydrostatic weighing Apparatus/ SICT-CP-20706-2
		0.800 ~ 0.900	0.000 072	
		0.900 ~ 1.000	0.000 075	
		1.000 ~ 1.100	0.000 078	
		1.100 ~ 1.200	0.000 082	
		1.200 ~ 1.300	0.000 086	
		1.300 ~ 1.400	0.000 091	
		1.400 ~ 1.500	0.000 096	
		1.500 ~ 1.600	0.000 10	
		1.600 ~ 1.800	0.000 11	
		1.800 ~ 2.000	0.000 12	
		2.000 ~ 2.020	0.000 26	
		2.020 ~ 2.500	0.000 60	
		2.500 ~ 3.000	0.000 61	

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706			
alcohol (Volumn)		(0 ~ 10) % (10 ~ 30) % (30 ~ 40) % (40 ~ 50) % (50 ~ 60) % (60 ~ 70) % (70 ~ 80) % (80 ~ 90) % (90 ~ 100) %	0.039 % 0.043 % 0.038 % 0.030 % 0.025 % 0.023 % 0.020 % 0.019 % 0.017 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-3
API		-1 ~ 51 51 ~ 101	0.013 0.014	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-4
Baumé-light		10 ~ 30 30 ~ 40 40 ~ 60 60 ~ 70 70 ~ 100	0.015 0.016 0.018 0.019 0.12	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
Baumé - heavy		0 ~ 40 40 ~ 75	0.014 0.013	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
sugar		(0 ~ 10) % (10 ~ 90) %	0.018 % 0.017 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-6
milk		(15 ~ 20) (20 ~ 40)	0.081 0.082	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-7
Bouyoucos		(-5.0 ~ 60.0) g/L	0.14 g/L	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-8
salinity		(0 ~ 26.4) %	0.025 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-9
LPG		(0.50 ~ 0.55) g/cm ³ (0.55 ~ 0.60) g/cm ³ (0.60 ~ 0.65) g/cm ³	0.000 065 g/cm ³ 0.000 066 g/cm ³ 0.000 068 g/cm ³	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-10

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706 Twadde11	0 ~ 12 12 ~ 74 74 ~ 102 102 ~ 170 170 ~ 200	0.016 0.059 0.060 0.061 0.062	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-11
Chloride meters	20707	(0 ~ 0.1) % (0.1 ~ 2.0) %	0.000 2 % 0.001 0 %	Chlorine standard liquid/ SICT-CP-20707
Others; Solid density	20799	(1 ~ 2) g (2 ~ 5) g (5 ~ 10) g (10 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 500) g	0.004 6 g/cm ³ 0.002 4 g/cm ³ 0.001 0 g/cm ³ 0.000 59 g/cm ³ 0.000 43 g/cm ³ 0.000 37 g/cm ³ 0.000 36 g/cm ³	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-11
		(1 ~ 2) g (2 ~ 5) g (5 ~ 10) g (10 ~ 20) g (20 ~ 500) g	0.000 46 g/cm ³ 0.000 25 g/cm ³ 0.000 14 g/cm ³ 0.000 12 g/cm ³ 0.000 11 g/cm ³	

* 20704, 20705, 20706, 20707 unit % is weight percent.

208. Viscosity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Kinematic viscometers; capillary, etc.	20801	(2.5 ~ 100 000) mPa·s	1.3×10^{-2}	Viscosity CRM/ SICT-CP-20801
Dynamic viscometers; rotaional, etc. Viscosity	20802	(2.5 ~ 200 000) mPa·s	1.7×10^{-2}	Viscosity CRM/ SICT-CP-20802

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Anemometers; hot-wire	20901	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20901
Anemometers; pitot tube, etc.	20902	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20902
Gas flowmeters; differential pressure	20908	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; differential pressure	20909	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Liquid flowmeters; electromagnetic	20910	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; thermal mass, etc.	20911	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; Coriolis, etc.	20912	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; positive displacement	20914	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; positive displacement	20915	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; turbine	20916	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; turbine	20917	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; ultrasonic	20918	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid flowmeters; ultrasonic	20919	(0.01 ~ 50) m^3/h	1.2×10^{-3}	Master Meter/SICT-CP-20926
		(0.000 12 ~ 0.01) m^3/h	2.6×10^{-3}	Weight measuring method/ SICT-CP-20927
		(0.01 ~ 50) m^3/h	7.0×10^{-4}	
Gas flowmeters; variable area	20920	(1.2×10^{-5} ~ 0.12) m^3/h	1.9×10^{-3}	Sonic Nozzle/SICT-CP-20928
		(0.12 ~ 300) m^3/h	2.0×10^{-3}	Master Meter/SICT-CP-20929
		(300 ~ 4 000) m^3/h	3.6×10^{-3}	
Liquid flowmeters; variable area	20921	(0.01 ~ 50) m^3/h	1.2×10^{-3}	Master Meter/SICT-CP-20926
		(0.000 12 ~ 0.01) m^3/h	2.6×10^{-3}	Weight measuring method/ SICT-CP-20927
		(0.01 ~ 50) m^3/h	7.0×10^{-4}	
Gas flowmeters; vortex	20922	(1.2×10^{-5} ~ 0.12) m^3/h	1.9×10^{-3}	Sonic Nozzle/SICT-CP-20928
		(0.12 ~ 300) m^3/h	2.0×10^{-3}	Master Meter/SICT-CP-20929
		(300 ~ 4 000) m^3/h	3.6×10^{-3}	
Liquid flowmeters; vortex	20923	(0.01 ~ 50) m^3/h	1.2×10^{-3}	Master Meter/SICT-CP-20926
		(0.000 12 ~ 0.01) m^3/h	2.6×10^{-3}	Weight measuring method/ SICT-CP-20927
		(0.01 ~ 50) m^3/h	7.0×10^{-4}	
Anemometers; vane, etc	20925	(0.1 ~ 1.0) m/s	8.7×10^{-2}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20925
		(1.0 ~ 2.0) m/s	8.4×10^{-3}	
		(2.0 ~ 70) m/s	4.8×10^{-3}	
Others; Anemometers; Ultrasonic current meter	20999	(0.1 ~ 1.0) m/s	8.7×10^{-2}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20999
		(1.0 ~ 2.0) m/s	8.4×10^{-3}	
		(2.0 ~ 70) m/s	4.8×10^{-3}	

210. Hardness

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Brinell hardness testers	21001	(100 ~ 250) HBW 10/3 000 (250 ~ 450) HBW 10/3 000 (450 ~ 650) HBW 10/3 000	3.1 HBW 10/3 000 4.9 HBW 10/3 000 8.2 HBW 10/3 000	Brinell Hardness CRM/ SICT-CP-21001
Rockwell hardness testers	21002	(20 ~ 70) HRC (20 ~ 100) HRBW (42 ~ 86) HR30N (29 ~ 82) HR30TW	0.45 HRC 0.80 HRBW 0.70 HR30N 1.1 HR30TW	Rockwell Hardness CRM/ SICT-CP-21002
Shore hardness testers	21003	(20 ~ 100) HS	0.9 HS	Shore Hardness CRM/ SICT-CP-21003
Vickers hardness testers	21004	(50 ~ 300) HV 0.2 (300 ~ 600) HV 0.2 (600 ~ 850) HV 0.2 (50 ~ 300) HV 0.3 (300 ~ 600) HV 0.3 (600 ~ 850) HV 0.5 (50 ~ 300) HV 0.5 (300 ~ 600) HV 0.5 (600 ~ 850) HV 1 (50 ~ 300) HV 10 (300 ~ 600) HV 10 (600 ~ 850) HV 10 (300 ~ 600) HV 30 (600 ~ 850) HV 30	6.0 HV 0.2 18 HV 0.2 27 HV 0.2 5.0 HV 0.3 14 HV 0.3 26 HV 0.5 6.0 HV 0.5 15 HV 0.5 20 HV 1 3.0 HV 10 8.0 HV 10 11 HV 10 8.0 HV 30 11 HV 30	Vickers Hardness CRM/ SICT-CP-21004
Durometer hardness testers	21005	(0 ~ 100) HDA (0 ~ 100) HDAM (0 ~ 100) HDAO (0 ~ 100) HDB (0 ~ 100) HDC (0 ~ 100) HDC2 (0 ~ 100) HDCS (0 ~ 100) HDD (0 ~ 100) HDDO (0 ~ 100) HDE (0 ~ 100) HDE2 (0 ~ 100) HDF (0 ~ 100) HDF0 (0 ~ 100) HDM (0 ~ 100) HDO (0 ~ 100) HD00 (0 ~ 100) HD000 (0 ~ 100) HD000-S	0.4 HDA 0.8 HDAM 0.4 HDAO 0.4 HDB 0.3 HDC 0.6 HDC2 0.3 HDCS 0.3 HDD 0.3 HDDO 0.4 HDE 0.6 HDE2 0.6 HDF 0.6 HDF0 0.8 HDM 0.3 HDO 0.4 HD00 0.4 HD000 0.3 HD000-S	Durometer Calibration device/ SICT-CP-21005
Leeb hardness testers	21006	(400 ~ 1 000) HLD	4.6 HLD	Leeb Hardness CRM/ SICT-CP-21006

211. Impact

측정량/장비	분류번호	교정 범위	측정불확도 (신뢰수준 약 95 %)	사용표준/측정방법 등
Charpy impact testers	21102	(0.5 ~ 900) J	-	Laser Distance Meter, Electronic Force Measuring Device/ SICT-CP-21102
Izod impact testers	21103	(0.5 ~ 900) J	-	Laser Distance Meter, Electronic Force Measuring Device/ SICT-CP-21103

301. Time/frequency

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency standards Time Base Frequency	30102	(0.1 ~ 10) MHz	1.0×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30102
General frequency sources Time Base Frequency	30103	10 kHz ~ 100 MHz	1.0×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30103
Frequency meters/counters Time Base Frequency Input Frequency	30104	(1 ~ 10) MHz 0.01 Hz ~ 1 Hz 1 Hz ~ 60 GHz	1.0×10^{-12} 64 pHz 6.4×10^{-11}	GPS Receiver, Universal Counter/ SICT-CP-30104
Time interval sources Period Time interval	30105	1 ns ~ 10 s (1 ~ 100) ns 100 ns ~ 1 ms 1 ms ~ 10 s	6.1×10^{-9} 0.15 ns 1.3 ns 2.1 ns	GPS Receiver, Universal Counter/ SICT-CP-30105
Time interval meters/stop watches/timers Trigger Voltage Period Reference Frequency Relative Time Difference Time rate Timer Count	30106	(-5 ~ 5) V (5 ~ 100) ns (1 ~ 10) MHz day month (-9.95 ~ 9.95) s / day (-300 ~ 300) s / month (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s ≥ 1	1.2×10^{-4} 6.2×10^{-5} ns 6.2×10^{-11} 1.1×10^{-7} 3.6×10^{-7} 6.1 ms 6.2 ms 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5} 0.58	Stop Watch Calibrator/ SICT-CP-30106

302. Velocity & revolution

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard RPM generators Revolution Velocity Measurement	30201	(1 ~ 10 000) min ⁻¹	0.004 0 min ⁻¹	GPS Receiver, Synthesizer Function Generator/ SICT-CP-30201
Revolution Velocity Measurement (Centrifuge)		(30 ~ 5 000) min ⁻¹ (5 000 ~ 8 500) min ⁻¹ (8 500 ~ 50 000) min ⁻¹ (50 000 ~ 80 000) min ⁻¹ (80 000 ~ 99 000) min ⁻¹	0.059 min ⁻¹ 0.099 min ⁻¹ 0.59 min ⁻¹ 0.93 min ⁻¹ 1.1 min ⁻¹	
Contact type tachometers Revolution Velocity Measurement	30202	(1 ~ 10) min ⁻¹ (10 ~ 1 000) min ⁻¹ (1 000 ~ 5 000) min ⁻¹	0.10 min ⁻¹ 0.016 min ⁻¹ 0.063 min ⁻¹	GPS Receiver, Tachometer Cal System/ SICT-CP-30202
Photo tachometers/stroboscopes Revolution Velocity Measurement (Photo-tachometer)	30203	(1 ~ 999.99) min ⁻¹ (1 000.0 ~ 99 999.9) min ⁻¹ (100 000 ~ 600 000) min ⁻¹	0.006 1 min ⁻¹ 0.061 min ⁻¹ 0.61 min ⁻¹	GPS Receiver, Photo Signal Detector/ SICT-CP-30203
Revolution Velocity Measurement (Stroboscope)		(30 ~ 9 000) min ⁻¹ (9 000 ~ 90 000) min ⁻¹ (90 000 ~ 500 000) min ⁻¹	0.005 8 min ⁻¹ 0.058 min ⁻¹ 0.58 min ⁻¹	
Speed meters Speed Test	30204	(0 ~ 400) km/h	6.1×10^{-3} km/h	GPS Receiver, Synthesizer Function Generator/ SICT-CP-30204
Wow-flutter generators Carrier Frequency	30205	10 Hz ~ 99.99 kHz	6.2×10^{-6}	GPS Receiver, Universal Counter/ SICT-CP-30205
Function Frequency		1 Hz ~ 10 kHz (10 ~ 30) kHz	6.2×10^{-6} 2.1×10^{-6}	
Wow/Flutter Deviation		(1 Hz ~ 100 Hz) (0 ~ 3) %	0.025 %	
Output Level		(1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 6) V	5.8×10^{-4} 1.7×10^{-4} 1.3×10^{-4}	
CCIR Pulse		10 ms 30 ms 60 ms 100 ms	1.0×10^{-2} ms 3.0×10^{-2} ms 6.0×10^{-2} ms 1.0×10^{-1} ms	
Wow-flutter meters Wow/Flutter Deviation	30206	(0.1 ~ 0.3) % (0.3 ~ 3) %	0.019 % 0.020 %	GPS Receiver, Wow Flutter Calibrator/ SICT-CP-30206
Carrier Frequency		3 kHz 3.15 kHz	6.2×10^{-5} kHz 6.2×10^{-5} kHz	
CCIR Pulse		(10 ~ 100) ms	0.59 %	
Output Voltage		(1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	6.8 μ V 9.8 μ V 76 μ V	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters	40101	(±) 0 pA (0 ~ 1) pA (1 ~ 10) pA (10 ~ 100) pA (0.1 ~ 100) nA (0.1 ~ 1) µA (1 ~ 10) µA (10 ~ 100) µA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (100 ~ 200) mA (0.2 ~ 1) A (1 ~ 10) A (10 ~ 100) A	13 fA 2.4×10^{-2} 9.4×10^{-3} 2.3×10^{-3} 8.5×10^{-4} 9.3×10^{-4} 4.5×10^{-5} 3.0×10^{-5} 2.5×10^{-5} 2.7×10^{-5} 2.0×10^{-5} 1.7×10^{-5} 3.0×10^{-5} 1.4×10^{-4} 1.5×10^{-4}	Calibrator/ SICT-CP-40101
Transconductance amplifiers	40102	(±) 10 µA ~ 10 A (10 ~ 50) A (50 ~ 100) A	1.3×10^{-5} 4.3×10^{-5} 4.4×10^{-5}	AC-DC Active Current Shunt/ SICT-CP-40102
		AC Current (10 µA) 10 Hz ~ 10 kHz (10 ~ 100) µA 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 µA ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	2.6×10^{-3} 3.6×10^{-4} 6.4×10^{-4} 9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5} 7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5} 8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5} 7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5} 8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Transconductance amplifiers AC Current	40102	(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5}	AC-DC Active Current Shunt/ SICT-CP-40102
		(10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5}	
		(20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4}	
		(50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-5}	
		(100 ~ 120) A 60 Hz	4.6×10^{-4}	
		(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	$0.29 \mu\text{V}$ 4.8×10^{-4} 2.4×10^{-4} 9.6×10^{-5} 4.9×10^{-5} 2.5×10^{-5} 1.1×10^{-5} 6.9×10^{-6} 3.7×10^{-6} 3.3×10^{-6} 5.4×10^{-6} 5.7×10^{-6}	Reference Multimeter/ SICT-CP-40103
		(±) 0 pA (0 ~ 1) pA (1 ~ 10) pA (10 ~ 100) pA (0.1 ~ 100) nA 100 nA ~ 10 A (10 ~ 100) A	9.6 fA 2.1×10^{-2} 6.8×10^{-3} 2.2×10^{-3} 8.5×10^{-4} 1.4×10^{-5} 4.5×10^{-5}	
		(-9.835 ~ 0.000) mV 0.000 mV (0.000 ~ 13.421) mV (13.421 ~ 37.006) mV (37.006 ~ 61.017) mV (61.017 ~ 76.373) mV	$0.42 \mu\text{V}$ $0.24 \mu\text{V}$ $0.42 \mu\text{V}$ $0.48 \mu\text{V}$ $0.53 \mu\text{V}$ $0.57 \mu\text{V}$	
Electrical temperature calibrators TEMPERATURE(SOURCE) T/C	40104			디지털 멀티미터/ SICT-CP-40104

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Electrical temperature calibrators	40104				
RTD		0.999 Ω (0.999 ~ 2.499) Ω (2.499 ~ 4.322) Ω (4.322 ~ 100.000) Ω (100.000 ~ 177.155) Ω (177.155 ~ 313.708) Ω (313.708 ~ 627.422) Ω (627.422 ~ 3 233.3) Ω (±) 0 mV (1 ~ 2) mV (2 ~ 3) mV (3 ~ 4) mV (4 ~ 5) mV (5 ~ 10) mV (10 ~ 50) mV (50 ~ 100) mV (0.1 ~ 0.2) V (0.2 ~ 0.3) V (0.3 ~ 0.6) V (0.6 ~ 1) V (1 ~ 6) V (6 ~ 10) V (10 ~ 70) V (70 ~ 100) V DC Current(SOURCE) (±) 0 mA (0 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 7) mA (7 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 100) mA Resistance(SOURCE)	0.063 mΩ 3.0×10^{-5} 1.9×10^{-5} 9.9×10^{-6} 8.8×10^{-6} 1.1×10^{-5} 9.2×10^{-6} 1.1×10^{-5} 0.24 μV 4.0×10^{-4} 2.0×10^{-4} 1.3×10^{-4} 1.0×10^{-4} 8.2×10^{-5} 4.2×10^{-5} 6.3×10^{-6} 6.2×10^{-5} 3.1×10^{-5} 3.1×10^{-5} 9.5×10^{-6} 3.1×10^{-5} 9.3×10^{-6} 6.2×10^{-5} 9.2×10^{-6} 0.064 μA 9.2×10^{-5} 6.2×10^{-5} 3.5×10^{-5} 2.3×10^{-5} 1.9×10^{-5} 3.3×10^{-5} 8.2×10^{-5} 7.0×10^{-5} 6.3×10^{-5} 0 Ω (0 ~ 0.6) Ω (0.6 ~ 1) Ω (1 ~ 10) Ω (10 ~ 20) Ω (20 ~ 30) Ω (30 ~ 50) Ω (50 ~ 70) Ω (70 ~ 100) Ω (0.1 ~ 0.2) kΩ (0.2 ~ 0.3) kΩ (0.3 ~ 0.5) kΩ (0.5 ~ 0.8) kΩ (0.8 ~ 1) kΩ	$0.063 \text{ m}\Omega$ 3.0×10^{-5} 1.9×10^{-5} 9.9×10^{-6} 8.8×10^{-6} 1.1×10^{-5} 9.2×10^{-6} 1.1×10^{-5} 0.24 μV 4.0×10^{-4} 2.0×10^{-4} 1.3×10^{-4} 1.0×10^{-4} 8.2×10^{-5} 4.2×10^{-5} 6.3×10^{-6} 6.2×10^{-5} 3.1×10^{-5} 3.1×10^{-5} 9.5×10^{-6} 3.1×10^{-5} 9.3×10^{-6} 6.2×10^{-5} 9.2×10^{-6} 0.064 μA 9.2×10^{-5} 6.2×10^{-5} 3.5×10^{-5} 2.3×10^{-5} 1.9×10^{-5} 3.3×10^{-5} 8.2×10^{-5} 7.0×10^{-5} 6.3×10^{-5} 0.061 mΩ 6.1×10^{-4} 8.9×10^{-5} 6.7×10^{-5} 3.3×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.2×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.1×10^{-5}	디지털 멀티미터/ SICT-CP-40104

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators				
Resistance(SOURCE)	40104	(1 ~ 2) kΩ	3.2×10^{-5}	디지털 멀티미터/ SICT-CP-40104
		(2 ~ 3) kΩ	2.3×10^{-5}	
		(3 ~ 5) kΩ	1.8×10^{-5}	
		(5 ~ 8) kΩ	1.4×10^{-5}	
		(8 ~ 10) kΩ	1.1×10^{-5}	
		(10 ~ 20) kΩ	3.2×10^{-5}	
		(20 ~ 30) kΩ	2.4×10^{-5}	
		(30 ~ 40) kΩ	1.9×10^{-5}	
		(40 ~ 50) kΩ	1.6×10^{-5}	
		(50 ~ 100) kΩ	1.1×10^{-5}	
TEMPERATURE(MEASURE)				
T/C		(-9.835 ~ 0.000) mV	0.59 μV	
		0.000 mV	0.50 μV	
		(0.000 ~ 13.422) mV	0.62 μV	
		(13.422 ~ 28.947) mV	0.75 μV	
		(28.947 ~ 45.094) mV	0.88 μV	
		(45.094 ~ 53.113) mV	0.95 μV	
		(53.113 ~ 76.374) mV	1.1 μV	
RTD		0.998 Ω	0.24 mΩ	
		(0.998 ~ 2.496) Ω	1.0×10^{-4}	
		(2.496 ~ 4.315) Ω	7.1×10^{-5}	
		(4.315 ~ 16.994) Ω	3.9×10^{-5}	
		(16.994 ~ 249.580) Ω	3.5×10^{-5}	
		(249.580 ~ 317.988) Ω	4.3×10^{-5}	
		(317.988 ~ 390.474) Ω	4.0×10^{-5}	
		(390.474 ~ 3 233.2) Ω	3.5×10^{-5}	
DC Voltage(MEASURE)		(±)		
		0 mV	0.50 μV	
		(1 ~ 5) mV	5.2×10^{-4}	
		(5 ~ 10) mV	9.3×10^{-5}	
		(10 ~ 100) mV	5.9×10^{-5}	
		(0.1 ~ 0.5) V	6.3×10^{-5}	
		(0.5 ~ 0.8) V	1.3×10^{-5}	
		(0.8 ~ 1) V	1.6×10^{-5}	
		(1 ~ 10) V	6.6×10^{-6}	
		(10 ~ 20) V	9.1×10^{-6}	
		(20 ~ 40) V	7.9×10^{-6}	
		(40 ~ 70) V	6.9×10^{-6}	
		(70 ~ 100) V	6.4×10^{-6}	
		(100 ~ 200) V	7.8×10^{-6}	
		(200 ~ 300) V	2.2×10^{-5}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators DC Current(MEASURE)	40104	(±) 0 mA (0 ~ 0.1) mA (0.1 ~ 0.2) mA (0.2 ~ 0.3) mA (0.3 ~ 0.7) mA (0.7 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 50) mA (50 ~ 100) mA (100 ~ 130) mA	0.062 μA 6.4×10^{-4} 3.2×10^{-4} 2.2×10^{-4} 1.7×10^{-4} 9.3×10^{-5} 9.9×10^{-5} 7.6×10^{-5} 5.8×10^{-5} 9.9×10^{-5} 8.2×10^{-5} 7.4×10^{-5} 7.0×10^{-5} 6.7×10^{-5} 8.7×10^{-5}	디지털 멀티미터/ SICT-CP-40104
Resistance(MEASURE)		0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 20) kΩ (20 ~ 30) kΩ (30 ~ 40) kΩ (40 ~ 50) kΩ (50 ~ 100) kΩ	0.098 mΩ 6.4×10^{-5} 1.1×10^{-5} 9.6×10^{-6} 6.5×10^{-5} 6.1×10^{-5} 4.7×10^{-5} 4.0×10^{-5} 4.2×10^{-5} 3.9×10^{-5} 3.4×10^{-5}	
DC current shunts Resistance	40105	1 μΩ (0.001 ~ 0.01) mΩ (0.01 ~ 0.2) mΩ (0.2 ~ 1) mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ	0.32 nΩ 2.8×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 1.1×10^{-6} 2.8×10^{-6} 8.1×10^{-7} 1.3×10^{-6} 6.7×10^{-7} 6.2×10^{-7} 7.9×10^{-7} 2.0×10^{-6} 1.4×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	Trans Conductance Amplifier/ SICT-CP-40105
Galvanometers/null detectors DC Voltage	40106	(±) (100 ~ 300) μV (0.3 ~ 1) mV 1 mV ~ 1 000 V	1.4×10^{-2} 1.2×10^{-2} 6.8×10^{-3}	Calibrator/ SICT-CP-40106
Potentiometers DC Voltage	40107	(100 ~ 300) μV (0.3 ~ 1) mV (1 ~ 3) mV 3 mV ~ 1 000 V	5.7×10^{-3} 2.2×10^{-3} 6.0×10^{-4} 3.0×10^{-4}	Calibrator/ SICT-CP-40107

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC power supplies	40108	(±)		DC Electronics Load/ SICT-CP-40108
		0 mV	5.8 μ V	
		(0 ~ 10) mV	5.8 $\times 10^{-4}$	
		(10 ~ 100) mV	5.8 $\times 10^{-5}$	
		(0.1 ~ 1) V	7.5 $\times 10^{-6}$	
		(1 ~ 10) V	3.3 $\times 10^{-6}$	
		(10 ~ 100) V	7.7 $\times 10^{-6}$	
		(100 ~ 600) V	1.3 $\times 10^{-5}$	
		(600 ~ 1 000) V	6.6 $\times 10^{-5}$	
DC Current	40108	(1 ~ 10) mA	5.8 $\times 10^{-3}$	
		(10 ~ 100) mA	5.9 $\times 10^{-4}$	
		(0.1 ~ 1) A	2.4 $\times 10^{-4}$	
		(1 ~ 10) A	3.1 $\times 10^{-4}$	
		(10 ~ 300) A	2.4 $\times 10^{-4}$	
		(300 ~ 500) A	2.6 $\times 10^{-4}$	
		(500 ~ 1 000) A	4.7 $\times 10^{-5}$	
		(1 000 ~ 3 000) A	5.1 $\times 10^{-4}$	
Load regulation	40108	(0 ~ 2) mV	0.16 mV	
		(2 ~ 20) mV	7.8 $\times 10^{-2}$	
		(20 ~ 200) mV	8.2 $\times 10^{-3}$	
Ripple	40108	(0.1 ~ 0.4) mV	3.8 $\times 10^{-1}$	
		(0.4 ~ 0.6) mV	1.1 $\times 10^{-1}$	
		(0.6 ~ 1) mV	7.3 $\times 10^{-2}$	
		(1 ~ 10) mV	4.4 $\times 10^{-2}$	
		(10 ~ 50) mV	7.1 $\times 10^{-2}$	
DC voltage dividers	40110			Calibrator/ SICT-CP-40110
		(±)		
		(0.01 ~ 1 000) V	4.5 $\times 10^{-6}$	
		(1 ~ 50) kV	8.8 $\times 10^{-5}$	
DC voltage standards	40111	(50 ~ 100) kV	8.4 $\times 10^{-5}$	Null Detector/ SICT-CP-40111
		1 V	1.6 μ V	
		1.018 V	0.8 μ V	
DC voltmeters	40112	10 V	3.1 μ V	Calibrator/ SICT-CP-40112
		(±)		
		0 mV	0.51 μ V	
		(0 ~ 1) mV	5.0 $\times 10^{-4}$	
		(1 ~ 2) mV	2.5 $\times 10^{-4}$	
		(2 ~ 5) mV	1.0 $\times 10^{-4}$	
		(5 ~ 10) mV	6.2 $\times 10^{-5}$	
		(10 ~ 20) mV	5.0 $\times 10^{-5}$	
		(20 ~ 50) mV	2.5 $\times 10^{-5}$	
		(50 ~ 80) mV	1.0 $\times 10^{-5}$	
		(80 ~ 100) mV	6.2 $\times 10^{-6}$	
Static/ionic voltmeters	40113	(0.1 ~ 1 000) V	8.0 $\times 10^{-6}$	DC Power Supply/ SICT-CP-40113
DC Voltage	40113	(±)		
		0 V	68 mV	
		0 V ~ 50 kV	1.3 $\times 10^{-2}$	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators	40201			Standard Capacitance Set / SICT-CP-40201
Frequency		50 Hz ~ 100 MHz	6.8×10^{-8}	
Capacitance		(1 pF)		
		50 Hz ~ 1 kHz	3.6×10^{-4}	
		1 kHz ~ 1 MHz	3.7×10^{-4}	
		2 MHz	4.2×10^{-4}	
		3 MHz	5.4×10^{-4}	
		4 MHz	7.2×10^{-4}	
		5 MHz	9.5×10^{-4}	
		10 MHz	2.5×10^{-3}	
		13 MHz	3.7×10^{-3}	
		(10 pF)		
		50 Hz ~ 5 MHz	3.6×10^{-4}	
		10 MHz	3.7×10^{-4}	
		13 MHz	3.9×10^{-4}	
		(100 pF)		
		(50 ~ 120) Hz	3.5×10^{-4}	
		120 Hz ~ 4 MHz	3.6×10^{-4}	
		5 MHz	3.8×10^{-4}	
		10 MHz	4.9×10^{-4}	
		13 MHz	6.1×10^{-4}	
		(1 000 pF)		
		50 Hz ~ 1 MHz	3.6×10^{-4}	
		2 MHz	3.8×10^{-4}	
		3 MHz	4.5×10^{-4}	
		4 MHz	5.7×10^{-4}	
		5 MHz	7.2×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	3.0×10^{-3}	
		(10 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.2×10^{-5}	
		(100 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.2×10^{-5}	
		(1 μF)		
		(50 ~ 100) Hz	7.0×10^{-4}	
		100 Hz ~ 10 kHz	8.2×10^{-5}	
		(10 ~ 100) kHz	1.1×10^{-4}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators Capacitance	40201	(10 μ F) 120 Hz ~ 1 kHz	1.3×10^{-3}	Standard Capacitance Set / SICT-CP-40201
		(100 μ F) 120 Hz	1.4×10^{-3}	
		(1 mF) 120 Hz	1.5×10^{-3}	
		(3 mF) 120 Hz	1.5×10^{-3}	
		(10 mF) 120 Hz	1.5×10^{-3}	
		(30 mF) 120 Hz	2.9×10^{-3}	
		(50 ~ 60) Hz		
		100 pF	7.7×10^{-5}	
		1 000 pF	3.9×10^{-5}	
		$\tan \delta$ 0.000 0 ~ 0.001 0	1.7×10^{-4}	
Decade capacitors Capacitance	40202	0.001 0 ~ 0.003 0	1.8×10^{-4}	Standard Capacitance Set / SICT-CP-40202
		0.003 0 ~ 0.005 0	1.9×10^{-4}	
		0.005 0 ~ 0.008 0	2.1×10^{-4}	
		0.008 0 ~ 0.010 0	2.2×10^{-4}	
		0.010 0 ~ 0.030 0	3.3×10^{-4}	
		0.030 0 ~ 0.050 0	4.4×10^{-4}	
		0.050 0 ~ 0.080 0	6.2×10^{-4}	
		0.080 0 ~ 0.100 0	7.4×10^{-4}	
		(50 Hz ~ 20 kHz)		
		1 pF	5.7×10^{-5}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard capacitors	40204	(50 Hz ~ 20 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F (1 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F (1 pF) 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 10) pF 1 kHz ~ 3 MHz (3 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (10 ~ 100) pF 1 kHz ~ 1 MHz (1 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (100 pF ~ 1 nF) 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz	5.2×10^{-5} 4.0×10^{-5} 3.0×10^{-5} 4.0×10^{-5} 2.9×10^{-4} 5.1×10^{-4} 9.1×10^{-6} 7.1×10^{-6} 6.1×10^{-6} 7.1×10^{-6} 5.0×10^{-5} 9.0×10^{-5} 2.4×10^{-4} 2.5×10^{-4} 3.3×10^{-4} 4.7×10^{-4} 6.7×10^{-4} 9.1×10^{-4} 2.5×10^{-3} 3.7×10^{-3} 2.3×10^{-4} 2.4×10^{-4} 2.6×10^{-4} 2.8×10^{-4} 2.3×10^{-4} 2.4×10^{-4} 2.5×10^{-4} 2.7×10^{-4} 4.0×10^{-4} 5.4×10^{-4} 2.3×10^{-4} 2.4×10^{-4} 2.8×10^{-4} 3.6×10^{-4} 5.0×10^{-4} 6.6×10^{-4} 1.9×10^{-3} 2.8×10^{-3}	Standard Capacitance Set / Capacitance Bridge SICT-CP-40204

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard capacitors	40204	(1 ~ 100) nF 120 Hz ~ 100 kHz	2.3×10^{-4}	Standard Capacitance Set / Capacitance Bridge SICT-CP-40204
		(100 nF ~ 1 μ F) 120 Hz	2.4×10^{-4}	
		120 Hz ~ 10 kHz	2.3×10^{-4}	
		(10 ~ 100) kHz	2.4×10^{-4}	
		(1 ~ 10) μ F 120 Hz ~ 1 kHz	1.2×10^{-3}	
		(30 μ F) 120 Hz	1.3×10^{-3}	
		(100 μ F) 120 Hz	1.3×10^{-3}	
		(300 μ F) 120 Hz	1.5×10^{-3}	
		(1 mF) 120 Hz	1.4×10^{-3}	
		(3 mF) 120 Hz	1.5×10^{-3}	
Earth testers	40205	(10 ~ 100) mF 120 Hz	1.4×10^{-3}	Decade Resistor/ SICT-CP-40205
		(30 mF) 120 Hz	1.4×10^{-3}	
		(300 mF) 120 Hz	2.9×10^{-3}	
		1 V	6.4×10^{-3}	
		(1 ~ 10) V	6.4×10^{-4}	
	Resistance	(10 ~ 50) V	1.3×10^{-3}	
		(50 ~ 100) V	6.4×10^{-4}	
		(100 ~ 500) V	1.6×10^{-4}	
	AC Current out	(500 ~ 1 000) V	6.4×10^{-4}	
		1 $\text{m}\Omega$	8.6×10^{-4}	
		(1 ~ 10) $\text{m}\Omega$	7.2×10^{-4}	
	AC Current out	10 $\text{m}\Omega$ ~ 100 k Ω	6.8×10^{-4}	
		1 A	1.2×10^{-3}	
		(1 ~ 3) A	1.5×10^{-3}	
		(3 ~ 20) A	9.7×10^{-4}	
		(20 ~ 30) A	1.0×10^{-3}	
		(30 ~ 60) A	8.4×10^{-4}	
		(60 ~ 100) A	1.0×10^{-3}	
	Timer	(100 ~ 150) A	4.6×10^{-3}	
		(150 ~ 200) A	3.7×10^{-3}	
		1 s	5.8×10^{-6}	
	Timer	(1 ~ 100) s	5.8×10^{-6}	
		(100 ~ 1 000) s	8.2×10^{-6}	
		(1 000 ~ 10 000) s	5.8×10^{-5}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inductance bridges/indicators	40206			Standard Inductor/ SICT-CP-40206
Frequency		50 Hz ~ 100 MHz	7.0×10^{-8}	
Inductance		(1 kHz) 100 µH 1 mH 10 mH 100 mH 1 H 10 H	1.9×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4}	
Inductors	40208			Standard Inductor/ SICT-CP-40208
Standard Inductance		(1 kHz) 100 µH 1 mH 10 mH 100 mH 1 H 10 H	28 nH 0.24 µH 2.4 µH 24 µH 0.24 mH 2.5 mH	
Decade Inductance		(1 kHz) 100 µH ~ 10 H	3.5×10^{-3}	
Insulation testers	40210			High Resistance Decade/ SICT-CP-40210
AC Voltage		1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	8.8×10^{-5} 9.0×10^{-5} 1.0×10^{-4} 1.1×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 5 000) V (5 000 ~ 10 000) V	6.4×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 6.5×10^{-3} 6.1×10^{-3}	
Insulation Resistance		1 kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ 10 TΩ	7.1×10^{-5} 3.7×10^{-5} 2.5×10^{-5} 3.1×10^{-5} 9.5×10^{-5} 2.4×10^{-5} 3.1×10^{-5} 6.1×10^{-5} 1.3×10^{-4} 2.6×10^{-4} 6.3×10^{-4}	
Q-meters	40211			Universal Counter/ SICT-CP-40211
Frequency Test		60 Hz ~ 100 MHz	7.0×10^{-8}	
Quality Factor		0 ~ 1 000	9.7×10^{-4}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance bridges & similar instruments	40213			
Resistance(Rheostat Arm)		1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ	1.4×10^{-6} 7.5×10^{-7} 4.0×10^{-6} 4.7×10^{-7} 5.1×10^{-7} 4.4×10^{-7} 2.4×10^{-7} 5.1×10^{-7} 1.1×10^{-6} 1.3×10^{-6} 5.3×10^{-6} 1.1×10^{-5}	Standard Resistance Set/ SICT-CP-40213
Resistance(Ratio Arm)		1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ	1.4×10^{-6} 7.5×10^{-7} 4.0×10^{-6} 4.7×10^{-7} 5.1×10^{-7} 4.4×10^{-7} 2.4×10^{-7} 5.1×10^{-7} 1.1×10^{-6} 1.3×10^{-6} 5.3×10^{-6} 1.1×10^{-5}	
Resistance Ratio		0.1 ~ 1.3	33×10^{-9}	
Resistance meters	40214			
DC Resistance		1 μΩ 5 μΩ 10 μΩ 25 μΩ 100 μΩ 1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ 10 TΩ	8.7×10^{-4} 5.8×10^{-4} 4.8×10^{-4} 5.2×10^{-4} 3.0×10^{-4} 1.2×10^{-6} 7.5×10^{-7} 4.0×10^{-6} 4.7×10^{-7} 5.1×10^{-7} 4.4×10^{-7} 2.4×10^{-7} 5.1×10^{-7} 1.1×10^{-6} 1.3×10^{-6} 5.3×10^{-6} 1.1×10^{-5} 2.5×10^{-4} 7.1×10^{-4} 9.4×10^{-4} 1.5×10^{-3} 7.1×10^{-4}	Standard Resistance Set/ SICT-CP-40214

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance meters	40214			Standard Resistance Set/ SICT-CP-40214
Frequency		1 kHz	6.8×10^{-8}	
AC Voltage		10 mV	2.4×10^{-4}	
		(10 ~ 100) mV	8.0×10^{-5}	
		(0.1 ~ 10) V	8.2×10^{-5}	
AC Resistance		(1 kHz)		
		1 mΩ	5.1×10^{-3}	
		(1 ~ 10) mΩ	5.2×10^{-4}	
		(10 ~ 100) mΩ	3.3×10^{-4}	
		(0.1 ~ 1) Ω	1.5×10^{-4}	
		1 Ω ~ 10 kΩ	1.3×10^{-4}	
		(10 ~ 100) kΩ	1.6×10^{-4}	
		(0.1 ~ 1) MΩ	3.1×10^{-4}	
		(1 ~ 10) MΩ	3.0×10^{-3}	
Resistors	40215			Standard Resistance Set/ SICT-CP-40215
DC Resistance		1 mΩ	1.6×10^{-6}	
		(1 ~ 10) mΩ	1.1×10^{-6}	
		(10 ~ 100) mΩ	2.8×10^{-6}	
		(0.1 ~ 1) Ω	8.1×10^{-7}	
		(1 ~ 10) Ω	1.3×10^{-6}	
		(10 ~ 100) Ω	6.7×10^{-7}	
		(0.1 ~ 1) kΩ	6.2×10^{-7}	
		(1 ~ 10) kΩ	7.9×10^{-7}	
		(10 ~ 100) kΩ	2.0×10^{-6}	
		(0.1 ~ 1) MΩ	1.4×10^{-6}	
		(1 ~ 10) MΩ	5.2×10^{-6}	
		(10 ~ 100) MΩ	9.7×10^{-6}	
		(0.1 ~ 1) GΩ	2.3×10^{-4}	
		(1 ~ 10) GΩ	6.9×10^{-4}	
		(10 ~ 100) GΩ	9.3×10^{-4}	
		(0.1 ~ 1) TΩ	1.4×10^{-3}	
		(1 ~ 10) TΩ	4.1×10^{-3}	
		(10 ~ 100) TΩ	7.6×10^{-3}	
AC Resistance		(50 Hz ~ 1 kHz)		
		1 mΩ	1.0×10^{-3}	
		(1 ~ 10) mΩ	5.9×10^{-4}	
		(10 ~ 100) mΩ	3.9×10^{-4}	
		(0.1 ~ 1) Ω	2.4×10^{-4}	
		(1 ~ 100) Ω	2.5×10^{-4}	
		(10 Ω)		
		1 kHz	2.4×10^{-4}	
		1 MHz	4.0×10^{-4}	
		2 MHz	5.6×10^{-4}	
		3 MHz	6.5×10^{-4}	
		4 MHz	7.5×10^{-4}	
		5 MHz	1.0×10^{-3}	
		10 MHz	4.0×10^{-3}	
		13 MHz	6.0×10^{-3}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors	40215	(10 ~ 100) Ω		Standard Resistance Set/ SICT-CP-40215
		1 kHz	2.4×10^{-4}	
		1 MHz	4.0×10^{-4}	
		2 MHz	4.8×10^{-4}	
		3 MHz	5.6×10^{-4}	
		4 MHz	5.6×10^{-4}	
		5 MHz	5.6×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	3.0×10^{-3}	
		(100 Ω ~ 1 k Ω)		
		1 kHz	2.4×10^{-4}	
		100 kHz	4.0×10^{-4}	
		1 MHz	4.0×10^{-4}	
		2 MHz	4.0×10^{-4}	
		3 MHz	4.0×10^{-4}	
		4 MHz	4.8×10^{-4}	
		5 MHz	5.6×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	3.0×10^{-3}	
Decade Resistance	40215	(1 ~ 10) k Ω		
		1 kHz	2.4×10^{-4}	
		100 kHz	3.3×10^{-4}	
		1 MHz	4.0×10^{-4}	
		(10 ~ 100) k Ω		
		1 kHz	2.4×10^{-4}	
		100 kHz	4.0×10^{-4}	
		1 MHz	4.0×10^{-4}	
		(100 k Ω ~ 1 M Ω)		
		1 kHz	3.8×10^{-4}	
		(1 ~ 10) M Ω		
		1 kHz	3.0×10^{-3}	
		0 Ω	$64 \mu\Omega$	
		(0 ~ 10) m Ω	$6.5 \mu\Omega$	
		(10 ~ 100) m Ω	$64 \mu\Omega$	
		(0.1 ~ 1) Ω	$66 \mu\Omega$	
		(1 ~ 7) Ω	3.9×10^{-5}	
		(7 ~ 10) Ω	1.3×10^{-5}	
		(10 ~ 70) Ω	2.0×10^{-5}	
		(70 ~ 100) Ω	9.8×10^{-6}	
		(100 ~ 700) Ω	1.9×10^{-5}	
		(0.7 ~ 1) k Ω	9.6×10^{-6}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors	40215	(1 ~ 7) kΩ (7 ~ 10) kΩ (10 ~ 70) kΩ (70 ~ 100) kΩ (100 ~ 600) kΩ (0.6 ~ 1) MΩ (1 ~ 7) MΩ (7 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ (1 ~ 10) TΩ	3.5×10^{-5} 1.2×10^{-5} 2.0×10^{-5} 9.8×10^{-6} 2.9×10^{-5} 1.2×10^{-5} 8.0×10^{-5} 2.7×10^{-5} 2.0×10^{-4} 2.5×10^{-4} 7.0×10^{-4} 1.0×10^{-3} 1.5×10^{-3} 4.3×10^{-3}	Standard Resistance Set/ SICT-CP-40215
Impedance bridges/LCR meters	40217	50 Hz ~ 100 MHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 10) V (10 ~ 100) V Capacitance (1 pF) 50 Hz ~ 1 kHz 1 kHz ~ 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (10 pF) 50 Hz ~ 5 MHz 10 MHz 13 MHz (100 pF) 50 Hz ~ 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz	7.0×10^{-8} 2.1×10^{-3} 4.4×10^{-4} 8.8×10^{-5} 8.2×10^{-5} 8.9×10^{-5} 3.5×10^{-4} 3.6×10^{-4} 4.2×10^{-4} 5.4×10^{-4} 7.2×10^{-4} 9.4×10^{-4} 2.5×10^{-3} 3.6×10^{-3} 3.5×10^{-4} 3.7×10^{-4} 3.8×10^{-4} 3.5×10^{-4} 3.6×10^{-4} 3.7×10^{-4} 4.8×10^{-4} 6.0×10^{-4}	Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217			Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
Capacitance		(1 000 pF)		
		50 Hz ~ 1 MHz	3.5×10^{-4}	
		2 MHz	3.8×10^{-4}	
		3 MHz	4.5×10^{-4}	
		4 MHz	5.6×10^{-4}	
		5 MHz	7.2×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	2.9×10^{-3}	
		(10 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.1×10^{-5}	
		(100 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.1×10^{-5}	
		(1 μ F)		
		(50 ~ 100) Hz	5.1×10^{-4}	
		100 Hz ~ 10 kHz	8.1×10^{-5}	
		(10 ~ 100) kHz	1.0×10^{-4}	
		(10 μ F)		
		120 Hz ~ 1 kHz	1.2×10^{-3}	
		(100 μ F)		
		120 Hz	1.3×10^{-3}	
		(1 mF)		
		120 Hz	1.4×10^{-3}	
		(3 mF)		
		120 Hz	1.4×10^{-3}	
		(10 mF)		
		120 Hz	1.4×10^{-3}	
		(30 mF)		
		120 Hz	2.9×10^{-3}	
		(1 pF)		
		1 kHz ~ 1 MHz	0.000 12	
		1 MHz ~ 5 MHz	0.000 23	
		5 MHz ~ 13 MHz	0.000 84	
		(10 pF)		
		1 kHz ~ 13 MHz	0.000 15	
		(100 pF)		
		1 kHz ~ 5 MHz	0.000 13	
		5 MHz ~ 13 MHz	0.000 27	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217			Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
Dissipation Factor		(1 pF) 1 kHz ~ 1 MHz (1 ~ 5) MHz (5 ~ 13) MHz	0.000 12 0.000 23 0.000 84	
		(10 pF) 1 kHz ~ 13 MHz	0.000 15	
		(100 pF) 1 kHz ~ 5 MHz (5 ~ 13) MHz	0.000 13 0.000 27	
		(1 000 pF) 1 kHz ~ 1 MHz (1 ~ 5) MHz (5 ~ 13) MHz	0.000 12 0.000 24 0.000 86	
		(10 nF) 120 Hz ~ 100 kHz	0.000 46	
		(100 nF) 120 Hz ~ 100 kHz	0.000 58	
		(1 µF) 120 Hz ~ 100 kHz	0.000 81	
Resistance		(1 mΩ) 50 Hz 50 Hz ~ 1 kHz	6.0 × 10 ⁻³ 5.0 × 10 ⁻³	
		(10 mΩ) 50 Hz 50 Hz ~ 1 kHz	1.0 × 10 ⁻³ 5.2 × 10 ⁻⁴	
		(100 mΩ) 50 Hz 50 Hz ~ 1 kHz	7.1 × 10 ⁻⁴ 3.3 × 10 ⁻⁴	
		(1 Ω) 50 Hz 50 Hz ~ 1 kHz	6.8 × 10 ⁻⁴ 1.3 × 10 ⁻⁴	
		(10 Ω) 50 Hz 50 Hz ~ 1 kHz	9.1 × 10 ⁻⁴ 1.3 × 10 ⁻⁴	
		1 kHz ~ 1 MHz	3.3 × 10 ⁻⁴	
		(1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz	5.2 × 10 ⁻⁴ 6.1 × 10 ⁻⁴ 7.1 × 10 ⁻⁴ 1.0 × 10 ⁻³ 4.0 × 10 ⁻³ 6.0 × 10 ⁻³	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217			Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
Resistance		(100 Ω)		
		50 Hz	6.2×10^{-4}	
		50 Hz ~ 1 kHz	1.3×10^{-4}	
		1 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 ~ 2) MHz	4.2×10^{-4}	
		(2 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(1 k Ω)		
		1 kHz	1.3×10^{-4}	
		1 kHz ~ 3 MHz	3.3×10^{-4}	
		(3 ~ 4) MHz	4.2×10^{-4}	
		(4 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(10 k Ω)		
		1 kHz	1.3×10^{-4}	
		(1 ~ 100) kHz	2.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(100 k Ω)		
		1 kHz	1.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 M Ω)		
		1 kHz	3.0×10^{-4}	
		(10 M Ω)		
		1 kHz	2.9×10^{-3}	
Inductance		(1 kHz)		
		100 μ H	1.9×10^{-4}	
		1 mH	1.3×10^{-4}	
		10 mH	1.3×10^{-4}	
		100 mH	1.3×10^{-4}	
		1 H	1.3×10^{-4}	
		10 H	1.3×10^{-4}	
DC Bias		(\pm)		
		0 μ V	0.68 μ V	
		0 μ V ~ 100 mV	1.1×10^{-5}	
		(0.1 ~ 1) V	7.5×10^{-6}	
		(1 ~ 10) V	7.2×10^{-6}	
		(10 ~ 100) V	8.2×10^{-6}	
DC Current		0 μ A	5.8 μ A	
		0 μ A ~ 200 mA	4.1×10^{-5}	
		(0.2 ~ 2) A	3.6×10^{-5}	
		(2 ~ 20) A	1.9×10^{-4}	
		(20 ~ 100) A	1.4×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC ammeters	40301	(100 μ A)		Power Calibrator, Calibrator/ SICT-CP-40301
		40 Hz ~ 1 kHz	2.2×10^{-4}	
		(1 ~ 5) kHz	4.9×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(100 μ A ~ 1 mA)		
		40 Hz ~ 1 kHz	1.8×10^{-4}	
		(1 ~ 5) kHz	3.8×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(1 ~ 10) mA		
		40 Hz ~ 1 kHz	1.8×10^{-4}	
		(1 ~ 5) kHz	3.4×10^{-4}	
		(5 ~ 10) kHz	1.9×10^{-3}	
		(10 ~ 100) mA		
		40 Hz ~ 1 kHz	1.7×10^{-4}	
		(1 ~ 5) kHz	3.2×10^{-4}	
		(5 ~ 10) kHz	1.5×10^{-3}	
Clamp ammeters/voltmeters	40302	(100 mA ~ 1 A)		Power Calibrator, Calibrator/ SICT-CP-40302
		40 Hz ~ 1 kHz	3.5×10^{-4}	
		(1 ~ 5) kHz	6.7×10^{-4}	
		(5 ~ 10) kHz	8.3×10^{-3}	
		(1 ~ 10) A		
		(40 ~ 100) Hz	2.1×10^{-4}	
		100 Hz ~ 5 kHz	5.8×10^{-4}	
		(10 ~ 20) A		
		(40 ~ 100) Hz	2.0×10^{-4}	
		100 Hz ~ 5 kHz	5.3×10^{-4}	
		(20 ~ 100) A		
		(40 ~ 100) Hz	1.9×10^{-4}	
		100 Hz ~ 5 kHz	5.9×10^{-4}	
		(100 ~ 200) A		
		60 Hz	8.5×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Clamp ammeters/voltmeters	40302	100 mA ~ 10 A		Power Calibrator, Calibrator/ SICT-CP-40302	
		40 Hz ~ 1 kHz	7.0×10^{-4}		
		(1 ~ 5) kHz	9.0×10^{-4}		
		(5 ~ 10) kHz	8.3×10^{-3}		
		(10 ~ 100) A			
		(40 ~ 100) Hz	6.3×10^{-4}		
		100 Hz ~ 5 kHz	8.3×10^{-4}		
		(5 ~ 10) kHz	4.3×10^{-2}		
		(100 ~ 1 000) A			
		(40 ~ 100) Hz	1.3×10^{-3}		
AC Current		100 Hz ~ 1 kHz	2.5×10^{-3}		
		(1 000 ~ 2 500) A			
		(40 ~ 60) Hz	1.3×10^{-3}		
		(2 500 ~ 3 000) A			
		60 Hz	1.3×10^{-3}		
		(3 000 ~ 6 000) A			
		60 Hz	3.9×10^{-4}		
		(6 000 ~ 10 000) A			
		60 Hz	3.3×10^{-4}		
		0 μA	9.2 nA		
DC Current		(0 ~ 100) μA	9.6×10^{-4}		
		100 μA ~ 1 A	6.2×10^{-4}		
		(1 ~ 100) A	6.4×10^{-4}		
		(100 ~ 500) A	2.7×10^{-4}		
		(500 ~ 1 000) A	3.2×10^{-4}		
		(1 000 ~ 2 000) A	5.6×10^{-4}		
		(2 000 ~ 2 500) A	1.7×10^{-3}		
		(1 ~ 10) mV			
AC Voltage		40 Hz ~ 10 kHz	4.8×10^{-3}		
		(10 ~ 50) kHz	4.9×10^{-3}		
		(50 ~ 100) kHz	6.5×10^{-3}		
		(10 ~ 100) mV			
		40 Hz ~ 10 kHz	8.5×10^{-4}		
		(10 ~ 50) kHz	9.6×10^{-4}		
		(50 ~ 100) kHz	1.4×10^{-3}		
		(100 mV ~ 1 V)			
		40 Hz ~ 10 kHz	1.6×10^{-4}		
		(10 ~ 50) kHz	2.4×10^{-4}		
		(50 ~ 100) kHz	5.7×10^{-4}		

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(1 ~ 10) V 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	9.0×10^{-5} 1.1×10^{-4} 1.5×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302
		(10 ~ 200) V 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	1.0×10^{-4} 1.3×10^{-4} 2.2×10^{-4}	
		(200 ~ 1 000) V 40 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz	1.8×10^{-4} 2.5×10^{-4} 2.0×10^{-3}	
		DC VOLTAGE 0 mV (0 ~ 100) mV 100 mV ~ 1 V (1 ~ 1 000) V	6.1 μ V 6.1×10^{-4} 6.3×10^{-5} 6.2×10^{-5}	
		Resistance 0 Ω (0 ~ 100) Ω 100 Ω ~ 100 k Ω 100 k Ω ~ 10 M Ω (10 ~ 100) M Ω	0.61 m Ω 6.2×10^{-4} 6.2×10^{-5} 7.7×10^{-5} 1.3×10^{-4}	
		Frequency 10 Hz ~ 10 MHz	6.5×10^{-5}	
AC voltage/current calibrators	40303	(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.7×10^{-3} 3.0×10^{-3} 1.2×10^{-2}	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		(1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.1×10^{-3} 1.7×10^{-3} 7.7×10^{-3}	
		(2 ~ 5) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3}	
		(5 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage
AC Voltage		(10 ~ 20) mV		Measurement Standard,
		10 Hz	1.8×10^{-4}	Reference Multimeter,
		10 Hz ~ 10 kHz	1.4×10^{-4}	Current Shunt/
		(10 ~ 100) kHz	2.2×10^{-4}	SICT-CP-40303
		100 kHz ~ 1 MHz	2.2×10^{-3}	
		(20 ~ 50) mV		
		10 Hz	1.4×10^{-4}	
		10 Hz ~ 10 kHz	9.2×10^{-5}	
		(10 ~ 100) kHz	1.6×10^{-4}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(50 ~ 100) mV		
		10 Hz	1.1×10^{-4}	
		10 Hz ~ 10 kHz	6.6×10^{-5}	
		(10 ~ 100) kHz	1.2×10^{-4}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(100 ~ 200) mV		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 10 kHz	3.9×10^{-5}	
		(10 ~ 100) kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(200 ~ 500) mV		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 10 kHz	3.6×10^{-5}	
		(10 ~ 100) kHz	7.1×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(500 mV ~ 1 V)		
		10 Hz	7.6×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	6.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(1 ~ 2) V		
		10 Hz	7.1×10^{-5}	
		10 Hz ~ 10 kHz	2.7×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-5}	
		100 kHz ~ 1 MHz	1.0×10^{-3}	
		(2 ~ 5) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.6×10^{-5}	
		10 kHz ~ 100 kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303	(5 ~ 20) V		Alternating Voltage
		10 Hz	7.2×10^{-5}	Measurement Standard,
		10 Hz ~ 10 kHz	2.8×10^{-5}	Reference Multimeter,
		(10 ~ 100) kHz	7.5×10^{-5}	Current Shunt /
		100 kHz ~ 1 MHz	1.4×10^{-3}	SICT-CP-40303
		(20 ~ 50) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	3.0×10^{-5}	
		(10 ~ 100) kHz	8.0×10^{-5}	
		(50 ~ 200) V		
		10 Hz	7.4×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	8.5×10^{-5}	
		(200 ~ 1 000) V		
		10 Hz	7.7×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-4}	
AC Current		(10 µA)		
		10 Hz ~ 10 kHz	2.6×10^{-3}	
		(10 ~ 100) µA		
		10 Hz ~ 1 kHz	3.6×10^{-4}	
		(1 ~ 10) kHz	6.4×10^{-4}	
		(100 µA ~ 1 mA)		
		10 Hz	9.8×10^{-5}	
		10 Hz ~ 1 kHz	7.5×10^{-5}	
		(1 ~ 10) kHz	9.4×10^{-5}	
		(1 ~ 100) mA		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 1 kHz	4.6×10^{-5}	
		(1 ~ 10) kHz	4.2×10^{-5}	
		(100 mA ~ 1 A)		
		10 Hz	8.1×10^{-5}	
		10 Hz ~ 1 kHz	4.9×10^{-5}	
		(1 ~ 10) kHz	4.4×10^{-5}	
		(1 ~ 2) A		
		10 Hz	7.9×10^{-5}	
		10 Hz ~ 1 kHz	4.7×10^{-5}	
		(1 ~ 10) kHz	4.5×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators AC Current	40303	(2 ~ 5) A		Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 1 kHz	5.2×10^{-5}	
		(1 ~ 10) kHz	5.0×10^{-5}	
		(5 ~ 10) A		
		10 Hz	8.6×10^{-5}	
		10 Hz ~ 1 kHz	5.9×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(10 ~ 20) A		
		10 Hz	9.3×10^{-5}	
Wattmeter calibrators AC Voltage	40304	10 Hz ~ 1 kHz	6.8×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(20 ~ 50) A		
		10 Hz	1.0×10^{-4}	
		10 Hz ~ 1 kHz	8.3×10^{-5}	
		(1 ~ 10) kHz	1.1×10^{-4}	
		(50 ~ 100) A		
		10 Hz	1.2×10^{-4}	
		10 Hz ~ 1 kHz	9.7×10^{-5}	
		(1 ~ 10) kHz	1.3×10^{-4}	
		(100 ~ 200) A		
		60 Hz	4.5×10^{-4}	
		(1 ~ 2) mV		Power Standard, Counter/ SICT-CP-40304
		10 Hz ~ 10 kHz	1.8×10^{-3}	
		(10 ~ 100) kHz	3.1×10^{-3}	
		100 kHz ~ 1 MHz	1.2×10^{-2}	
		(2 ~ 3) mV		
		10 Hz ~ 10 kHz	1.1×10^{-3}	
		(10 ~ 100) kHz	1.7×10^{-3}	
		100 kHz ~ 1 MHz	7.7×10^{-3}	
		(3 ~ 4) mV		
		(10 ~ 40) Hz	6.4×10^{-4}	
		40 Hz ~ 10 kHz	5.8×10^{-4}	
		(10 ~ 100) kHz	1.0×10^{-3}	
		100 kHz ~ 1 MHz	5.4×10^{-3}	
		(4 ~ 6) mV		
		(10 ~ 40) Hz	5.3×10^{-4}	
		40 Hz ~ 10 kHz	4.6×10^{-4}	
		(10 ~ 100) kHz	8.1×10^{-4}	
		100 kHz ~ 1 MHz	4.6×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304			
AC Voltage		(6 ~ 10) mV		
		(10 ~ 40) Hz	4.2×10^{-4}	
		40 Hz ~ 10 kHz	3.5×10^{-4}	
		(10 ~ 100) kHz	5.8×10^{-4}	
		100 kHz ~ 1 MHz	3.8×10^{-3}	
		(10 ~ 60) mV		
		(10 ~ 40) Hz	2.5×10^{-4}	
		40 Hz ~ 10 kHz	2.1×10^{-4}	
		(10 ~ 100) kHz	3.6×10^{-4}	
		100 kHz ~ 1 MHz	2.7×10^{-3}	
		(60 ~ 100) mV		
		(10 ~ 40) Hz	1.1×10^{-4}	
		40 Hz ~ 10 kHz	6.5×10^{-5}	
		(10 ~ 100) kHz	1.2×10^{-4}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(100 ~ 600) mV		
		(10 ~ 40) Hz	9.1×10^{-5}	
		40 Hz ~ 10 kHz	4.8×10^{-5}	
		(10 ~ 100) kHz	9.0×10^{-5}	
		100 kHz ~ 1 MHz	1.2×10^{-3}	
		(0.6 ~ 20) V		
		(10 ~ 40) Hz	7.6×10^{-5}	
		40 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	7.4×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(20 ~ 1 000) V		
		(10 ~ 40) Hz	7.7×10^{-5}	
		40 Hz ~ 10 kHz	3.2×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-4}	
AC Current		(10 ~ 100) μ A		
		10 Hz ~ 10 kHz	2.6×10^{-3}	
		(0.1 ~ 1) mA		
		(10 ~ 40) Hz	3.8×10^{-4}	
		40 Hz ~ 1 kHz	3.7×10^{-4}	
		(1 ~ 10) kHz	6.5×10^{-4}	
		(1 ~ 100) mA		
		(10 ~ 40) Hz	9.7×10^{-5}	
		40 Hz ~ 1 kHz	7.3×10^{-5}	
		(1 ~ 10) kHz	9.0×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304			Power Standard, Counter/ SICT-CP-40304
AC Current		(0.1 ~ 5) A (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (5 ~ 20) A (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 100) A (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 200) A 60 Hz	7.8×10^{-5} 4.6×10^{-5} 4.3×10^{-5} 8.9×10^{-5} 6.3×10^{-5} 7.7×10^{-5} 1.1×10^{-4} 9.3×10^{-5} 1.3×10^{-4} 2.5×10^{-3}	
Active Power		(50 ~ 60) Hz 0 W (0 ~ 1) W (1 ~ 4.8) W 4.8 W ~ 5 kW (5 ~ 11) kW (11 ~ 52.8) kW	0.10 mW 5.0×10^{-4} 3.0×10^{-4} 1.6×10^{-4} 1.5×10^{-4} 1.4×10^{-4}	
Power Factor		(50 ~ 60) Hz -1 ~ 1	0.000 2	
THD-V		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 20) %	0.015 % 0.019 %	
THD-I		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 10) % (10 ~ 20) %	0.015 % 0.019 % 0.038 %	
Reactive Power		(50 ~ 60) Hz 0 var (0 ~ 1) var (1 ~ 4.8) var (4.8 ~ 48) var 48 var ~ 4.8 kvar (4.8 ~ 52.8) kvar	0.10 mvar 5.0×10^{-4} 2.2×10^{-4} 1.7×10^{-4} 1.6×10^{-4} 1.5×10^{-4}	
Apparent Power		(50 ~ 60) Hz (24 ~ 96 000) VA	5.9×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304	(±)		
	DC Voltage	0 mV	0.29 μ V	
		(0 ~ 20) mV	0.50 μ V	
		(20 ~ 50) mV	0.56 μ V	
		(50 ~ 100) mV	0.68 μ V	
		(0.1 ~ 1) V	6.8×10^{-6}	
		(1 ~ 10) V	3.6×10^{-6}	
		(10 ~ 100) V	5.2×10^{-6}	
		(100 ~ 1 000) V	5.6×10^{-6}	
	DC Current	(±)		
		0 μ A	5.5 nA	
		(0 ~ 100) μ A	5.8 nA	
		(0.1 ~ 100) mA	2.2×10^{-5}	
		(0.1 ~ 1) A	2.3×10^{-5}	
		(1 ~ 20) A	2.4×10^{-5}	
		(20 ~ 50) A	2.6×10^{-5}	
		(50 ~ 100) A	2.7×10^{-5}	
	DC Power	0 mW	0.1 μ W	
		(0 ~ 1) mW	6.2 μ W	
		(1 ~ 10) mW	30 μ W	
		10 mW ~ 10 W	3.1×10^{-3}	
		10 W ~ 1 kW	1.8×10^{-3}	
		(1 ~ 15) kW	5.4×10^{-3}	
		(15 ~ 30) kW	4.1×10^{-3}	
	Frequency	10 Hz ~ 1 MHz	1.0×10^{-6}	
Flicker(P_{st})		(50 Hz)		
		(1 ~ 4 000) cpm		
		1	0.003	
Flicker($P_{inst,max,Sinusoidal}$)		(50 Hz)		
		(0.5 ~ 1.5) Hz		
		1	0.007	
		(1.5 ~ 20) Hz		
		1	0.009	
		(20 ~ 33.333) Hz		
		1	0.008	
Flicker($P_{inst,max,Square}$)		(50 Hz)		
		(0.5 ~ 18) Hz		
		1	0.007	
		(18 ~ 25) Hz		
		1	0.009	
		(25 ~ 28) Hz		
		1	0.005	
		(28 ~ 33.333) Hz		
		1	0.015	
Flicker($P_{st, Range}$)		(50 Hz)		
		1 620 cpm		
		0.25	0.002	
		0.25 ~ 5	0.013	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC current shunts	40305	(40 ~ 60) Hz (1 ~ 10) mΩ (10 ~ 100) mΩ (60 Hz ~ 1 kHz) (1 ~ 100) mΩ (40 Hz ~ 1 kHz) 100 mΩ ~ 1 Ω (1 ~ 100) Ω 100 Ω ~ 10 kΩ	1.8×10^{-4} 3.5×10^{-4} 5.6×10^{-4} 3.5×10^{-4} 1.8×10^{-4} 2.3×10^{-4}	Reference Multimeter, Calibrator/ SICT-CP-40305
AC Voltage dividers		(50 Hz) (1 ~ 100) kV (60 Hz) (1 ~ 100) kV	1.9×10^{-4} 1.9×10^{-4}	
Phase angle generators, synchro resolve generators	40306	(-180 ~ 180) ° 50 Hz (50 ~ 500) Hz (500 ~ 1 000) Hz	0.001 6 ° 0.003 1 ° 0.010 °	Power Calibrator/ SICT-CP-40307
Voltage/current phase angle meters/synchro resolve meters	40307	(50 ~ 60) Hz (-180 ~ 180) °	0.008 8 °	Power Calibrator/ SICT-CP-40307
Potential transformer test sets	40308	60 Hz (110 ~ 1 100) V (-19.99 ~ 19.99) % 60 Hz (1 100 ~ 100 000) V (-19.99 ~ 19.99) % 60 Hz (110 ~ 100 000) V (-680 ~ 680) ′ 60 Hz 5 ~ 600 (50 ~ 60) Hz (1.25 ~ 2.5) VA (2.5 ~ 3.75) VA (3.75 ~ 10) VA (10 ~ 50) VA (50 ~ 100) VA (50 ~ 60) Hz 0.8 ~ 1	0.012 % 0.015 % 0.6 ′ 1.6×10^{-4} 6.7×10^{-4} 4.6×10^{-4} 3.7×10^{-4} 3.0×10^{-4} 2.6×10^{-4} 0.000 14	Standard Potential transforme, Ratio transformers/ SICT-CP-40308
Potential transformer	40309	(110 ~ 1 100) V (-19.99 ~ 19.99) % (-680 ~ 680) ′ (1 100 ~ 100 000) V (-19.99 ~ 19.99) % (-680 ~ 680) ′	0.020 % 0.70 ′ 0.016 % 0.50 ′	Standard Potential transforme/ SICT-CP-40309

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power factor meters AC Power Factor	40310	(50 Hz, 60 Hz) -1 ~ 1	1.1×10^{-4}	Power Calibrator/ SICT-CP-40310
AC power meters AC Voltage	40311	(1 ~ 2) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (2 ~ 6) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (6 ~ 20) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (20 ~ 60) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (60 ~ 200) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (200 ~ 600) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.6 ~ 2) V 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (2 ~ 6) V 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (6 ~ 20) V 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	4.8 $\times 10^{-3}$ 4.9 $\times 10^{-3}$ 6.5 $\times 10^{-3}$ 2.4 $\times 10^{-3}$ 2.6 $\times 10^{-3}$ 3.5 $\times 10^{-3}$ 9.0 $\times 10^{-4}$ 1.0 $\times 10^{-3}$ 1.6 $\times 10^{-3}$ 3.6 $\times 10^{-4}$ 4.7 $\times 10^{-4}$ 1.1 $\times 10^{-3}$ 2.1 $\times 10^{-4}$ 2.9 $\times 10^{-4}$ 7.1 $\times 10^{-4}$ 1.1 $\times 10^{-4}$ 1.8 $\times 10^{-4}$ 4.6 $\times 10^{-4}$ 7.1 $\times 10^{-5}$ 1.0 $\times 10^{-4}$ 1.6 $\times 10^{-4}$ 1.3 $\times 10^{-4}$ 1.5 $\times 10^{-4}$ 2.2 $\times 10^{-4}$ 8.1 $\times 10^{-5}$ 1.1 $\times 10^{-4}$ 1.5 $\times 10^{-4}$	Power Calibrator, Calibrator/ SICT-CP-40311

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
AC Voltage		(20 ~ 60) V		
		40 Hz ~ 50 kHz	1.6×10^{-4}	
		(50 ~ 100) kHz	3.4×10^{-4}	
		(60 ~ 200) V		
		40 Hz ~ 10 kHz	9.8×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.4×10^{-4}	
		(200 ~ 600) V		
		40 Hz ~ 1 kHz	1.8×10^{-4}	
		(1 ~ 10) kHz	2.5×10^{-4}	
		(10 ~ 20) kHz	2.0×10^{-3}	
		(600 ~ 1 000) V		
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	1.0×10^{-3}	
AC Current		(100 ~ 200) μ A		
		40 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 5) kHz	4.9×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(200 ~ 600) μ A		
		40 Hz ~ 1 kHz	3.2×10^{-4}	
		(1 ~ 5) kHz	7.7×10^{-4}	
		(5 ~ 10) kHz	4.0×10^{-3}	
		(0.6 ~ 2) mA		
		40 Hz ~ 1 kHz	2.1×10^{-4}	
		(1 ~ 5) kHz	4.9×10^{-4}	
		(5 ~ 10) kHz	2.6×10^{-3}	
		(2 ~ 6) mA		
		40 Hz ~ 1 kHz	3.2×10^{-4}	
		(1 ~ 5) kHz	6.9×10^{-4}	
		(5 ~ 10) kHz	3.8×10^{-3}	
		(6 ~ 20) mA		
		40 Hz ~ 1 kHz	2.1×10^{-4}	
		(1 ~ 5) kHz	4.3×10^{-4}	
		(5 ~ 10) kHz	2.4×10^{-3}	
		(20 ~ 60) mA		
		40 Hz ~ 1 kHz	3.0×10^{-4}	
		(1 ~ 5) kHz	6.5×10^{-4}	
		(5 ~ 10) kHz	2.6×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(60 ~ 200) mA		Power Calibrator, Calibrator/ SICT-CP-40311
		40 Hz ~ 1 kHz	2.0×10^{-4}	
		(1 ~ 5) kHz	4.0×10^{-4}	
		(5 ~ 10) kHz	1.8×10^{-3}	
		(200 ~ 600) mA		
		(40 ~ 50) Hz	5.8×10^{-4}	
		(50 ~ 60) Hz	1.4×10^{-4}	
		60 Hz ~ 1 kHz	5.3×10^{-4}	
		(1 ~ 5) kHz	1.2×10^{-3}	
		(5 ~ 10) kHz	9.3×10^{-3}	
		(0.6 ~ 2) A		
		(40 ~ 50) Hz	4.0×10^{-4}	
		(50 ~ 60) Hz	9.1×10^{-5}	
		60 Hz ~ 1 kHz	3.9×10^{-4}	
		(1 ~ 5) kHz	8.0×10^{-4}	
		(5 ~ 10) kHz	8.6×10^{-3}	
		(2 ~ 6) A		
		(40 ~ 50) Hz	3.3×10^{-4}	
		(50 ~ 60) Hz	9.7×10^{-5}	
		60 Hz ~ 1 kHz	7.9×10^{-4}	
		(1 ~ 5) kHz	8.2×10^{-4}	
		(5 ~ 10) kHz	8.2×10^{-3}	
		(6 ~ 20) A		
		(40 ~ 50) Hz	2.1×10^{-4}	
		(50 ~ 60) Hz	1.1×10^{-4}	
		60 Hz ~ 1 kHz	6.2×10^{-4}	
		(1 ~ 5) kHz	6.3×10^{-4}	
		(5 ~ 10) kHz	3.2×10^{-2}	
		(20 ~ 60) A		
		(40 ~ 50) Hz	2.4×10^{-4}	
		(50 ~ 60) Hz	1.8×10^{-4}	
		60 Hz ~ 1 kHz	6.5×10^{-4}	
		(1 ~ 5) kHz	6.6×10^{-4}	
		(5 ~ 10) kHz	6.2×10^{-2}	
		(60 ~ 100) A		
		(40 ~ 50) Hz	1.9×10^{-4}	
		(50 ~ 60) Hz	1.8×10^{-4}	
		60 Hz ~ 1 kHz	5.6×10^{-4}	
		(1 ~ 5) kHz	5.7×10^{-4}	
		(5 ~ 10) kHz	4.9×10^{-2}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(100 ~ 1 000) A (40 ~ 100) Hz 100 Hz ~ 1 kHz	1.3×10^{-3} 2.5×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40311
		(1 000 ~ 2 500) A (40 ~ 60) Hz	1.3×10^{-3}	
		(2 500 ~ 3 000) A 60 Hz	1.4×10^{-3}	
		(50 ~ 60) Hz 0 W	0.10 mW	
		(0 ~ 1) W	4.9×10^{-4}	
		(1 ~ 4.8) W	3.0×10^{-4}	
		4.8 W ~ 24 kW	1.6×10^{-4}	
		(24 ~ 52.8) kW	1.4×10^{-4}	
		(52.8 ~ 600) kW	1.3×10^{-3}	
		0 mV	0.79 μ V	
		(0 ~ 10) mV	0.85 μ V	
		(10 ~ 100) mV	1.5 μ V	
		(100 ~ 500) mV	1.5×10^{-5}	
		(500 ~ 900) mV	7.8×10^{-6}	
		(0.9 ~ 5) V	9.0×10^{-6}	
		(5 ~ 9) V	4.8×10^{-6}	
		(9 ~ 50) V	7.9×10^{-6}	
		(50 ~ 90) V	6.9×10^{-6}	
DC Voltage	40311	(90 ~ 500) V	9.9×10^{-6}	
		(500 ~ 900) V	8.8×10^{-6}	
		(900 ~ 1 000) V	1.0×10^{-5}	
		0 μ A	9.0 nA	
		(0 ~ 30) μ A	10 nA	
		(30 ~ 200) μ A	17 nA	
		(200 ~ 600) μ A	8.6×10^{-5}	
		(0.6 ~ 2) mA	6.0×10^{-5}	
		(2 ~ 6) mA	7.3×10^{-5}	
		(6 ~ 20) mA	5.3×10^{-5}	
		(20 ~ 60) mA	9.1×10^{-5}	
		(60 ~ 200) mA	6.9×10^{-5}	
		(200 ~ 600) mA	1.7×10^{-4}	
		(0.6 ~ 2) A	1.2×10^{-4}	
		(2 ~ 6) A	4.0×10^{-4}	
		(6 ~ 60) A	2.5×10^{-4}	
		(60 ~ 100) A	1.7×10^{-4}	
		(100 ~ 2 500) A	1.3×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			
DC Power		0 mW	0.06 μ W	
		(0 ~ 1) mW	0.08 μ W	
		(1 ~ 10) mW	8.0×10^{-5}	
		(10 ~ 100) mW	6.1×10^{-5}	
		(0.1 ~ 100) W	1.1×10^{-4}	
		100 W ~ 20 kW	1.9×10^{-4}	
		(20 ~ 1 000) kW	1.0×10^{-3}	
		(1 000 ~ 2 500) kW	1.3×10^{-3}	
Power Factor		(50 ~ 60) Hz		
		-1 ~ 1	0.000 2	
Frequency		10 Hz ~ 10 MHz	1.3×10^{-4}	
THD-V		(50 ~ 60) Hz		
		(0.5 ~ 1) %	0.015 %	
		(1 ~ 3) %	0.016 %	
		(3 ~ 5) %	0.022 %	
		(5 ~ 10) %	0.027 %	
		(10 ~ 20) %	0.046 %	
THD-I		(50 ~ 60) Hz		
		(0.5 ~ 1) %	0.015 %	
		(1 ~ 3) %	0.016 %	
		(3 ~ 5) %	0.019 %	
		(5 ~ 10) %	0.038 %	
		(10 ~ 20) %	0.060 %	
Flicker(P_{st})		(50 Hz)		
		(1 ~ 4 000) cpm		
		1	0.003	
Flicker($P_{inst,max,Sinusoidal}$)		(50 Hz)		
		(0.5 ~ 1.5) Hz		
		1	0.007	
		(1.5 ~ 20) Hz		
		1	0.009	
		(20 ~ 33.333) Hz		
		1	0.008	
Flicker($P_{inst,max,Square}$)		(50 Hz)		
		(0.5 ~ 8.8) Hz		
		1	0.007	
		(8.8 ~ 18) Hz		
		1	0.008	
		(18 ~ 25) Hz		
		1	0.010	
		(25 ~ 28) Hz		
		1	0.005	
		(28 ~ 33.333) Hz		
		1	0.015	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
Flicker ($P_{st, range}$)		(50 Hz) 1 620 cpm 0.25 0.25 ~ 5	0.002 0.013	
Reactive Power		(50 ~ 60) Hz 0 var (0 ~ 0.5) var (0.5 ~ 1) var (1 ~ 4.8) var (4.8 ~ 48) var 48 var ~ 4.8 kvar (4.8 ~ 52.8) kvar	92 μ var 4.6×10^{-4} 4.1×10^{-4} 1.9×10^{-4} 1.7×10^{-4} 1.6×10^{-4} 1.5×10^{-4}	
Apparent Power		(50 ~ 60) Hz 24 VA ~ 52.8 kVA	6.1×10^{-5}	
AC power supplies	40312			Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312
AC Voltage		(10 mV) 40 Hz ~ 5 kHz (10 ~ 100) mV 40 Hz ~ 5 kHz (100 mV ~ 1 V) 40 Hz ~ 5 kHz (1 ~ 10) V 40 Hz ~ 5 kHz (10 ~ 100) V 40 Hz ~ 5 kHz (100 ~ 600) V 40 Hz ~ 5 kHz (600 ~ 1 000) V 40 Hz ~ 5 kHz	2.2×10^{-4} 7.9×10^{-5} 6.7×10^{-5} 6.8×10^{-5} 7.0×10^{-5} 4.1×10^{-5} 7.1×10^{-5}	
Frequency		10 Hz (10 ~ 50) Hz (50 ~ 100) Hz (0.1 ~ 1) kHz (1 ~ 5) kHz	9.5×10^{-5} 1.9×10^{-5} 7.7×10^{-6} 8.4×10^{-7} 3.8×10^{-7}	
AC Current		(1 mA) (50 ~ 60) Hz (1 ~ 10) mA (50 ~ 60) Hz	6.4×10^{-4} 3.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
AC power supplies	AC Current	(10 ~ 100) mA (50 ~ 60) Hz	2.4×10^{-4}	Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312	
		(100 mA ~ 1 A) (50 ~ 60) Hz	2.1×10^{-4}		
		(1 ~ 10) A (50 ~ 60) Hz	2.3×10^{-4}		
		(10 ~ 20) A (50 ~ 60) Hz	4.0×10^{-4}		
		(20 ~ 30) A (50 ~ 60) Hz	6.4×10^{-4}		
		(30 ~ 50) A (50 ~ 60) Hz	4.2×10^{-4}		
	DC Voltage	(±)			
		0 mV	$5.8 \mu\text{V}$		
		(0 ~ 10) mV	5.8×10^{-4}		
		(10 ~ 100) mV	5.8×10^{-5}		
		(0.1 ~ 100) V	7.7×10^{-6}		
		(100 ~ 600) V	1.3×10^{-5}		
		(600 ~ 1 000) V	6.6×10^{-5}		
Load Regulation	DC Current	(1 ~ 10) mA	5.8×10^{-3}		
		(10 ~ 100) mA	5.9×10^{-4}		
		(0.1 ~ 1) A	2.4×10^{-4}		
		(1 ~ 10) A	3.1×10^{-4}		
		(10 ~ 300) A	2.4×10^{-4}		
		(300 ~ 500) A	2.6×10^{-4}		
		(500 ~ 1 000) A	4.7×10^{-5}		
		(1 000 ~ 3 000) A	5.1×10^{-4}		
Ripple		(0 ~ 2) mV	0.16 mV		
		(2 ~ 20) mV	7.8×10^{-2}		
		(20 ~ 200) mV	8.2×10^{-3}		
		(0.1 ~ 0.4) mV	3.8×10^{-1}		
		(0.4 ~ 0.6) mV	1.1×10^{-1}		
Harmonic Voltage		(0.6 ~ 1) mV	7.3×10^{-2}		
		(1 ~ 10) mV	4.4×10^{-2}		
		(10 ~ 50) mV	7.1×10^{-2}		
		(50 ~ 60) Hz			
Harmonic Current		0.5 %	0.050 %		
		(0.5 ~ 10) %	0.051 %		
		(10 ~ 20) %	0.082 %		
		(50 ~ 60) Hz			
		0.5 %	0.050 %		
		(0.5 ~ 20) %	0.051 %		

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
DC Voltage		(±) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 3.0×10^{-4} 2.3×10^{-4} 1.2×10^{-2}	
AC Voltage		(50 ~ 60) Hz 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.7×10^{-4} 1.2×10^{-2}	
AC Breaking Current		(50 ~ 60) Hz 0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	5.3×10^{-4} 4.4×10^{-4} 7.3×10^{-4} 7.1×10^{-4} 4.4×10^{-4} 3.6×10^{-4} 7.1×10^{-4} 4.4×10^{-4} 7.3×10^{-4}	
DC Breaking Current		0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	3.9×10^{-4} 1.3×10^{-4} 6.4×10^{-4} 3.2×10^{-4} 1.3×10^{-4} 6.5×10^{-5} 3.3×10^{-4} 1.4×10^{-4} 6.4×10^{-4}	
Resistance		1 mΩ (1 ~ 10) mΩ 10 mΩ ~ 100 kΩ	8.6×10^{-4} 7.2×10^{-4} 6.8×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 5 000) V (5 000 ~ 10 000) V	6.4×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 6.5×10^{-3} 6.1×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers				
Insulation Resistance	40313	1 kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ 10 TΩ	7.1×10^{-5} 3.7×10^{-5} 2.5×10^{-5} 3.1×10^{-5} 9.5×10^{-5} 2.4×10^{-5} 3.1×10^{-5} 6.1×10^{-5} 1.3×10^{-4} 2.6×10^{-4} 6.3×10^{-4}	AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
Leakage current(DC)		0 μA (0 ~ 1) μA (1 ~ 2) μA (2 ~ 5) μA (5 ~ 10) μA (10 ~ 20) μA (20 ~ 50) μA (50 ~ 100) μA (100 ~ 200) μA (0.2 ~ 100) mA	7.0 nA 7.0×10^{-3} 3.6×10^{-3} 1.4×10^{-3} 7.4×10^{-4} 4.0×10^{-4} 1.8×10^{-4} 1.3×10^{-4} 8.5×10^{-5} 6.1×10^{-4}	
Leakage current(AC)		(20 μA) 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (20 ~ 50) μA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (50 ~ 100) μA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (100 ~ 200) μA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.3×10^{-3} 8.5×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 5.5×10^{-3} 6.8×10^{-4} 4.4×10^{-4} 3.4×10^{-4} 6.8×10^{-4} 2.8×10^{-3} 4.9×10^{-4} 3.2×10^{-4} 2.3×10^{-4} 4.9×10^{-4} 4.0×10^{-4} 3.9×10^{-4} 2.5×10^{-4} 1.7×10^{-4} 4.0×10^{-4} 1.7×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313	(200 ~ 500) μ A		
		10 Hz	4.4×10^{-4}	AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
		(10 ~ 20) Hz	3.2×10^{-4}	
		20 Hz ~ 1 kHz	2.4×10^{-4}	
		(1 ~ 5) kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		500 μ A ~ 1 mA		
		10 Hz	7.0×10^{-4}	
		(10 ~ 20) Hz	6.6×10^{-4}	
		20 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		Leakage current(AC)		
		(1 ~ 100) mA	7.0×10^{-4}	
Output AC Current	40313	10 Hz	6.6×10^{-4}	
		(10 ~ 20) Hz	6.3×10^{-4}	
		20 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.0×10^{-4}	
		(5 ~ 10) kHz	2.7×10^{-3}	
		Output AC Current		
		(60 Hz)		
		1 A	1.2×10^{-3}	
		(1 ~ 3) A	1.5×10^{-3}	
		(3 ~ 20) A	9.7×10^{-4}	
Timer	40313	(20 ~ 30) A	1.0×10^{-3}	
		(30 ~ 60) A	8.4×10^{-4}	
		(60 ~ 100) A	1.0×10^{-3}	
		(100 ~ 150) A	4.6×10^{-3}	
Power recorders	40314	(150 ~ 200) A	3.7×10^{-3}	
		1 s	5.8×10^{-6}	
		(1 ~ 100) s	5.8×10^{-6}	
		(100 ~ 1 000) s	8.2×10^{-6}	
		(1 000 ~ 10 000) s	5.8×10^{-5}	
AC Wattage	40314	(50 ~ 60) Hz		
		0 mW	70 μ W	Power Energy Calibrator/ SICT-CP-40314
		(0 ~ 0.22) mW	2.1×10^{-1}	
		(0.22 ~ 1.1) mW	4.1×10^{-2}	
		(1.1 ~ 2.2) mW	2.1×10^{-2}	
		(2.2 ~ 11) mW	4.1×10^{-3}	
		(11 ~ 22) mW	2.1×10^{-3}	
		(22 ~ 44) mW	1.0×10^{-3}	
		(44 ~ 66) mW	7.0×10^{-4}	
		(66 ~ 88) mW	5.3×10^{-4}	
		(88 ~ 110) mW	4.3×10^{-4}	
		(110 ~ 480) mW	2.1×10^{-4}	
		480 mW ~ 12 kW	1.2×10^{-4}	
		(12 ~ 24) kW	6.8×10^{-4}	
		(24 ~ 300) kW	1.2×10^{-3}	
		(300 ~ 600) kW	1.4×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power recorders	40314	0 mW (0 ~ 1) mW (1 ~ 10) mW (10 ~ 100) mW (0.1 ~ 100) W (0.1 ~ 10) kW (10 ~ 20) kW (20 ~ 1 000) kW (1 000 ~ 2 500) kW	61 nW 7.7×10^{-5} 4.8×10^{-5} 6.1×10^{-5} 1.1×10^{-4} 1.9×10^{-4} 1.5×10^{-4} 1.0×10^{-3} 1.3×10^{-3}	Power Energy Calibrator/ SICT-CP-40314
Current transformer test sets	40315	60 Hz 5 A ~ 10 kA (-19.99 ~ 19.99) %	0.012 %	Current transforme, Ratio transformers/ SICT-CP-40315
Phase Angle Error		60 Hz 5 A ~ 10 kA (-680 ~ 680) °	0.6°	
Ratio		60 Hz 5 ~ 600	1.2×10^{-4}	
Burden		(50 ~ 60) Hz (1.25 ~ 2.5) VA (2.5 ~ 3.75) VA (3.75 ~ 5) VA (5 ~ 10) VA (10 ~ 20) VA (20 ~ 50) VA (50 ~ 100) VA	1.0×10^{-3} 6.6×10^{-4} 5.3×10^{-4} 4.9×10^{-4} 3.9×10^{-4} 3.5×10^{-4} 3.2×10^{-4}	
Power Factor		(50 ~ 60) Hz 0.8 ~ 1	0.001 4	
Current/turn current coil transformers	40316	(5 ~ 10 000) A (-19.99 ~ 19.99) %	0.020 %	Current transforme/ SICT-CP-40316
Ratio		(-680 ~ 680) °	0.70°	
Phase		(AC) 2 ~ 50	0.10 %	
Current Coil		(DC) 2 ~ 50	0.10 %	
transducers		(±) (10 A) 50 : 1 ~ 5 000 : 1	1.2×10^{-4}	
		(10 ~ 1 000) A 50 : 1 ~ 5 000 : 1	1.9×10^{-4}	
		(1 000 ~ 2 000) A 50 : 1 ~ 5 000 : 1	2.5×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters	40318			Reference Multimeter, Calibrator/ SICT-CP-40318
	AC Voltage	(600 μ V) 1 kHz	7.8×10^{-3}	
		(600 μ V ~ 1 mV) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	5.0×10^{-3} 4.8×10^{-3} 6.5×10^{-3}	
		(1 ~ 3) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	2.0×10^{-3} 1.7×10^{-3} 2.8×10^{-3}	
		(3 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.7×10^{-4} 5.7×10^{-4} 1.2×10^{-3}	
		(10 ~ 30) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	8.0×10^{-4} 3.7×10^{-4} 1.1×10^{-3}	
		(30 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	4.3×10^{-4} 1.7×10^{-4} 5.7×10^{-4}	
		(100 mV ~ 10 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	4.9×10^{-4} 1.1×10^{-4} 2.6×10^{-4}	
		(10 ~ 100) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	5.3×10^{-4} 1.3×10^{-4} 3.6×10^{-4}	
		(100 ~ 1 000) V 50 Hz 50 Hz ~ 1 kHz	3.7×10^{-4} 1.1×10^{-4}	
		(1 mV) 100 kHz 100 kHz ~ 1 MHz	6.5×10^{-3} 2.7×10^{-2}	
		(1 ~ 10) mV 100 kHz 100 kHz ~ 1 MHz	1.2×10^{-3} 5.6×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters	40318			Reference Multimeter, Calibrator/ SICT-CP-40318
AC Voltage		(10 ~ 100) mV		
		100 kHz	5.7×10^{-4}	
		100 kHz ~ 1 MHz	3.7×10^{-3}	
		(100 mV ~ 1 V)		
		100 kHz	1.5×10^{-4}	
		100 kHz ~ 1 MHz	2.3×10^{-3}	
		(1 ~ 10) V		
		100 kHz	4.7×10^{-5}	
		100 kHz ~ 1 MHz	7.0×10^{-4}	
		(10 ~ 20) V		
		100 kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(25 mV)		
		1 MHz	1.9×10^{-2}	
		(1 ~ 30) MHz	2.3×10^{-2}	
		(25 ~ 100) mV		
		1 MHz	2.3×10^{-2}	
		(1 ~ 30) MHz	2.8×10^{-2}	
		(100 ~ 300) mV		
		1 MHz	3.7×10^{-2}	
		(1 ~ 30) MHz	4.0×10^{-2}	
		(300 mV ~ 1 V)		
		1 MHz	2.4×10^{-2}	
		(1 ~ 30) MHz	2.7×10^{-2}	
		(1 ~ 2) V		
		1 MHz	1.5×10^{-2}	
		(1 ~ 30) MHz	1.8×10^{-2}	
AC Output Voltage		(1 mV)		
		10 Hz ~ 10 kHz	1.9×10^{-3}	
		(10 ~ 100) kHz	3.1×10^{-3}	
		(1 ~ 10) mV		
		10 Hz ~ 10 kHz	2.6×10^{-4}	
		(10 ~ 100) kHz	3.7×10^{-4}	
		(10 ~ 100) mV		
		10 Hz	9.2×10^{-5}	
		10 Hz ~ 10 kHz	4.8×10^{-5}	
		(10 ~ 100) kHz	9.1×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters	40318			Reference Multimeter, Calibrator/ SICT-CP-40318
AC Output Voltage		(100 mV ~ 1 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.2 × 10 ⁻⁵ 1.8 × 10 ⁻⁵ 5.9 × 10 ⁻⁵	
DC Output Voltage		1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V	7.5 × 10 ⁻⁴ 7.6 × 10 ⁻⁵ 9.4 × 10 ⁻⁶ 2.2 × 10 ⁻⁵	
Watt hour meters	40319			Power Calibrator/ SICT-CP-40319
Electric-Energy(Active Power)		(50 ~ 60) Hz 0 Wh (0 ~ 0.02) Wh (0.02 ~ 0.05) Wh (0.05 ~ 0.1) Wh (0.1 ~ 0.48) Wh (0.48 ~ 4.8) Wh 4.8 Wh ~ 5.28 kWh (5.28 ~ 15.84) kWh (15.84 ~ 26.4) kWh (26.4 ~ 42.24) kWh (42.24 ~ 52.8) kWh	0.10 mWh 0.10 mWh 0.15 mWh 0.46 mWh 3.0 × 10 ⁻⁴ 2.2 × 10 ⁻⁴ 2.1 × 10 ⁻⁴ 2.2 × 10 ⁻⁴ 2.9 × 10 ⁻⁴ 3.9 × 10 ⁻⁴ 4.6 × 10 ⁻⁴	
Electric-Energy(DC Power)		0 mWh (0 ~ 1) mWh (1 ~ 10) mWh (10 ~ 100) mWh 100 mWh ~ 100 Wh 100 Wh ~ 20 kWh (20 ~ 100) kWh	2 µWh 2 µWh 5 µWh 5.0 × 10 ⁻⁴ 4.5 × 10 ⁻⁴ 5.4 × 10 ⁻⁴ 5.1 × 10 ⁻⁴	
Electric-Energy(Reactive Power)		(50 ~ 60) Hz 0 varh (0 ~ 0.12) varh (0.12 ~ 0.3) varh (0.3 ~ 0.6) varh (0.6 ~ 2.88) varh 2.88 varh ~ 2.88 kvarh (2.88 ~ 31.68) kvarh (31.68 ~ 52.8) kvarh	0.10 mvarh 0.11 mvarh 0.18 mvarh 5.1 × 10 ⁻⁴ 3.8 × 10 ⁻⁴ 3.4 × 10 ⁻⁴ 3.3 × 10 ⁻⁴ 4.6 × 10 ⁻⁴	
Electric-Energy(Apparent Power)		(50 ~ 60) Hz 24 VAh ~ 52.8 kVAh	4.4 × 10 ⁻⁴	
Error rate in electric-energy meas. (Active Power)		(50 ~ 60) Hz, Single phase (120 ~ 480) V (0.2 ~ 200) A (-60 ~ 60) ° (-100 ~ 100) %	0.011 %	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Watt hour meters Error rate in electric-energy meas. (Active Power)	40319	(60 ~ 480) V (0.05 ~ 120) A (-0.25 ~ 0.5) (-100 ~ 100) %	0.023 %	Power Calibrator/ SICT-CP-40319
		(60 ~ 480) V (0.05 ~ 120) A (0.5 ~ 0.25) (-100 ~ 100) %	0.023 %	
		(60 ~ 480) V (0.05 ~ 120) A -1 (-100 ~ 100) %	0.021 %	
Error rate in electric-energy meas. (Active Power)		(50 ~ 60) Hz, Three phase (120 ~ 450) V (0.2 ~ 120) A (-60 ~ 60) ° (-100 ~ 100) %	0.012 %	
Error rate in electric-energy meas. (Reactive Power)		(50 ~ 60) Hz, Single phase (120 ~ 480) V (0.2 ~ 200) A (-90 ~ -30) ° (-100 ~ 100) %	0.011 %	
		(60 ~ 480) V (0.05 ~ 1) A -(0.5 ~ 0.25) (-100 ~ 100) %	0.032 %	
		(60 ~ 480) V (0.05 ~ 1) A (0.25 ~ 1) (-100 ~ 100) %	0.032 %	
		(60 ~ 480) V (1 ~ 200) A -(0.5 ~ 0.25) (-100 ~ 100) %	0.023 %	
		(60 ~ 480) V (1 ~ 200) A (0.25 ~ 1) (-100 ~ 100) %	0.023 %	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Watt hour meters Error rate in electric-energy meas. (Reactive Power)	40319	(50 ~ 60) Hz, Three phase (120 ~ 450) V (0.2 ~ 120) A (-90 ~ -30) ° (-100 ~ 100) %	0.012 %	Power Calibrator/ SICT-CP-40319
Error rate in electric-energy meas. (Apparent Power)		(50 ~ 60) Hz, Single phase (120 ~ 480) V (0.2 ~ 200) A (-60 ~ 60) ° (-100 ~ 100) %	0.007 %	
		(50 ~ 60) Hz, Three phase (120 ~ 450) V (0.2 ~ 120) A (-60 ~ 60) ° (-100 ~ 100) %	0.007 %	
Ratio transformers Ratio	40321	PT (±) (0.1 ~ 1.000) % (1.000 ~ 19.00) %	0.006 % 0.01 %	Calibrator/ SICT-CP-40321
		CT (±) (0.1 ~ 1.000) % (1.000 ~ 19.00) %	0.019 % 0.02 %	
Phase		PT (±) (0.040 ~ 1.999)' (1.999 ~ 19.99)' (19.99 ~ 199.9)' (199.9 ~ 600)'	0.060' 0.06' 0.2' 1'	
		CT (±) (0.040 ~ 1.999)' (1.999 ~ 19.99)' (19.99 ~ 199.9)' (199.9 ~ 600)'	0.060' 0.06' 0.2' 1'	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF amplifiers	40401 Amplifier	(DC)		Reference Multimeter/ SICT-CP-40401
		1 mV	0.4 μ V	
		(1 ~ 10) mV	7.4×10^{-5}	
		(10 ~ 100) mV	6.1×10^{-5}	
		(0.1 ~ 1) V	6.0×10^{-5}	
		(1 ~ 1 000) V	1.0×10^{-4}	
		(10 Hz ~ 10 kHz)		
		1 mV	1.7 μ V	
		(1 ~ 10) mV	2.6×10^{-4}	
		(10 ~ 100) mV	1.1×10^{-4}	
		(0.1 ~ 1) V	9.0×10^{-5}	
		(1 ~ 1 000) V	1.0×10^{-4}	
		(10 ~ 100) kHz		
		1 mV	3.1 μ V	
		(1 ~ 10) mV	3.7×10^{-4}	
		(10 ~ 100) mV	1.1×10^{-4}	
		(0.1 ~ 1) V	8.0×10^{-5}	
		(1 ~ 1 000) V	1.0×10^{-4}	
DC/LF attenuators	40402 Attenuation	10 Hz ~ 100 kHz		Reference Multimeter/ SICT-CP-40402
		(0 ~ -20) dB	0.001 9 dB	
		(-20 ~ -60) dB	0.001 7 dB	
		(-60 ~ -70) dB	0.005 5 dB	
		(-70 ~ -80) dB	0.008 7 dB	
Multimeter calibrators	40403 DC Voltage	(\pm)		Reference Multimeter/ SICT-CP-40403
		0 mV	0.05 μ V	
		(0 ~ 100) mV	3.3×10^{-6}	
		(0.1 ~ 1) V	1.4×10^{-6}	
		(1 ~ 10) V	1.0×10^{-6}	
		(10 ~ 100) V	1.6×10^{-6}	
		(100 ~ 1 000) V	2.1×10^{-6}	
	DC Current	(\pm)		
		1 nA	7.0 pA	
		(1 ~ 100) nA	4.7×10^{-3}	
		100 nA ~ 10 A	1.2×10^{-5}	
		(10 ~ 50) A	4.0×10^{-5}	
	AC Voltage	(50 ~ 100) A	4.4×10^{-5}	
		(1 mV)		
		10 Hz ~ 10 kHz	1.7×10^{-3}	
		(10 ~ 100) kHz	3.0×10^{-3}	
		100 kHz ~ 1 MHz	1.2×10^{-2}	
		(1 ~ 2) mV		
		10 Hz ~ 10 kHz	1.1×10^{-3}	
		(10 ~ 100) kHz	1.7×10^{-3}	
		100 kHz ~ 1 MHz	7.7×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
AC Voltage		(2 ~ 5) mV		
		10 Hz	6.4×10^{-4}	
		10 Hz ~ 10 kHz	5.8×10^{-4}	
		(10 ~ 100) kHz	1.0×10^{-3}	
		100 kHz ~ 1 MHz	5.4×10^{-3}	
		(5 ~ 10) mV		
		10 Hz	4.2×10^{-4}	
		10 Hz ~ 10 kHz	3.5×10^{-4}	
		(10 ~ 100) kHz	5.8×10^{-4}	
		100 kHz ~ 1 MHz	3.9×10^{-3}	
		(10 ~ 20) mV		
		10 Hz	1.8×10^{-4}	
		10 Hz ~ 10 kHz	1.4×10^{-4}	
		(10 ~ 100) kHz	2.2×10^{-4}	
		100 kHz ~ 1 MHz	2.2×10^{-3}	
		(20 ~ 50) mV		
		10 Hz	1.4×10^{-4}	
		10 Hz ~ 10 kHz	9.2×10^{-5}	
		(10 ~ 100) kHz	1.6×10^{-4}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(50 ~ 100) mV		
		10 Hz	1.1×10^{-4}	
		10 Hz ~ 10 kHz	6.6×10^{-5}	
		(10 ~ 100) kHz	1.2×10^{-4}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(100 ~ 200) mV		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 10 kHz	3.9×10^{-5}	
		(10 ~ 100) kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(200 ~ 500) mV		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 10 kHz	3.6×10^{-5}	
		(10 ~ 100) kHz	7.1×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(0.5 ~ 1) V		
		10 Hz	7.6×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	6.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(1 ~ 2) V		
		10 Hz	7.1×10^{-5}	
		10 Hz ~ 10 kHz	2.7×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-5}	
		100 kHz ~ 1 MHz	1.0×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
AC Voltage		(2 ~ 5) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.6×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(5 ~ 20) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.8×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	3.0×10^{-5}	
		(10 ~ 100) kHz	8.0×10^{-5}	
		(50 ~ 200) V		
		10 Hz	7.4×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	8.5×10^{-5}	
		(200 ~ 1 000) V		
		10 Hz	7.7×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-4}	
AC Current		(10 µA)		
		10 Hz ~ 10 kHz	2.6×10^{-3}	
		(10 ~ 100) µA		
		10 Hz ~ 1 kHz	3.6×10^{-4}	
		(1 ~ 10) kHz	6.4×10^{-4}	
		(0.1 ~ 1) mA		
		10 Hz	9.8×10^{-5}	
		10 Hz ~ 1 kHz	7.5×10^{-5}	
		(1 ~ 10) kHz	9.4×10^{-5}	
		(1 ~ 100) mA		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 1 kHz	4.6×10^{-5}	
		(1 ~ 10) kHz	4.2×10^{-5}	
		(0.1 ~ 1) A		
		10 Hz	8.1×10^{-5}	
		10 Hz ~ 1 kHz	4.9×10^{-5}	
		(1 ~ 10) kHz	4.4×10^{-5}	
		(1 ~ 2) A		
		10 Hz	7.9×10^{-5}	
		10 Hz ~ 1 kHz	4.7×10^{-5}	
		(1 ~ 10) kHz	4.5×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
AC Current		(2 ~ 5) A		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 1 kHz	5.2×10^{-5}	
		(1 ~ 10) kHz	5.0×10^{-5}	
		(5 ~ 10) A		
		10 Hz	8.6×10^{-5}	
		10 Hz ~ 1 kHz	5.9×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(10 ~ 20) A		
		10 Hz	9.3×10^{-5}	
		10 Hz ~ 1 kHz	6.8×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(20 ~ 50) A		
		10 Hz	1.0×10^{-4}	
		10 Hz ~ 1 kHz	8.3×10^{-5}	
		(1 ~ 10) kHz	1.1×10^{-4}	
		(50 ~ 100) A		
		10 Hz	1.2×10^{-4}	
		10 Hz ~ 1 kHz	9.7×10^{-5}	
		(1 ~ 10) kHz	1.3×10^{-4}	
		(100 ~ 200) A		
		60 Hz	4.5×10^{-4}	
Resistance		0 Ω	$0.14 \mu\Omega$	
		(0 ~ 1) Ω	6.6×10^{-6}	
		(1 ~ 1.9) Ω	8.4×10^{-6}	
		(1.9 ~ 10) Ω	3.6×10^{-6}	
		(10 ~ 19) Ω	2.6×10^{-6}	
		(19 ~ 100) Ω	2.8×10^{-6}	
		(0.1 ~ 1) kΩ	2.5×10^{-6}	
		(1 ~ 1.9) kΩ	3.8×10^{-6}	
		(1.9 ~ 10) kΩ	2.0×10^{-6}	
		(10 ~ 19) kΩ	1.3×10^{-6}	
		(19 ~ 100) kΩ	1.9×10^{-6}	
		(100 ~ 190) kΩ	2.0×10^{-6}	
		(0.19 ~ 1) MΩ	2.9×10^{-6}	
		(1 ~ 1.9) MΩ	3.1×10^{-6}	
		(1.9 ~ 10) MΩ	3.6×10^{-6}	
		(10 ~ 19) MΩ	2.9×10^{-6}	
		(19 ~ 100) MΩ	1.5×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
Multimeter calibrators(property) (Digital sampling)				
AC Voltage		(1 mV) 0.1 Hz ~ 3 kHz	8.4×10^{-4}	
		(1 ~ 2) mV 0.1 Hz ~ 3 kHz	4.2×10^{-4}	
		(2 ~ 3) mV 0.1 Hz ~ 3 kHz	2.8×10^{-4}	
		(3 ~ 5) mV 0.1 Hz ~ 3 kHz	1.7×10^{-4}	
		(5 ~ 10) mV 0.1 Hz ~ 3 kHz	8.8×10^{-5}	
		(10 ~ 20) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(20 ~ 30) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 ~ 50) mV 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(50 ~ 100) mV 0.1 Hz ~ 3 kHz	2.6×10^{-5}	
		(100 ~ 200) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 ~ 300) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(300 ~ 500) mV 0.1 Hz ~ 3 kHz	2.8×10^{-5}	
		(500 mV ~ 1 V) 0.1 Hz ~ 3 kHz	2.4×10^{-5}	
		(1 ~ 2) V 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(2 ~ 3) V 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(3 ~ 5) V 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(5 ~ 10) V 0.1 Hz ~ 3 kHz	2.6×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
Multimeter calibrators(property) (Digital sampling)				
AC Voltage		(10 ~ 30) V 10 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 ~ 50) V 10 Hz ~ 3 kHz	2.8×10^{-5}	
		(50 ~ 100) V 10 Hz ~ 3 kHz	2.4×10^{-5}	
		(100 ~ 200) V 10 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 ~ 1 000) V 50 Hz ~ 1 kHz	2.4×10^{-5}	
Oscilloscope calibrators	40404			Calibrator/ SICT-CP-40404
DC Voltage Amplitude		(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V	0.50 μ V 4.2×10^{-4} 2.1×10^{-5} 8.5×10^{-5} 4.3×10^{-5} 2.1×10^{-5} 8.5×10^{-6} 7.2×10^{-6} 3.8×10^{-6} 2.8×10^{-6} 6.3×10^{-6} 3.7×10^{-6} 2.9×10^{-6} 6.3×10^{-6} 3.8×10^{-6} 3.3×10^{-6} 6.4×10^{-6} 3.9×10^{-6} 3.8×10^{-6}	
AC Voltage Amplitude		(10 Hz ~ 10 kHz) 1 mV (1 ~ 2) mV (2 ~ 10) mV (10 ~ 500) mV (0.5 ~ 100) V (100 ~ 200) V	0.76 μ V 8.3×10^{-5} 8.4×10^{-5} 6.0×10^{-5} 5.8×10^{-5} 4.0×10^{-5}	
Sine Wave Generator		(100 ~ 600) mV 50 kHz (50 ~ 500) kHz 0.5 MHz ~ 1 GHz (1 ~ 6) GHz	0.58 mV 1.0×10^{-3} 1.7×10^{-2} 1.9×10^{-2}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscope calibrators	40404			Calibrator/ SICT-CP-40404
Sine Wave Generator		(600 mV ~ 1 V) 50 kHz (50 ~ 500) kHz 0.5 MHz ~ 1 GHz (1 ~ 6) GHz	0.58 mV 1.0×10^{-3} 1.7×10^{-2} 1.9×10^{-2}	
Time Marker Generator		(0.1 ~ 1) ns (1 ~ 10) ns (10 ~ 100) ns 0.1 μ s ~ 10 ms (10 ~ 100) ms (0.1 ~ 1) s (1 ~ 5) s	5.8×10^{-8} 6.5×10^{-9} 3.1×10^{-9} 5.8×10^{-8} 6.1×10^{-9} 5.8×10^{-8} 1.2×10^{-8}	
Impedance Mesurement		(50 ~ 75) Ω 75 Ω ~ 1 M Ω	1.7×10^{-4} 2.1×10^{-4}	
CD/DVD meters/analyzers	40405			Modulation Domain Analyzer/ SICT-CP-40405
Jitter		(1.0 ~ 60.0) ns 1 % 2 % 4 % 8 % 10 % 15 %	1.7×10^{-3} 0.05 % 0.09 % 0.19 % 0.36 % 0.44 % 0.67 %	
Video signal generators	40406			Video Measurement/ SICT-CP-40406
NTSC, PAL Multiburst		(0.1 ~ 1) MHz (1 ~ 2) MHz (2 ~ 6) MHz	6.0×10^{-2} 6.2×10^{-3} 3.1×10^{-3}	
NTSC, PAL, SECAM Pulse and Bar		(0 ~ 300) ns (0 ~ 1 000) mV	4.2×10^{-4} 3.5×10^{-3}	
NTSC, PAL , SECAM Frequency		1 Hz ~ 10 MHz	1.6×10^{-9}	
Video frequency		(10 ~ 100) Hz 100 Hz ~ 500 MHz	6.2×10^{-8} 6.2×10^{-9}	
Video level		(30 ~ 600) mV (600 ~ 1 200) mV	2.6×10^{-3} 2.3×10^{-3}	
TTL Sync level		(1 ~ 5) V	2.7×10^{-3}	
D-TV Level		(30 ~ 600) mV (600 ~ 1 200) mV	2.6×10^{-3} 2.3×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal generators NTSC, PAL, H-Timming(Level) (Time)	40406	(0 ~ 100) mV	2.6×10^{-3}	Video Measurement/ SICT-CP-40406
		(100 ~ 1 000) mV	3.4×10^{-3}	
		(0 ~ 254) ns	1.2×10^{-2}	
		(254 ~ 300) ns	3.8×10^{-3}	
		300 ns ~ 3 μ s	3.2×10^{-3}	
		(3 ~ 7) μ s	7.4×10^{-3}	
		(7 ~ 10) μ s	4.2×10^{-3}	
		NTSC, PAL Color Bar(Luminance Level)	0.06 mV	
		(0 ~ 100) mV	3.4×10^{-3}	
		(100 ~ 1 000) mV	0.06 mV	
		(100 ~ 1 000) mV	3.4×10^{-3}	
NTSC, PAL Color Bar(Chrominance Level)		(0 ~ 360) $^{\circ}$	0.13 $^{\circ}$	
		SECAM Color Bar Frequency	1.2×10^{-3}	
		(D'R & D'B) (3 ~ 5) MHz	1.2×10^{-3}	
RF Output frequency		10 kHz ~ 10 MHz	6.0×10^{-4}	
		(10 ~ 100) MHz	6.0×10^{-5}	
		(100 ~ 1 000) MHz	6.0×10^{-6}	
RF Output level		(0.1 ~ 10) mV	1.4×10^{-2}	
		(10 ~ 500) mV	1.3×10^{-2}	
Sound Frequency		10 Hz ~ 100 kHz	6.1×10^{-8}	
		100 kHz ~ 1 MHz	6.1×10^{-7}	
Audio distortion analyzers/meters Input Frequency	40407	1 Hz ~ 200 kHz	6.1×10^{-7}	Calibrator/ SICT-CP-40407
		(10 ~ 100) kHz	0.008 3 dB	
		Input DC Voltage	0.27 μ V	
		0 mV	5.8×10^{-3}	
		(0 ~ 1) mV	5.8×10^{-4}	
		(1 ~ 10) mV	5.8×10^{-4}	
		10 mV ~ 300 V	5.8×10^{-4}	
		Input Distortion	(100 Hz ~ 10 kHz)	
		(-10 ~ -40) dB	0.005 8 dB	
		(-40 ~ -50) dB	0.006 0 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Audio distortion analyzers/meters				
Input Distortion	40407	(10 kHz ~ 50 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 9 dB 0.006 3 dB 0.008 2 dB 0.019 dB 0.052 dB	Calibrator/ SICT-CP-40407
Input AC Voltage		(10 ~ 100) Hz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 100 V (100 ~ 300) V (100 Hz ~ 1 kHz) (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (100 ~ 300) V (1 ~ 10) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (10 ~ 100) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	9.0 × 10 ⁻⁴ 4.0 × 10 ⁻⁴ 4.2 × 10 ⁻⁴ 5.3 × 10 ⁻⁴ 8.4 × 10 ⁻⁴ 1.8 × 10 ⁻⁴ 1.1 × 10 ⁻⁴ 1.0 × 10 ⁻⁴ 2.3 × 10 ⁻⁴ 8.4 × 10 ⁻⁴ 1.8 × 10 ⁻⁴ 1.5 × 10 ⁻⁴ 2.7 × 10 ⁻⁴ 1.4 × 10 ⁻³ 7.6 × 10 ⁻⁴ 4.8 × 10 ⁻⁴ 4.1 × 10 ⁻⁴ 3.4 × 10 ⁻⁴	
Input Attenuation		(10 Hz) (30 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB (10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (10 ~ 100) kHz (30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -80) dB	0.006 8 dB 0.016 dB 0.052 dB 0.008 3 dB 0.014 dB 0.042 dB 0.009 1 dB 0.023 dB 0.057 dB	
Input Impedance		300 Ω ~ 200 kΩ	3.1 × 10 ⁻⁴	
Input Filter		(10 Hz ~ 100 kHz) 1 V	8.3 × 10 ⁻⁴	
(Distortion meter calibrator)				
Distortion		(400 Hz , 1 kHz) (-10 ~ -20) dB (-20 ~ -40) dB (-40 ~ -60) dB (-60 ~ -80) dB	0.15 dB 0.14 dB 0.17 dB 0.26 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF filters	40408	10 Hz ~ 50 kHz (50 ~ 100) kHz (100 ~ 150) kHz	5.8×10^{-4} 1.2×10^{-3} 5.8×10^{-3}	Audio Analyzer/ SICT-CP-40408
LF/audio signal analyzers	40409	1 Hz ~ 200 kHz Output Frequency AC Output Level AC Output Level Flatness Output Attenuation Output DC Offset Output Impedance Input Frequency AC Input Level Flatness DC Input Level Input Distortion	5.8 $\times 10^{-6}$ (10 ~ 100) Hz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm (100 Hz ~ 10 kHz) (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm (10 ~ 100) kHz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm 10 Hz ~ 100 kHz (0 ~ -60) dB (\pm) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 50 V 5 Ω (10 ~ 600) Ω 1 Hz ~ 200 kHz 10 Hz ~ 100 kHz (\pm) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 300 V (100 Hz ~ 10 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	Calibrator, Reference Multimeter/ SICT-CP-40409 8.7 $\times 10^{-4}$ 9.4×10^{-5} 0.005 8 dB 8.7 $\times 10^{-4}$ 6.5×10^{-5} 0.005 8 dB 8.7 $\times 10^{-4}$ 9.4×10^{-5} 0.005 8 dB 0.007 1 dB 0.005 8 dB 0.7 μ V 1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5} 1.2×10^{-3} 6.0×10^{-4} 6.1×10^{-7} 0.008 3 dB 0.27 μ V 5.8×10^{-3} 5.8×10^{-4} 5.8×10^{-4} 0.005 8 dB 0.006 0 dB 0.006 8 dB 0.012 dB 0.028 dB

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF/audio signal analyzers				Calibrator, Reference
Input Distortion	40409	(10 kHz ~ 50 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 9 dB 0.006 3 dB 0.008 2 dB 0.019 dB 0.052 dB	Multimeter/ SICT-CP-40409
AC Input Level		(10 ~ 100) Hz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 100 V (100 ~ 300) V (100 Hz ~ 1 kHz) (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (100 ~ 300) V (1 ~ 10) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (10 ~ 100) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	9.0×10^{-4} 4.0×10^{-4} 4.2×10^{-4} 5.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.1×10^{-4} 1.0×10^{-4} 2.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 2.7×10^{-4} 1.4×10^{-3} 7.6×10^{-4} 4.1×10^{-4} 3.4×10^{-4} 2.6×10^{-4}	
Input Attenuation		(10 Hz) (30 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB (10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (10 ~ 100) kHz (30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -80) dB	0.006 8 dB 0.016 dB 0.052 dB 0.008 3 dB 0.014 dB 0.042 dB 0.009 1 dB 0.023 dB 0.057 dB	
Input Impedance		300 Ω ~ 200 kΩ	3.1×10^{-4}	
Input Filter		(10 Hz ~ 100 kHz) 1 V	8.3×10^{-4}	
Line frequency meters				Calibrator/ SICT-CP-40410
Frequency	40410	16 Hz ~ 1 kHz	1.3×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	Frequency	40411	$(0.01 \sim 0.1)$ Hz	5.8×10^{-6}
			$(0.1 \sim 1)$ Hz	5.8×10^{-7}
			1 Hz ~ 1 GHz	5.8×10^{-9}
			$(1 \sim 4)$ GHz	1.5×10^{-8}
	Output Level		$(10 \sim 100)$ Hz	
			1 mV	1.0×10^{-3}
			$(1 \sim 10)$ mV	1.0×10^{-4}
			$10 \text{ mV} \sim 100$ V	7.0×10^{-5}
			$(100 \text{ Hz} \sim 10 \text{ kHz})$	
			1 mV	1.0×10^{-3}
			$(1 \sim 10)$ mV	1.0×10^{-4}
			$10 \text{ mV} \sim 100$ V	3.0×10^{-5}
	DC Offset		$(10 \sim 100)$ kHz	
			1 mV	1.0×10^{-3}
			$(1 \sim 10)$ mV	1.0×10^{-4}
			$10 \text{ mV} \sim 100$ V	8.0×10^{-5}
			(\pm)	
Level Flatness	Level Flatness		0 mV	$0.7 \mu\text{V}$
			$(0 \sim 1)$ mV	$0.7 \mu\text{V}$
			$(1 \sim 10)$ mV	1.0×10^{-4}
			$10 \text{ mV} \sim 20$ V	6.0×10^{-5}
			(100 mV)	
			$(10 \sim 100)$ Hz	0.099 dB
			100 Hz ~ 10 kHz	0.083 dB
			$(10 \sim 100)$ kHz	0.095 dB
			$(100 \text{ mV} \sim 1 \text{ V})$	
			$(10 \sim 100)$ Hz	0.005 4 dB
Attenuation	Attenuation		100 Hz ~ 10 kHz	0.001 1 dB
			$(10 \sim 100)$ kHz	0.007 2 dB
			$(1 \sim 30)$ V	
			$(10 \sim 100)$ Hz	0.021 dB
			100 Hz ~ 10 kHz	0.015 dB
			$(10 \sim 100)$ kHz	0.027 dB
			$(10 \text{ Hz} \sim 100 \text{ kHz})$	
			$(0 \sim 80)$ dB	0.006 1 dB
	Distortion		$(20 \text{ Hz} \sim 1 \text{ kHz})$	
			$(3.16 \sim 0.010)$ %	1.5×10^{-1}
			$(1 \sim 100)$ kHz	
			$(3.16 \sim 0.010)$ %	3.2×10^{-1}

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	40411	100 μ s ~ 100 ns	7.0×10^{-4}	Audio Analyzer, Digital Multimeter/ SICT-CP-40411
		(100 ~ 10) ns	7.8×10^{-4}	
		(10 ~ 1) ns	4.7×10^{-3}	
		1 ns ~ 100 ps	4.6×10^{-2}	
		Duty cycle (1 ~ 99) %	0.006 1 %	
High Frequency Flatness Test		(100 kHz ~ 80 MHz)		
		(0 ~ 20) dBm	0.11 dB	
		FM Modulation (0.1 ~ 400) kHz	1.2×10^{-2}	
AM Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Genescopes	40412	9 kHz ~ 10 MHz	2.8×10^{-6}	Signal Generator/ SICT-CP-40412
		(10 ~ 200) MHz	6.4×10^{-7}	
		9 kHz ~ 200 MHz		
		(100 ~ 50) dB μ V	0.31 dB	
AC/DC high voltages voltmeters	40413	(\pm)		Calibrator/ SICT-CP-40413
		0 kV	0.58 V	
		(0 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.1×10^{-4}	
		(1 ~ 2) kV	4.4×10^{-4}	
		(2 ~ 100) kV	3.4×10^{-4}	
		(50 Hz)		
		0.01 kV	0.58 V	
		(0.01 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.2×10^{-4}	
		(1 ~ 2) kV	5.5×10^{-4}	
		(2 ~ 3) kV	5.3×10^{-4}	
		(3 ~ 15 kV	5.0×10^{-4}	
		(15 ~ 100) kV	5.7×10^{-4}	
		(60 Hz)		
		0.01 kV	0.58 V	
		(0.01 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.2×10^{-4}	
		(1 ~ 2) kV	5.5×10^{-4}	
		(2 ~ 3) kV	4.7×10^{-4}	
		(3 ~ 15 kV	4.5×10^{-4}	
		(15 ~ 100) kV	5.4×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Jitter meters	40415			Modulation Domain Analyzer/ SICT-CP-40415
CD/DVD Jitter		(1 ~ 20) ns (20 ~ 60) ns	1.7×10^{-3} 1.6×10^{-3}	
VTR Jitter		0.05 μs (0.05 ~ 0.1) μs (0.1 ~ 0.2) μs (0.2 ~ 0.5) μs (0.5 ~ 0.7) μs 1 % 2 % 4 % 8 % 10 % 15 %	0.66 ns 0.77 ns 1.2 ns 2.8 ns 4.3 ns 0.05 % 0.09 % 0.19 % 0.36 % 0.44 % 0.67 %	
Leakage current testers	40416			Calibrator/ SICT-CP-40416
DC Current		0 μA (0 ~ 1) μA (1 ~ 2) μA (2 ~ 5) μA (5 ~ 10) μA (10 ~ 20) μA (20 ~ 50) μA (50 ~ 100) μA (100 ~ 200) μA (0.2 ~ 100) mA	7.0 nA 2.4×10^{-3} 3.6×10^{-3} 1.4×10^{-3} 7.4×10^{-4} 4.0×10^{-4} 1.8×10^{-4} 1.3×10^{-4} 8.5×10^{-5} 6.1×10^{-4}	
AC Current		(20 μA) 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (20 ~ 50) μA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (50 ~ 100) μA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz	26 nA 8.5×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 5.5×10^{-3} 6.8×10^{-4} 4.4×10^{-4} 3.4×10^{-4} 6.8×10^{-4} 2.8×10^{-3} 4.9×10^{-4} 3.2×10^{-4} 2.3×10^{-4} 4.9×10^{-4} 4.0×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416			Calibrator/ SICT-CP-40416
AC Current		(100 ~ 200) μ A 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (200 ~ 500) μ A 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.5 ~ 1) mA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 100) mA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz	3.9×10^{-4} 2.5×10^{-4} 1.7×10^{-4} 4.0×10^{-4} 1.7×10^{-3} 4.4×10^{-4} 3.2×10^{-4} 2.4×10^{-4} 5.4×10^{-4} 2.8×10^{-3} 7.0×10^{-4} 6.6×10^{-4} 6.3×10^{-4} 7.2×10^{-4} 2.1×10^{-3} 7.0×10^{-4} 6.6×10^{-4} 6.3×10^{-4} 7.0×10^{-4} 2.7×10^{-3}	
DC Voltage		0 V (0 ~ 0.1) V (0.1 ~ 0.2) V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 300) V (300 ~ 500) V (500 ~ 1 000) V	0.06 mV 6.0×10^{-4} 3.0×10^{-4} 1.2×10^{-4} 6.0×10^{-5} 3.1×10^{-4} 1.2×10^{-4} 6.1×10^{-5} 3.1×10^{-5} 1.4×10^{-5} 8.8×10^{-6} 3.1×10^{-5} 2.3×10^{-5} 1.2×10^{-4} 5.8×10^{-5}	
AC Voltage		0.1 V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	0.074 mV 6.5×10^{-4} 8.3×10^{-4} 1.2×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416			Calibrator/ SICT-CP-40416
AC Voltage		(0.1 ~ 0.2) V		
		10 Hz	4.5×10^{-4}	
		(0.01 ~ 50) kHz	3.5×10^{-4}	
		(50 ~ 100) kHz	5.5×10^{-4}	
		(100 ~ 300) kHz	9.5×10^{-4}	
		(300 ~ 500) kHz	1.8×10^{-3}	
		(0.5 ~ 1) MHz	3.4×10^{-3}	
		(0.2 ~ 0.5) V		
		10 Hz	4.0×10^{-4}	
		(10 ~ 20) Hz	2.0×10^{-4}	
		(0.02 ~ 50) kHz	1.7×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 300) kHz	6.0×10^{-4}	
		(300 ~ 500) kHz	1.7×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	
		(0.5 ~ 1) V		
		10 Hz	3.3×10^{-4}	
		(10 ~ 20) Hz	1.4×10^{-4}	
		(0.02 ~ 20) kHz	9.1×10^{-5}	
		(20 ~ 100) kHz	1.5×10^{-4}	
		(100 ~ 300) kHz	4.9×10^{-4}	
		(300 ~ 500) kHz	1.4×10^{-3}	
		(0.5 ~ 1) MHz	2.4×10^{-3}	
		(1 ~ 2) V		
		10 Hz	4.3×10^{-4}	
		(0.01 ~ 100) kHz	3.3×10^{-4}	
		(100 ~ 300) kHz	5.5×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.1×10^{-3}	
		(2 ~ 5) V		
		10 Hz	4.0×10^{-4}	
		(0.01 ~ 100) kHz	2.0×10^{-4}	
		(100 ~ 300) kHz	5.0×10^{-4}	
		(300 ~ 500) kHz	1.7×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416			Calibrator/ SICT-CP-40416
AC Voltage		(5 ~ 10) V		
		10 Hz	3.3×10^{-4}	
		(10 ~ 20) Hz	1.4×10^{-4}	
		(0.02 ~ 20) kHz	8.9×10^{-5}	
		(20 ~ 100) kHz	1.4×10^{-4}	
		(100 ~ 300) kHz	3.9×10^{-4}	
		(300 ~ 500) kHz	1.4×10^{-3}	
		(0.5 ~ 1) MHz	2.2×10^{-3}	
		(10 ~ 20) V		
		10 Hz	3.1×10^{-4}	
		(10 ~ 20) Hz	1.2×10^{-4}	
		(0.02 ~ 20) kHz	6.0×10^{-5}	
		(20 ~ 50) kHz	9.0×10^{-5}	
		(50 ~ 100) kHz	1.1×10^{-4}	
		(20 ~ 50) V		
		10 Hz	4.2×10^{-4}	
		(10 ~ 20) Hz	2.2×10^{-4}	
		(0.02 ~ 50) kHz	1.8×10^{-4}	
		(50 ~ 100) kHz	2.8×10^{-4}	
		(50 ~ 100) V		
		10 Hz	3.4×10^{-4}	
		(0.01 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 1 000) V		
		(0.05 ~ 1) kHz	1.1×10^{-4}	
Resistance		100 mΩ	$7.7 \mu\Omega$	
		1 Ω ~ 10 kΩ	6.2×10^{-5}	
Input Voltage to Output Current Display(U1)		20 Hz		
		(4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz		
		(4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz		
		(4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz		
		(4.85 ~ 5.36) mA	0.006 1 mA	
		200 Hz		
		(5.11 ~ 5.65) mA	0.006 1 mA	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U1)	40416	500 Hz (6.64 ~ 7.34) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
		1 kHz (9.70 ~ 10.73) mA	0.006 1 mA	
		2 kHz (14.07 ~ 15.56) mA	0.006 2 mA	
		5 kHz (17.82 ~ 19.70) mA	0.006 2 mA	
		10 kHz (18.66 ~ 20.63) mA	0.006 2 mA	
		20 kHz (18.92 ~ 20.92) mA	0.006 2 mA	
		50 kHz (19.00 ~ 21.00) mA	0.006 4 mA	
		100 kHz (19.00 ~ 21.00) mA	0.006 7 mA	
		200 kHz (19.00 ~ 21.00) mA	0.010 mA	
		500 kHz (19.00 ~ 21.00) mA	0.030 mA	
		1 MHz (19.00 ~ 21.00) mA	0.046 mA	
Input Voltage to Output Current Display(U2)		20 Hz (4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA	
		200 Hz (4.92 ~ 5.44) mA	0.006 1 mA	
		500 Hz (5.37 ~ 5.93) mA	0.006 1 mA	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U2)	40416	1 kHz (5.56 ~ 6.14) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
		2 kHz (4.68 ~ 5.17) mA	0.006 1 mA	
		5 kHz (2.53 ~ 2.80) mA	0.000 63 mA	
		10 kHz (1.35 ~ 1.49) mA	0.000 62 mA	
		20 kHz (0.683 ~ 0.755) mA	0.000 61 mA	
		50 kHz (274.57 ~ 303.47) mA	0.029 μ A	
		100 kHz (137.48 ~ 151.95) μ A	0.020 μ A	
		200 kHz (68.82 ~ 76.06) μ A	0.030 μ A	
		500 kHz (27.43 ~ 30.32) μ A	0.042 μ A	
		1 MHz (13.71 ~ 15.16) μ A	0.033 μ A	
Input Voltage to Output Current Display(U3)		20 Hz (4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA	
		200 Hz (4.95 ~ 5.47) mA	0.006 1 mA	
		500 Hz (5.65 ~ 6.25) mA	0.006 1 mA	
		1 kHz (6.60 ~ 7.29) mA	0.006 1 mA	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U3)	40416	2 kHz (7.14 ~ 7.89) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
		5 kHz (5.31 ~ 5.87) mA	0.006 1 mA	
		10 kHz (3.12 ~ 3.45) mA	0.000 64 mA	
		20 kHz (1.63 ~ 1.81) mA	0.000 62 mA	
		50 kHz (0.664 ~ 0.734) mA	0.000 62 mA	
		100 kHz (322.16 ~ 367.12) µA	0.046 µA	
		200 kHz (166.03 ~ 183.81) µA	0.070 µA	
		500 kHz (66.37 ~ 73.35) µA	0.10 µA	
		1 MHz (33.14 ~ 36.63) µA	0.08 µA	
Input Voltage to Output Voltage Ratio(U1)		4.00 (20 Hz)	1.3×10^{-4}	
		3.98 (50 Hz)	6.5×10^{-5}	
		3.97 (60 Hz)	6.5×10^{-5}	
		3.92 (100 Hz)	6.5×10^{-5}	
		3.72 (200 Hz)	6.5×10^{-5}	
		2.87 (500 Hz)	6.4×10^{-5}	
		1.96 (1 kHz)	6.4×10^{-5}	
		1.96 (2 kHz)	6.4×10^{-5}	
		1.96 (5 kHz)	6.4×10^{-5}	
		1.96 (10 kHz)	6.4×10^{-5}	
		1.00 (20 kHz)	6.7×10^{-5}	
		1.00 (50 kHz)	9.6×10^{-5}	
		1.00 (100 kHz)	1.2×10^{-4}	
		1.00 (200 kHz)	4.2×10^{-4}	
		1.00 (500 kHz)	1.5×10^{-3}	
		1.00 (1 MHz)	2.6×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers				
Input Voltage to Output Voltage Ratio(U2)	40416	4.00 (20 Hz) 3.99 (50 Hz) 3.99 (60 Hz) 3.96 (100 Hz) 3.87 (200 Hz) 3.54 (500 Hz) 3.43 (1 kHz) 4.06 (2 kHz) 7.50 (5 kHz) 14.1 (10 kHz) 27.8 (20 kHz) 69.2 (50 kHz) 138 (100 kHz) 272 (200 kHz) 691 (500 kHz) 1 382 (1 MHz)	1.3×10^{-4} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.6×10^{-5} 7.0×10^{-5} 3.6×10^{-5} 4.8×10^{-5} 1.0×10^{-4} 2.7×10^{-4} 1.1×10^{-3} 3.0×10^{-3}	Calibrator/ SICT-CP-40416
Input Voltage to Output Voltage Ratio(U3)		4.00 (20 Hz) 3.99 (50 Hz) 3.98 (60 Hz) 3.95 (100 Hz) 3.83 (200 Hz) 2.36 (500 Hz) 2.87 (1 kHz) 2.65 (2 kHz) 3.57 (5 kHz) 6.09 (10 kHz) 11.6 (20 kHz) 28.7 (50 kHz) 57.2 (100 kHz) 114 (200 kHz) 286 (500 kHz) 572 (1 MHz)	1.3×10^{-4} 6.5×10^{-5} 6.5×10^{-5} 6.6×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.4×10^{-5} 6.4×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.9×10^{-5} 3.8×10^{-5} 7.9×10^{-5} 1.9×10^{-4} 6.1×10^{-4} 2.3×10^{-3}	
mAs Meter		1 mAs (1 ~ 2 000) mAs (2 000 ~ 9 999) mAs	1.2×10^{-3} 1.0×10^{-3} 1.1×10^{-3}	
Electronic AC/DC loads				
DC Voltage	40417	0 mV (0 ~ 5) mV (5 ~ 20) mV (20 ~ 100) mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 4) V (4 ~ 7) V (7 ~ 9) V (9 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 400) V (400 ~ 1 000) V	0.058 mV 5.8×10^{-2} 5.8×10^{-3} 1.2×10^{-3} 6.2×10^{-5} 3.2×10^{-5} 2.1×10^{-5} 1.3×10^{-5} 9.1×10^{-6} 7.9×10^{-6} 3.1×10^{-5} 1.0×10^{-5} 3.4×10^{-5} 2.5×10^{-5} 1.6×10^{-5}	Calibrator/ SICT-CP-40417

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	40417			
DC Current		1 mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA (0.1 ~ 0.2) A (0.2 ~ 0.4) A (0.4 ~ 0.6) A (0.6 ~ 0.8) A (0.8 ~ 1) A (1 ~ 3) A (3 ~ 6) A (6 ~ 10) A (10 ~ 40) A (40 ~ 100) A (100 ~ 1 000) A (1 000 ~ 2 000) A	5.8 μ A 2.9×10^{-3} 1.2×10^{-3} 5.8×10^{-4} 1.2×10^{-4} 5.9×10^{-5} 2.9×10^{-4} 1.9×10^{-4} 1.2×10^{-4} 8.4×10^{-5} 6.6×10^{-5} 5.1×10^{-5} 2.6×10^{-5} 1.6×10^{-5} 4.0×10^{-5} 2.9×10^{-5} 1.4×10^{-4} 4.9×10^{-4}	Calibrator/ SICT-CP-40417
Charge voltage		0 mV (0 ~ 5) mV (5 ~ 20) mV (20 ~ 100) mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 4) V (4 ~ 7) V (7 ~ 9) V (9 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 400) V (400 ~ 1 000) V (1 000 ~ 1 200) V (1 200 ~ 1 400) V (1 400 ~ 1 500) V	0.058 mV 5.8×10^{-2} 5.8×10^{-3} 1.2×10^{-3} 6.2×10^{-5} 3.2×10^{-5} 2.1×10^{-5} 1.3×10^{-5} 9.1×10^{-6} 7.9×10^{-6} 3.1×10^{-5} 1.0×10^{-5} 3.4×10^{-5} 2.5×10^{-5} 1.6×10^{-5} 1.0×10^{-3} 9.2×10^{-4} 8.7×10^{-4}	
Charge and Discharge Current		(±) 1 mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA (0.1 ~ 0.2) A (0.2 ~ 0.4) A (0.4 ~ 0.6) A (0.6 ~ 0.8) A (0.8 ~ 1) A (1 ~ 3) A (3 ~ 6) A (6 ~ 10) A (10 ~ 40) A (40 ~ 100) A (100 ~ 1 000) A (1 000 ~ 3 000) A	5.8 μ A 2.9×10^{-3} 1.2×10^{-3} 5.8×10^{-4} 1.2×10^{-4} 5.9×10^{-5} 2.9×10^{-4} 1.9×10^{-4} 1.2×10^{-4} 8.4×10^{-5} 6.6×10^{-5} 5.1×10^{-5} 2.6×10^{-5} 1.6×10^{-5} 4.0×10^{-5} 2.9×10^{-5} 1.4×10^{-4} 4.9×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	40417			Calibrator/ SICT-CP-40417
Resistance		0.1 Ω (0.1 ~ 1) Ω (1 ~ 2) Ω (2 ~ 4) Ω (4 ~ 500) Ω (0.5 ~ 2) kΩ (2 ~ 10) kΩ	0.58 mΩ 2.1×10^{-3} 4.0×10^{-3} 2.3×10^{-3} 1.6×10^{-3} 9.0×10^{-4} 2.0×10^{-4}	
AC Voltage		(0.001 V) (40 ~ 400) Hz (0.001 ~ 0.1) V (40 ~ 400) Hz (0.1 ~ 0.2) V (40 ~ 400) Hz (0.2 ~ 0.5) V (40 ~ 400) Hz (0.5 ~ 2) V (40 ~ 400) Hz (2 ~ 3) V (40 ~ 400) Hz (3 ~ 7) V (40 ~ 400) Hz (7 ~ 20) V (40 ~ 50) Hz (50 ~ 400) Hz (20 ~ 80) V (40 ~ 50) Hz (50 ~ 400) Hz (80 ~ 200) V (40 ~ 400) Hz (200 ~ 500) V (50 ~ 400) Hz	0.61 mV 6.1×10^{-3} 3.1×10^{-3} 1.2×10^{-3} 6.2×10^{-4} 2.8×10^{-4} 2.3×10^{-4} 1.5×10^{-4} 9.8×10^{-5} 2.1×10^{-4} 1.2×10^{-4} 1.3×10^{-4} 1.8×10^{-4}	
AC Current		(1 mA) (40 ~ 400) Hz (1 ~ 100) mA (40 ~ 400) Hz (100 mA ~ 0.2 A) (40 ~ 400) Hz	0.58 mA 5.8×10^{-2} 5.8×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	AC Current	40417 (0.2 ~ 0.6) A (40 ~ 400) Hz	2.0×10^{-3}	Calibrator/ SICT-CP-40417
		(0.6 ~ 2) A (40 ~ 400) Hz	9.4×10^{-4}	
		(2 ~ 5) A (40 ~ 400) Hz	1.2×10^{-3}	
		(5 ~ 20) A (40 ~ 400) Hz	1.0×10^{-3}	
	AC Resistance	(1 Ω) (40 ~ 400) Hz	$1.0 \text{ m}\Omega$	
		(1 ~ 50) Ω (40 ~ 400) Hz	1.5×10^{-3}	
		(50 ~ 100) Ω (40 ~ 400) Hz	1.1×10^{-3}	
		(100 Ω ~ 10 kΩ) (40 ~ 400) Hz	1.7×10^{-3}	
Modulation meters	Frequency Modulation	40418 0 kHz (0 ~ 400) kHz	1 Hz 1.2×10^{-2}	Measuring Receiver/ SICT-CP-40418
		0 % (0 ~ 100) %	0.01 % 1.2×10^{-2}	
		0 rad (0 ~ 400) rad	1.2 mrad 1.2×10^{-2}	
	Amplitude Modulation			
Analogue/digital multimeters	DC Voltage	40419 (±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 15) mV (15 ~ 20) mV (20 ~ 50) mV (0.05 ~ 0.2) V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V (500 ~ 1 000) V	$0.43 \mu\text{V}$ 5.0×10^{-4} 2.5×10^{-4} 1.0×10^{-4} 5.0×10^{-5} 3.3×10^{-5} 2.5×10^{-6} 1.2×10^{-5} 8.0×10^{-6} 4.8×10^{-6} 3.8×10^{-6} 4.0×10^{-6} 2.6×10^{-6} 2.3×10^{-6} 6.0×10^{-6} 4.0×10^{-6} 3.5×10^{-6} 8.0×10^{-6} 5.2×10^{-6} 4.5×10^{-6}	Calibrator/ SICT-CP-40419

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters				
AC Voltage	40419	(0.6 mV) 1 kHz (1 mV) 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (1 ~ 2) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (2 ~ 5) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (5 ~ 10) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (10 ~ 15) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	4.1 μ V 4.2 μ V 4.2 μ V 4.1 μ V 4.2 μ V 5.5 μ V 11 μ V 21 μ V 23 μ V 2.2 \times 10 ⁻³ 2.2 \times 10 ⁻³ 2.1 \times 10 ⁻³ 2.2 \times 10 ⁻³ 3.0 \times 10 ⁻³ 6.0 \times 10 ⁻³ 1.1 \times 10 ⁻² 1.3 \times 10 ⁻² 1.1 \times 10 ⁻³ 9.2 \times 10 ⁻⁴ 9.0 \times 10 ⁻⁴ 1.0 \times 10 ⁻³ 1.5 \times 10 ⁻³ 3.0 \times 10 ⁻³ 5.2 \times 10 ⁻³ 6.8 \times 10 ⁻³ 6.3 \times 10 ⁻⁴ 5.0 \times 10 ⁻⁴ 4.9 \times 10 ⁻⁴ 5.9 \times 10 ⁻⁴ 9.5 \times 10 ⁻⁴ 1.9 \times 10 ⁻³ 3.2 \times 10 ⁻³ 4.3 \times 10 ⁻³ 4.8 \times 10 ⁻⁴ 3.6 \times 10 ⁻⁴ 3.5 \times 10 ⁻⁴ 4.5 \times 10 ⁻⁴ 7.6 \times 10 ⁻⁴ 1.5 \times 10 ⁻³ 2.5 \times 10 ⁻³ 3.7 \times 10 ⁻³	Calibrator/ SICT-CP-40419

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Voltage		(15 ~ 20) mV		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	2.9×10^{-4}	
		(0.04 ~ 20) kHz	2.8×10^{-4}	
		(20 ~ 50) kHz	3.8×10^{-4}	
		(50 ~ 100) kHz	6.7×10^{-4}	
		(100 ~ 300) kHz	1.4×10^{-3}	
		(300 ~ 500) kHz	2.2×10^{-3}	
		(0.5 ~ 1) MHz	3.4×10^{-3}	
		(20 ~ 50) mV		
		10 Hz	4.6×10^{-4}	
		(10 ~ 40) Hz	2.6×10^{-4}	
		(0.04 ~ 20) kHz	2.1×10^{-4}	
		(20 ~ 50) kHz	2.7×10^{-4}	
		(50 ~ 100) kHz	6.6×10^{-4}	
		(100 ~ 300) kHz	1.0×10^{-3}	
		(300 ~ 500) kHz	1.6×10^{-3}	
		(0.5 ~ 1) MHz	3.3×10^{-3}	
		(50 ~ 100) mV		
		10 Hz	3.3×10^{-4}	
		(10 ~ 40) Hz	1.6×10^{-4}	
		(0.04 ~ 20) kHz	1.3×10^{-4}	
		(20 ~ 50) kHz	1.9×10^{-4}	
		(50 ~ 100) kHz	4.8×10^{-4}	
		(100 ~ 300) kHz	7.6×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	
		(100 ~ 150) mV		
		10 Hz	2.9×10^{-4}	
		(10 ~ 40) Hz	1.3×10^{-4}	
		(0.04 ~ 20) kHz	1.0×10^{-4}	
		(20 ~ 50) kHz	1.6×10^{-4}	
		(50 ~ 100) kHz	4.1×10^{-4}	
		(100 ~ 300) kHz	6.8×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	
		(150 ~ 200) mV		
		10 Hz	2.7×10^{-4}	
		(10 ~ 40) Hz	1.2×10^{-4}	
		(0.04 ~ 20) kHz	9.2×10^{-5}	
		(20 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	3.9×10^{-4}	
		(100 ~ 300) kHz	6.4×10^{-4}	
		(300 ~ 500) kHz	1.2×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Voltage		(0.2 ~ 0.5) V		
		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.3×10^{-4}	
		(20 ~ 40) Hz	9.0×10^{-5}	
		(0.04 ~ 20) kHz	6.6×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(100 ~ 300) kHz	4.6×10^{-4}	
		(300 ~ 500) kHz	1.2×10^{-3}	
		(0.5 ~ 1) MHz	2.7×10^{-3}	
		(0.5 ~ 1) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		(20 ~ 40) Hz	5.8×10^{-5}	
		(0.04 ~ 20) kHz	4.9×10^{-5}	
		(20 ~ 50) kHz	7.7×10^{-5}	
		(50 ~ 100) kHz	1.2×10^{-4}	
		(100 ~ 300) kHz	3.6×10^{-4}	
		(300 ~ 500) kHz	1.0×10^{-3}	
		(0.5 ~ 1) MHz	2.5×10^{-3}	
		(1 ~ 2) V		
		10 Hz	2.2×10^{-4}	
		(10 ~ 20) Hz	8.5×10^{-5}	
		(20 ~ 40) Hz	4.5×10^{-5}	
		(0.04 ~ 20) kHz	4.2×10^{-5}	
		(20 ~ 50) kHz	6.8×10^{-5}	
		(50 ~ 100) kHz	9.6×10^{-5}	
		(100 ~ 300) kHz	3.2×10^{-4}	
		(300 ~ 500) kHz	9.0×10^{-4}	
		(0.5 ~ 1) MHz	2.4×10^{-3}	
		(2 ~ 5) V		
		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.3×10^{-4}	
		(20 ~ 40) Hz	8.2×10^{-5}	
		(0.04 ~ 20) kHz	6.2×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(100 ~ 300) kHz	4.4×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.2×10^{-3}	
		(5 ~ 10) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	9.8×10^{-5}	
		(20 ~ 40) Hz	5.4×10^{-5}	
		(0.04 ~ 20) kHz	4.7×10^{-5}	
		(20 ~ 50) kHz	7.7×10^{-5}	
		(50 ~ 100) kHz	1.1×10^{-4}	
		(100 ~ 300) kHz	3.2×10^{-4}	
		(300 ~ 500) kHz	1.0×10^{-3}	
		(0.5 ~ 1) MHz	1.7×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Voltage		(10 ~ 15) V		
		10 Hz	2.3×10^{-4}	
		(10 ~ 20) Hz	8.7×10^{-5}	
		(20 ~ 40) Hz	4.6×10^{-5}	
		(0.04 ~ 20) kHz	4.3×10^{-5}	
		(20 ~ 50) kHz	7.1×10^{-5}	
		(50 ~ 100) kHz	9.5×10^{-5}	
		(100 ~ 300) kHz	2.9×10^{-4}	
		(300 ~ 500) kHz	9.4×10^{-4}	
		(0.5 ~ 1) MHz	1.5×10^{-3}	
		(15 ~ 20) V		
		10 Hz	2.2×10^{-4}	
		(10 ~ 20) Hz	8.5×10^{-5}	
		(20 ~ 40) Hz	4.3×10^{-5}	
		(0.04 ~ 20) kHz	4.1×10^{-5}	
		(20 ~ 50) kHz	6.8×10^{-5}	
		(50 ~ 100) kHz	9.0×10^{-5}	
		(100 ~ 300) kHz	2.8×10^{-4}	
		(300 ~ 500) kHz	9.1×10^{-4}	
		(0.5 ~ 1) MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	3.2×10^{-4}	
		(10 ~ 20) Hz	1.4×10^{-4}	
		(20 ~ 40) Hz	9.8×10^{-5}	
		(0.04 ~ 20) kHz	7.4×10^{-5}	
		(20 ~ 50) kHz	1.1×10^{-5}	
		(50 ~ 100) kHz	2.1×10^{-4}	
		(50 ~ 100) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		(20 ~ 40) Hz	6.5×10^{-5}	
		(0.04 ~ 20) kHz	5.6×10^{-5}	
		(20 ~ 50) kHz	8.5×10^{-5}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(100 ~ 200) V		
		10 Hz	2.3×10^{-4}	
		(10 ~ 20) Hz	9.3×10^{-5}	
		(20 ~ 40) Hz	5.6×10^{-5}	
		(0.04 ~ 20) kHz	5.1×10^{-5}	
		(20 ~ 50) kHz	7.9×10^{-5}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(200 ~ 500) V		
		50 Hz ~ 1 kHz	6.7×10^{-5}	
		(500 ~ 1 000) V		
		50 Hz ~ 1 kHz	6.3×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	Resistance	0 Ω	0.001 0 mΩ	Calibrator/ SICT-CP-40419
		0 Ω ~ 10 kΩ	1.2×10^{-6}	
		(10 ~ 100) kΩ	1.4×10^{-6}	
	DC Current	(0.1 ~ 1) MΩ	7.2×10^{-6}	
		(1 ~ 10) MΩ	7.7×10^{-6}	
		(10 ~ 100) MΩ	1.2×10^{-5}	
		(0.1 ~ 1) GΩ	3.2×10^{-4}	
		(±)		
		0 nA	6.0 nA	
		(0 ~ 1) nA	6.9×10^{-3}	
		(1 ~ 100) nA	4.6×10^{-3}	
		(0.1 ~ 1) μA	6.0×10^{-3}	
		(1 ~ 2) μA	3.1×10^{-3}	
		(2 ~ 5) μA	1.2×10^{-3}	
		(5 ~ 10) μA	6.3×10^{-4}	
		(10 ~ 20) μA	3.5×10^{-4}	
		(20 ~ 50) μA	1.6×10^{-4}	
		(50 ~ 100) μA	9.4×10^{-5}	
		(100 ~ 200) μA	6.3×10^{-5}	
		(0.2 ~ 0.5) mA	4.8×10^{-5}	
		(0.5 ~ 1) mA	3.5×10^{-5}	
		(1 ~ 1.5) mA	3.1×10^{-5}	
		(1.5 ~ 2) mA	3.0×10^{-5}	
		(2 ~ 5) mA	4.4×10^{-5}	
		(5 ~ 10) mA	3.2×10^{-5}	
		(10 ~ 15) mA	2.9×10^{-5}	
		(15 ~ 20) mA	2.8×10^{-5}	
		(20 ~ 50) mA	5.6×10^{-5}	
		(50 ~ 100) mA	4.4×10^{-5}	
		(100 ~ 150) mA	4.1×10^{-5}	
		(150 ~ 200) mA	3.9×10^{-5}	
		(0.2 ~ 0.5) A	9.4×10^{-5}	
		(0.5 ~ 1) A	6.9×10^{-5}	
		(1 ~ 1.5) A	6.1×10^{-5}	
		(1.5 ~ 2) A	5.8×10^{-5}	
		(2 ~ 3) A	3.3×10^{-4}	
		(3 ~ 5) A	2.4×10^{-4}	
		(5 ~ 10) A	1.6×10^{-4}	
		(10 ~ 20) A	1.2×10^{-4}	
		(20 ~ 30) A	2.4×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Current		(20 μ A)		
		1 kHz	11 nA	
		10 kHz	81 nA	
		(20 ~ 50) μ A		
		10 Hz	1.4×10^{-4}	
		(10 ~ 20) Hz	9.0×10^{-5}	
		20 Hz ~ 1 kHz	7.3×10^{-5}	
		(1 ~ 5) kHz	1.5×10^{-4}	
		(5 ~ 10) kHz	5.5×10^{-4}	
		(50 ~ 100) μ A		
		10 Hz	1.9×10^{-4}	
		(10 ~ 20) Hz	1.2×10^{-4}	
		20 Hz ~ 1 kHz	9.5×10^{-5}	
		(1 ~ 5) kHz	2.0×10^{-4}	
		(5 ~ 10) kHz	7.5×10^{-4}	
		(0.1 ~ 0.2) mA		
		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.9×10^{-4}	
		20 Hz ~ 1 kHz	1.4×10^{-5}	
		(1 ~ 5) kHz	3.1×10^{-4}	
		(5 ~ 10) kHz	1.2×10^{-3}	
		(0.2 ~ 0.5) mA		
		10 Hz	9.0×10^{-5}	
		(10 ~ 20) Hz	7.0×10^{-5}	
		20 Hz ~ 1 kHz	6.0×10^{-5}	
		(1 ~ 5) kHz	1.2×10^{-4}	
		(5 ~ 10) kHz	5.7×10^{-4}	
		(0.5 ~ 1) mA		
		10 Hz	1.4×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		20 Hz ~ 1 kHz	8.0×10^{-5}	
		(1 ~ 5) kHz	1.6×10^{-4}	
		(5 ~ 10) kHz	7.6×10^{-4}	
		(1 ~ 2) mA		
		10 Hz	2.4×10^{-4}	
		(10 ~ 20) Hz	1.6×10^{-4}	
		20 Hz ~ 1 kHz	1.2×10^{-4}	
		(1 ~ 5) kHz	2.4×10^{-4}	
		(5 ~ 10) kHz	1.2×10^{-3}	
		(2 ~ 5) mA		
		10 Hz	9.0×10^{-5}	
		(10 ~ 20) Hz	7.0×10^{-5}	
		20 Hz ~ 1 kHz	5.2×10^{-5}	
		(1 ~ 5) kHz	1.1×10^{-4}	
		(5 ~ 10) kHz	5.4×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Current		(5 ~ 10) mA		
		10 Hz	1.4×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		20 Hz ~ 1 kHz	7.3×10^{-5}	
		(1 ~ 5) kHz	1.4×10^{-4}	
		(5 ~ 10) kHz	7.2×10^{-4}	
		(10 ~ 20) mA		
		10 Hz	2.4×10^{-4}	
		(10 ~ 20) Hz	1.6×10^{-4}	
		20 Hz ~ 1 kHz	1.2×10^{-4}	
		(1 ~ 5) kHz	2.2×10^{-4}	
		(5 ~ 10) kHz	1.1×10^{-3}	
		(20 ~ 50) mA		
		10 Hz	1.0×10^{-4}	
		(10 ~ 20) Hz	8.0×10^{-5}	
		20 Hz ~ 1 kHz	4.8×10^{-5}	
		(1 ~ 5) kHz	1.1×10^{-4}	
		(5 ~ 10) kHz	4.0×10^{-4}	
		(50 ~ 100) mA		
		10 Hz	1.4×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		20 Hz ~ 1 kHz	6.8×10^{-5}	
		(1 ~ 5) kHz	1.4×10^{-4}	
		(5 ~ 10) kHz	6.0×10^{-4}	
		(0.1 ~ 0.2) A		
		10 Hz	2.4×10^{-4}	
		(10 ~ 20) Hz	1.6×10^{-4}	
		20 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 5) kHz	2.1×10^{-4}	
		(5 ~ 10) kHz	1.0×10^{-3}	
		(0.2 ~ 1) A		
		40 Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.4×10^{-4}	
		(1 ~ 5) kHz	2.6×10^{-4}	
		(5 ~ 10) kHz	2.7×10^{-3}	
		(1 ~ 2) A		
		40 Hz ~ 1 kHz	2.4×10^{-4}	
		(1 ~ 5) kHz	4.2×10^{-4}	
		(5 ~ 10) kHz	5.2×10^{-3}	
		(2 ~ 3) A		
		(40 ~ 100) Hz	1.8×10^{-4}	
		100 Hz ~ 1 kHz	1.9×10^{-4}	
		(1 ~ 10) kHz	9.9×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters				
AC Current	40419	(3 ~ 5) A (40 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (5 ~ 10) A (40 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 30) A (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz	2.4×10^{-4} 2.5×10^{-4} 1.6×10^{-3} 4.2×10^{-4} 4.2×10^{-4} 3.1×10^{-3} 6.0×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 2.3×10^{-3} 8.0×10^{-4} 1.2×10^{-3} 3.9×10^{-3} 4.5×10^{-3}	Calibrator/ SICT-CP-40419
Frequency (Digital Smapling)		10 Hz ~ 10 MHz	6.4×10^{-7}	
AC Voltage		(1 mV) 0.1 Hz ~ 3 kHz (1 mV ~ 2 mV) 0.1 Hz ~ 3 kHz (2 mV ~ 3 mV) 0.1 Hz ~ 3 kHz (3 mV ~ 5 mV) 0.1 Hz ~ 3 kHz (5 mV ~ 10 mV) 0.1 Hz ~ 3 kHz (10 mV ~ 20 mV) 0.1 Hz ~ 3 kHz (20 mV ~ 30 mV) 0.1 Hz ~ 3 kHz (30 mV ~ 50 mV) 0.1 Hz ~ 3 kHz (50 mV ~ 100 mV) 0.1 Hz ~ 3 kHz	8.4×10^{-4} 4.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 8.8×10^{-5} 4.8×10^{-5} 3.6×10^{-5} 3.0×10^{-5} 2.6×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters (Digital Smapling)	40419			Calibrator/ SICT-CP-40419
AC Voltage		(100 mV ~ 200 mV) 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 mV ~ 300 mV) 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(300 mV ~ 500 mV) 0.1 Hz ~ 3 kHz	2.8×10^{-5}	
		(500 mV ~ 1 V) 0.1 Hz ~ 3 kHz	2.4×10^{-5}	
		(1 V ~ 2 V) 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(2 V ~ 3 V) 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(3 V ~ 5 V) 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(5 V ~ 10 V) 0.1 Hz ~ 3 kHz	2.6×10^{-5}	
		(10 V ~ 30 V) 10 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 V ~ 50 V) 10 Hz ~ 3 kHz	2.8×10^{-5}	
		(50 V ~ 100 V) 10 Hz ~ 3 kHz	2.4×10^{-5}	
		(100 V ~ 200 V) 10 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 V ~ 1 000 V) 50 Hz ~ 1 kHz	2.4×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters	40420			Calibrator/ SICT-CP-40420
AC Voltage Test		(600 μ V) 1 kHz	7.8×10^{-3}	
		(600 μ V ~ 20 mV) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	2.6 $\times 10^{-3}$ 2.5 $\times 10^{-3}$ 3.5 $\times 10^{-3}$ 7.0 $\times 10^{-3}$ 1.5 $\times 10^{-2}$	
		(20 ~ 200) mV (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz	5.2 $\times 10^{-4}$ 3.4 $\times 10^{-4}$ 8.8 $\times 10^{-4}$ 4.3 $\times 10^{-3}$	
		(200 mV ~ 2 V) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.5 $\times 10^{-4}$ 1.5 $\times 10^{-4}$ 4.6 $\times 10^{-4}$ 8.8 $\times 10^{-4}$ 3.4 $\times 10^{-3}$	
		(2 ~ 20) V (10 ~ 40) Hz 40 Hz ~ 100 kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.0 $\times 10^{-4}$ 1.2 $\times 10^{-4}$ 4.4 $\times 10^{-4}$ 2.2 $\times 10^{-3}$	
		(20 ~ 200) V (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz	3.0 $\times 10^{-4}$ 1.2 $\times 10^{-4}$ 1.8 $\times 10^{-4}$	
		(200 ~ 500) V 50 Hz ~ 1 kHz	3.8×10^{-4}	
		(500 ~ 1 000) V 50 Hz ~ 1 kHz	3.7×10^{-4}	
		(25 ~ 500) mV (1 ~ 30) MHz	2.1×10^{-2}	
		(500 mV ~ 2 V) (0.1 ~ 30) MHz	2.1×10^{-2}	
Weighting Test		(DIN/NOISE) 31.5 Hz ~ 10 kHz (JIS A) 31.5 Hz ~ 16 kHz (CCIR) 31.5 Hz ~ 31.5 kHz (CCIR/ARM) 31.5 Hz ~ 31.5 kHz	0.12 dB 0.12 dB 0.12 dB 0.12 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters	40420			Calibrator/ SICT-CP-40420
AC Voltage Output		(10 mV) 1 kHz (10 mV ~ 1 V) 20 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz	2.8×10^{-3} 8.5×10^{-5} 1.3×10^{-4} 6.9×10^{-4}	
DC Voltage Output		0 mV 100 mV ~ 1 V	0.99 μ V 1.1×10^{-5}	
Oscilloscopes	40421			Calibration Generator/ SICT-CP-40421
Impedance Measure		50 Ω 75 Ω 1 M Ω	3.5×10^{-5} 2.7×10^{-5} 2.5×10^{-5}	
DC Voltage		(\pm) 0 mV (0 ~ 1) mV (1 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 900) mV (0.9 ~ 9) V (9 ~ 200) V	0.79 μ V 8.0×10^{-4} 4.1×10^{-4} 1.7×10^{-4} 8.5×10^{-5} 1.5×10^{-5} 9.1×10^{-6} 9.5×10^{-6}	
AC Voltage(Square wave)		(1 kHz) 1 mV (1 ~ 25) mV (0.025 ~ 0.5) V (0.5 ~ 2.2) V (2.2 ~ 130) V	6.5×10^{-3} 8.8×10^{-4} 9.1×10^{-4} 6.8×10^{-4} 8.4×10^{-4}	
Time Marker		100 ps (100 ~ 200) ps 200 ps ~ 20 ms 20 ms ~ 5 s	6.2×10^{-7} 3.1×10^{-7} 1.7×10^{-6} 1.6×10^{-5}	
CAL Output Amplitude		(40 Hz ~ 20 kHz) 100 mV 100 mV ~ 12 V	3.2×10^{-5} 1.9×10^{-5}	
CAL Output Frequency		100 Hz ~ 10 MHz	6.2×10^{-7}	
Sinewave Signal Generator Level		50 kHz 50 kHz ~ 1 MHz 1 MHz ~ 1 GHz (1 ~ 4) GHz (4 ~ 18) GHz (18 ~ 25) GHz (25 ~ 33) GHz (33 ~ 40) GHz	2.3×10^{-2} 4.7×10^{-2} 1.5×10^{-2} 1.8×10^{-2} 3.2×10^{-2} 5.5×10^{-2} 5.8×10^{-2} 6.0×10^{-2}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes	40421	(10 ~ 40) Hz		Calibration Generator/ SICT-CP-40421
AC Voltage(Sine wave)		2 mV	5.3 μ V	
		(2 ~ 20) mV	1.9×10^{-3}	
		(20 ~ 200) mV	6.8×10^{-4}	
		(0.2 ~ 20) V	4.3×10^{-4}	
		(20 ~ 100) V	4.5×10^{-4}	
		(40 Hz ~ 1 kHz)		
		2 mV	5.0 μ V	
		(2 ~ 20) mV	1.7×10^{-3}	
		(20 ~ 800) mV	3.3×10^{-4}	
		(0.8 ~ 20) V	1.7×10^{-4}	
		(20 ~ 200) V	1.8×10^{-4}	
		(1 ~ 50) kHz		
		2 mV	5.2 μ V	
		(2 ~ 20) mV	1.9×10^{-3}	
		(20 ~ 200) mV	4.8×10^{-4}	
		(0.2 ~ 20) V	1.3×10^{-4}	
		(20 ~ 100) V	1.7×10^{-4}	
		(50 ~ 100) kHz		
		2 mV	7.1 μ V	
		(2 ~ 20) mV	2.7×10^{-3}	
		(20 ~ 200) mV	9.0×10^{-4}	
		(0.2 ~ 20) V	1.9×10^{-4}	
		(20 ~ 100) V	3.0×10^{-4}	
LF phase meters	40422	(1 Hz ~ 200 kHz)		Multi Function Generator/ SICT-CP-40422
Phase Test		(-180 ~ 180) °	0.074 °	
Volt/Current recorders	40424	(±)		Calibrator/ SICT-CP-40424
DC Voltage		(0 ~ 100) μ V	0.51 μ V	
		(0.1 ~ 1) mV	5.2×10^{-4}	
		(1 ~ 10) mV	5.9×10^{-5}	
		(0.01 ~ 1) V	6.7×10^{-6}	
		(1 ~ 10) V	4.3×10^{-6}	
		(10 ~ 100) V	6.3×10^{-6}	
		(100 ~ 1 000) V	8.7×10^{-6}	
DC Current		(±)		
		(0 ~ 1) nA	6.9 pA	
		(1 ~ 100) nA	4.6×10^{-3}	
		(0.1 ~ 1) μ A	2.3×10^{-3}	
		(1 ~ 10) μ A	7.2×10^{-4}	
		(10 ~ 100) μ A	1.4×10^{-4}	
		(0.1 ~ 10) mA	7.6×10^{-5}	
		(10 ~ 100) mA	8.4×10^{-5}	
		(0.1 ~ 1) A	1.2×10^{-4}	
		(1 ~ 100) A	2.1×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relay test sets	40425			Digital Multimeter/ SICT-CP-40425
DC Voltage		1 mV 1 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	6 μ V 7.0×10^{-4} 7.0×10^{-5} 2.2×10^{-5}	
DC Current		1 mA 1 mA ~ 1 A (1 ~ 20) A (20 ~ 100) A	58 μ A 6.0×10^{-4} 2.5×10^{-4} 4.0×10^{-4}	
AC Voltage		(1 mV) 20 Hz ~ 100 kHz (1 ~ 100) mV 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 10) V 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 100) V 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 1 000) V 50 Hz ~ 10 kHz (10 ~ 30) kHz	58 μ V 6.1×10^{-4} 1.2×10^{-3} 7.6×10^{-4} 1.3×10^{-3} 2.0×10^{-2} 2.2×10^{-4} 1.1×10^{-3} 2.1×10^{-2} 1.8×10^{-4} 1.1×10^{-3} 1.9×10^{-4} 4.5×10^{-4}	
AC Current		(1 mA) 40 Hz ~ 10 kHz (1 ~ 100) mA 40 Hz ~ 10 kHz (100 mA ~ 1 A) 40 Hz ~ 10 kHz (1 ~ 10) A 40 Hz ~ 10 kHz (10 ~ 100) A 40 Hz ~ 10 kHz	58 μ A 8.6×10^{-4} 9.1×10^{-4} 9.9×10^{-4} 2.3×10^{-4}	
Timer		(1 ~ 100) s	5.8×10^{-6}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF signal generators	40426			Audio Analyzer, Digital Multimeter/ SICT-CP-40426
Frequency Test		(0.1 ~ 1) Hz 1 Hz ~ 100 MHz	5.8×10^{-7} 5.8×10^{-9}	
Output Level Test		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		(100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5}	
		(10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 8.0×10^{-5}	
DC Offset		(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 20 V	0.7 μ V 0.7 μ V 1.0×10^{-4} 6.0×10^{-5}	
Output Level Flatness Test		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB	
		(100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.005 4 dB 0.001 1 dB 0.007 2 dB	
		(1 ~ 30) V (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.021 dB 0.015 dB 0.027 dB	
Attenuator Test		(10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	
Rise/Fall Time		100 μ s ~ 100 ns (100 ~ 10) ns (10 ~ 1) ns 1 ns ~ 100 ps	7.0×10^{-4} 7.8×10^{-4} 4.7×10^{-3} 4.6×10^{-2}	
Duty cycle		(1 ~ 99) %	0.006 1 %	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function
Reference Frequency		10 MHz	7.7×10^{-12}	Generator/ SICT-CP-40427
Center Frequency		10 Hz (10 ~ 100) Hz 100 Hz ~ 1 kHz 1 kHz ~ 1 MHz (1 ~ 100) MHz 100 MHz ~ 1 GHz	6.1×10^{-5} 6.1×10^{-6} 6.1×10^{-7} 6.1×10^{-8} 6.1×10^{-9} 6.1×10^{-8}	
Frequency Range		10 Hz (10 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 100) MHz 100 MHz ~ 1 GHz	1.1×10^{-3} 1.1×10^{-4} 1.1×10^{-5} 1.1×10^{-4} 1.1×10^{-6} 1.1×10^{-7} 1.1×10^{-9}	
Resolution Bandwidth		100 Hz 100 Hz ~ 3 kHz (3 ~ 300) kHz 300 kHz ~ 1 MHz	3.3×10^{-2} 3.1×10^{-2} 3.3×10^{-2} 3.4×10^{-2}	
Absolute Amplitude		(-60 dBV) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-60 ~ -50) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-50 ~ -40) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-40 ~ -30) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-30 ~ 30) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	0.043 dB 0.042 dB 0.056 dB 0.017 dB 0.015 dB 0.022 dB 0.009 dB 0.009 8 dB 0.012 dB 0.009 dB 0.006 8 dB 0.011 dB 0.016 dB 0.006 3 dB 0.007 3 dB	
Referency Level		(-60 dBV) 10 Hz ~ 100 kHz (-60 ~ 30) dBV 10 Hz ~ 100 kHz	0.17 dB 0.16 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function Generator/ SICT-CP-40427
Frequency Response		10 Hz 10 Hz ~ 100 kHz	0.009 1 dB 0.008 7 dB	
Logscale Fidelity		(0 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (-80 ~ -90) dB	0.009 2 dB 0.012 dB 0.016 dB 0.042 dB	
Output frequency		10 Hz ~ 300 MHz	6.1×10^{-11}	
Input Impedance		1 MΩ (50 ~ 75) Ω	0.000 12 MΩ 0.000 7 Ω	
Output Voltage		10 mV 10 mV ~ 5 V	0.000 38 mV 8.8×10^{-5}	
Output Offset Voltage		(-30 ~ 30) V	6.7×10^{-6}	
Output Voltage Flatness		10 Hz ~ 100 kHz	0.000 67 dB	
Spot generators	40428			Audio Analyzer, Digital Multimeter/ SICT-CP-40428
Frequency		(0.1 ~ 1) Hz 1 Hz ~ 100 kHz	5.8×10^{-7} 5.8×10^{-9}	
Output Level		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		(100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5}	
		(10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
Output Level Flatness		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB	
		(100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.005 4 dB 0.001 1 dB 0.007 2 dB	
		(1 ~ 10) V (10 ~ 100) Hz 100 Hz ~ 100 kHz	0.010 dB 0.011 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spot generators	40428			Audio Analyzer, Digital Multimeter/ SICT-CP-40428
Attenuation		(10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	
Sweep generators	40429			Audio Analyzer, Digital Multimeter/ SICT-CP-40429
Frequency		(0.1 ~ 1) Hz 1 Hz ~ 100 kHz	5.8×10^{-7} 5.8×10^{-9}	
Output Level Test		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 100 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		(100 Hz ~ 10 kHz) 1 mV (1 ~ 100) mV 100 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
Output Level Flatness		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB	
		(100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.005 4 dB 0.001 1 dB 0.007 2 dB	
		(1 ~ 10) V (10 ~ 100) Hz 100 Hz ~ 100 kHz	0.010 dB 0.011 dB	
Attenuation		(10 Hz ~ 10 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Signal transducers	Voltage	40430	1 mV $(1 \sim 10)$ mV 10 mV ~ 100 V $(100 \sim 300)$ V	5.2×10^{-4} 8.8×10^{-5} 3.1×10^{-5} 4.3×10^{-3}
			10 μ A $(10 \sim 100)$ μ A 100 μ A ~ 100 mA 100 mA ~ 20 A	9.5×10^{-4} 9.7×10^{-5} 7.0×10^{-5} 2.0×10^{-4}
			(1 ~ 10) Hz 10 Hz ~ 100 kHz	3.1×10^{-4} 7.2×10^{-5}
	Current			
	Frequency			
Transistor curve tracers	DC Voltage(Source)	40432	0 mV $(0 \sim 100)$ mV 100 mV ~ 1 V $(1 \sim 10)$ V $(10 \sim 100)$ V $(100 \sim 1\,000)$ V	1.0μ V 1.1×10^{-5} 1.0×10^{-5} 9.8×10^{-6} 1.1×10^{-5} 8.0×10^{-6}
	DC Current(Source)		0 nA $(0 \sim 1)$ nA $(1 \sim 100)$ nA $(0.1 \sim 1)$ μ A $(1 \sim 10)$ μ A 10 μ A ~ 10 mA $(10 \sim 100)$ mA 100 mA ~ 10 A	0.12 nA 1.0×10^{-2} 8.0×10^{-3} 8.0×10^{-4} 9.0×10^{-5} 1.6×10^{-5} 5.0×10^{-5} 2.5×10^{-4}
AC/DC high voltage generators	DC Voltage	40434	(\pm) 0 kV $(0 \sim 0.5)$ kV $(0.5 \sim 1)$ kV $(1 \sim 2)$ kV $(2 \sim 100)$ kV $(100 \sim 200)$ kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 3.0×10^{-4} 2.3×10^{-4} 1.2×10^{-2}
	AC Voltage		(50 Hz) 0.01 kV $(0.01 \sim 0.5)$ kV $(0.5 \sim 1)$ kV $(1 \sim 100)$ kV $(100 \sim 200)$ kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 5.7×10^{-4} 1.2×10^{-2}
			(60 Hz) 0.01 kV $(0.01 \sim 0.5)$ kV $(0.5 \sim 1)$ kV $(1 \sim 100)$ kV $(100 \sim 200)$ kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.2×10^{-4} 1.2×10^{-2}

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC/DC high voltage probes	40435			
DC Voltage		(±) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV	0.06 V 4.2×10^{-4} 2.1×10^{-4} 3.5×10^{-4}	DC Power Supply/ SICT-CP-40435
AC Voltage		(50 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV	2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4}	
		(60 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV	2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4}	
Logic analyzers	40436			Calibrator/ SICT-CP-40436
DC Voltage		(0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	1.5 μ V 8.8×10^{-6} 7.2×10^{-6}	
Clock frequency		10 MHz	7.7×10^{-12}	
Telephone testers	40437			Tone Pulse Simulator/ SICT-CP-40437
L1, L2 Output Voltage		(1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 1 000) V	9.9×10^{-4} 1.1×10^{-5} 7.1×10^{-6} 8.5×10^{-6}	
Loop Current		(0.1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A	2.5×10^{-5} 5.2×10^{-5} 2.2×10^{-4} 4.7×10^{-4}	
Ring Output Voltage		(10 Hz ~ 20 kHz) 100 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	4.7×10^{-4} 2.9×10^{-4} 3.1×10^{-4}	
Ring Frequency		(1 ~ 1 000) Hz	7.0×10^{-5}	
D.T.M.F & Pulse		(+10 ~ -39.9) dBm	0.09 dB	
D.T.M.F & Frequency		(697 ~ 1 477) Hz	0.59 Hz	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal analyzers				
Color Bar Decoding Accuracy(Gain)	40438	(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV (500 ~ 1 000) mV	2.5×10^{-1} 5.0×10^{-2} 8.2×10^{-2} 4.1×10^{-2} 2.1×10^{-2} 8.5×10^{-3}	Video Amplitude Calibration Fixture/ SICT-CP-40438
Frequency		20 Hz ~ 5 MHz	5.8×10^{-6}	
Color Bar Decoding Accuracy(Phase)		(0 ~ 360)°	0.70°	
Measure Square Wave		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 300) mV (300 ~ 400) mV (400 ~ 600) mV (600 ~ 999.9) mV	9.4×10^{-2} 2.0×10^{-2} 9.9×10^{-3} 2.1×10^{-3} 1.5×10^{-3} 1.2×10^{-3} 9.4×10^{-4}	
Measure Sine Wave		No Filter , PAL NTS BW Lim, NTSC,PAL Chroma BP, NTSC,PAL (10 kHz ~ 10 MHz) 500 mV	7.0×10^{-3}	
Burst Frequency		(3 ~ 5) MHz	4.0×10^{-7}	
vertical Gain		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 300) mV (300 ~ 600) mV (600 ~ 999.9) mV	9.4×10^{-2} 2.0×10^{-2} 9.9×10^{-3} 2.1×10^{-3} 1.5×10^{-3} 9.4×10^{-4}	
Horizontal Frequency		(20 ~ 100) Hz 100 Hz ~ 10 kHz 10 kHz ~ 10 MHz	3.1×10^{-3} 6.1×10^{-4} 6.1×10^{-5}	
Gain Frequency Response		Flat, Luminance, Chroma at (20 Hz ~ 20 MHz) 700 mV	7.0×10^{-3}	
Transient Response		(0 ~ 1 000) mV	1.3×10^{-2}	
(Video Noise)				
Luminance Volt Level		(0 ~ -30) dB	4.8×10^{-1}	
Chrominance AM/PM Level		(0 ~ -30) dB	6.7×10^{-1}	
Luminance Volt Level		(0 ~ 1 000) mV	1.7×10^{-5}	
Luminance Inputt Level		(0 ~ 1 000) mV	1.8×10^{-5}	
Chrominance Input Level		(0 ~ 1 000) mV	1.7×10^{-5}	

405. Low frequency electric & magnetic fields

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Flux meters	40503			Flux sources/ SICT-CP-40503
Flux sources	40504	0.1 mWb ~ 10 Wb (0.1 ~ 50) mWb (0.05 ~ 0.1) Wb (0.1 ~ 10) Wb	5.8 × 10 ⁻³ 6.6 × 10 ⁻⁴ 2.3 × 10 ⁻⁵ 1.4 × 10 ⁻⁵	GPS receiver, Frequency counter/ SICT-CP-40504
Magnetometers	40508	(0 ~ 0.1) mT (0.1 ~ 0.5) mT (0.5 ~ 3) mT (3 ~ 5) mT (5 ~ 20) mT (20 ~ 30) mT (30 ~ 1 700) mT	7.1 × 10 ⁻² 1.4 × 10 ⁻² 7.0 × 10 ⁻³ 4.0 × 10 ⁻³ 3.0 × 10 ⁻³ 6.7 × 10 ⁻³ 6.4 × 10 ⁻³	Helmholtz coil, Standard magnets/ SICT-CP-40508
Reference/standard magnets	40510	Gauss (1.5 ~ 30) mT (30 ~ 1 000) mT	7.3 × 10 ⁻³ 2.6 × 10 ⁻³	Gaussmeters/ SICT-CP-40510

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF amplifiers	40601	10 Hz ~ 10 kHz (0 ~ 80) dB	0.07 dB	Power Sensor, Attenuator/ SICT-CP-40601
		10 kHz ~ 10 GHz (0 ~ 40) dB (40 ~ 80) dB	0.08 dB 0.11 dB	
		(10 ~ 18) GHz (0 ~ 40) dB (40 ~ 80) dB	0.10 dB 0.13 dB	
		(18 ~ 30) GHz (0 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB	0.20 dB 0.31 dB 0.75 dB	
		(30 ~ 40) GHz (0 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB	0.26 dB 0.35 dB 0.76 dB	
		(40 ~ 50) GHz (0 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB	0.41 dB 0.47 dB 0.82 dB	
		(50 ~ 67) GHz (0 ~ 20) dB (20 ~ 45) dB	0.36 dB 0.45 dB	
		(9 kHz ~ 40 GHz) (0 ~ 100) dBc	0.90 dB	
Reflection coefficient		(0 ~ 1) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	4.7×10^{-3} 9.2×10^{-3} 1.5×10^{-2} 1.9×10^{-2} 3.3×10^{-2}	
SWR		(1 ~ ∞) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	9.5×10^{-3} 1.9×10^{-2} 3.1×10^{-2} 3.9×10^{-2} 6.7×10^{-2}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial attenuators	40602	(10 Hz ~ 1 kHz) (0 ~ 40) dB (40 ~ 50) dB (50 ~ 70) dB (1 kHz ~ 9 kHz) (0 ~ 40) dB (40 ~ 50) dB (50 ~ 70) dB (9 kHz ~ 26.5 GHz) (0 ~ 10) dB (10 ~ 30) dB (30 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 110) dB (110 ~ 120) dB (26.5 ~ 45) GHz (0 ~ 20) dB (20 ~ 70) dB (45 ~ 67) GHz (0 ~ 10) dB (10 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB	0.063 dB 0.070 dB 0.12 dB 0.063 dB 0.068 dB 0.11 dB 0.064 dB 0.069 dB 0.073 dB 0.077 dB 0.081 dB 0.086 dB 0.090 dB 0.095 dB 0.11 dB 0.12 dB 0.21 dB 0.24 dB 0.23 dB 0.25 dB 0.27 dB 0.32 dB	Power Sensor, Directional Coupler/ SICT-CP-40602
Attenuation		(0 ~ 1) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 7 0.009 2 0.015 0.019 0.033	
Reflection coefficient		(1 ~ ∞) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 5 0.019 0.031 0.039 0.067	
SWR				

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Burst pulse generators	40605	50 Ω (±) 5 V (5 ~ 100) V (0.1 ~ 8) kV	2.0 × 10 ⁻² 1.6 × 10 ⁻² 2.0 × 10 ⁻²	Digital Oscilloscope/ SICT-CP-40605	
		1 kΩ (±) 5 V ~ 8 kV	4.0 × 10 ⁻²		
		1 ns (1 ~ 2) ns (2 ~ 4) ns 4 ns ~ 1 μs (1 ~ 2) μs (2 ~ 4) μs (4 ~ 10) μs	2.0 × 10 ⁻² 6.8 × 10 ⁻³ 2.6 × 10 ⁻³ 1.5 × 10 ⁻³ 6.2 × 10 ⁻³ 2.6 × 10 ⁻³ 1.3 × 10 ⁻³		
		Pulse Width	6.0 × 10 ⁻³ 3.1 × 10 ⁻³ 1.5 × 10 ⁻³		
		Time measurement by section	6.0 × 10 ⁻³ 3.1 × 10 ⁻³ 1.5 × 10 ⁻³ 1.2 × 10 ⁻³		
		Repeat Frequency	1.6 × 10 ⁻³		
Attenuator calibrators	40606	Attenuation	(0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (30 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (100 ~ 110) dB (110 ~ 120) dB	0.024 dB 0.025 dB 0.027 dB 0.029 dB 0.031 dB 0.034 dB 0.036 dB 0.039 dB 0.042 dB 0.045 dB 0.048 dB 0.052 dB	Verification Kit/ SICT-CP-40606
RF power meter calibrators	40607	Output Power	3 μW 10 μW 30 μW 100 μW 300 μW 1 mW 3 mW 10 mW 30 mW 100 mW	0.1 nW 0.2 nW 0.5 nW 1 nW 4 nW 0.18 μW 0.19 μW 0.2 μW 0.3 μW 1 μW	Digital Multimeter/ SICT-CP-40607

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC transducers ; current probes, absorbing clamps, etc.	40608			Power Senso, Network analyzer/ SICT-CP-40608
Transfer Impedance		10 Hz ~ 50 MHz (50 ~ 200) MHz 200 MHz ~ 3 GHz	0.60 dB 1.1 dB 1.9 dB	
Insertion Loss		30 MHz ~ 1 GHz	1.9 dB	
Electric Magnetic Near-Field		100 kHz ~ 1 GHz	1.9 dB	
Reflection coefficient		(0 ~ 1) 10 Hz ~ 1 GHz (1 ~ 3) GHz	4.2×10^{-3} 6.0×10^{-3}	
SWR		(1 ~ ∞) 10 Hz ~ 1 GHz (1 ~ 3) GHz	9.0×10^{-3} 1.3×10^{-2}	
Coaxial directional couplers/ splitters	40610			Power Sensor, Synthesized Sweeper/ SICT-CP-40610
Coupling Factor		(10 Hz ~ 10 kHz) (0 ~ 40) dB (40 ~ 50) dB (50 ~ 70) dB	0.063 dB 0.070 dB 0.12 dB	
		(10 kHz ~ 100 kHz) (0 ~ 30) dB (30 ~ 50) dB (50 ~ 70) dB	0.064 dB 0.087 dB 0.098 dB	
		(100 kHz ~ 15 GHz) (0 ~ 40) dB (40 ~ 50) dB (50 ~ 70) dB	0.090 dB 0.093 dB 0.11 dB	
		(15 GHz ~ 18 GHz) (0 ~ 70) dB	0.12 dB	
		(18 GHz ~ 26.5 GHz) (0 ~ 70) dB	0.17 dB	
		(26.5 GHz ~ 45 GHz) (0 ~ 20) dB (20 ~ 70) dB	0.21 dB 0.24 dB	
		(45 GHz ~ 67 GHz) (0 ~ 10) dB (10 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB	0.23 dB 0.25 dB 0.27 dB 0.32 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial directional couplers/ splitters	40610			Power Sensor, Synthesized Sweeper/
Reflection coefficient		(0 ~ 1) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 7 0.009 2 0.015 0.019 0.033	
SWR		(1 ~ ∞) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 5 0.019 0.031 0.039 0.067	
Electrostatic discharge generators	40613			Digital Oscilloscope/ SICT-CP-40613
Peak Current(I_p)		(\pm) (3.75 ~ 7.5) A (7.5 ~ 15) A (15 ~ 22.5) A (22.5 ~ 56.3) A (56.3 ~ 93.8) A (93.8 ~ 112.5) A (112.5 ~ 150) A	5.1×10^{-2} 5.3×10^{-2} 4.6×10^{-2} 5.2×10^{-2} 4.9×10^{-2} 5.7×10^{-2} 5.2×10^{-2}	
Current I_1 (30 ~ 60) ns		(\pm) 2 A (2 ~ 4) A (4 ~ 8) A (8 ~ 16) A (16 ~ 36) A (36 ~ 50) A (50 ~ 60) A (60 ~ 80) A	4.5×10^{-2} 5.0×10^{-2} 5.3×10^{-2} 4.9×10^{-2} 5.0×10^{-2} 4.4×10^{-2} 5.7×10^{-2} 5.2×10^{-2}	
Current I_2 (60 ~ 130) ns		(\pm) 1 A (1 ~ 2) A (2 ~ 4) A (4 ~ 6) A (6 ~ 8) A (8 ~ 15) A (15 ~ 25) A (25 ~ 30) A (30 ~ 40) A	5.0×10^{-2} 5.4×10^{-2} 5.7×10^{-2} 4.9×10^{-2} 5.4×10^{-2} 6.5×10^{-2} 5.2×10^{-2} 6.7×10^{-2} 6.1×10^{-2}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrostatic discharge generators	40613			Digital Oscilloscope/ SICT-CP-40613
Current I3 (360 ~ 800) ns		(±) 0.275 A (0.275 ~ 0.55) A (0.55 ~ 1.1) A (1.1 ~ 1.65) A (1.65 ~ 4.13) A (4.13 ~ 6.88) A (6.88 ~ 8.25) A (8.25 ~ 11) A	1.5×10^{-1} 2.2×10^{-1} 1.9×10^{-1} 1.6×10^{-1} 2.3×10^{-1} 1.5×10^{-1} 2.4×10^{-1} 1.9×10^{-1}	
Current I4 (30 ~ 65) ns		(±) 0.15 A (0.15 ~ 0.3) A (0.3 ~ 0.6) A (0.6 ~ 1.2) A (1.2 ~ 2.25) A (2.25 ~ 2.7) A (2.7 ~ 3.75) A (3.75 ~ 4.5) A (4.5 ~ 6) A	1.3×10^{-1} 1.9×10^{-1} 3.0×10^{-1} 2.1×10^{-1} 2.5×10^{-1} 2.0×10^{-1} 1.5×10^{-1} 2.9×10^{-1} 2.0×10^{-1}	
Semiconductor Peak Current HBM		(±) (20 ~ 83.3) mA (0.083 3 ~ 1.33) A (1.33 ~ 6.66) A	3.1×10^{-2} 3.6×10^{-2} 2.7×10^{-2}	
Semiconductor Peak Current MM		(±) (0.219 ~ 14) A (14 ~ 35) A	3.5×10^{-2} 3.1×10^{-2}	
Time		0.1 ns 0.1 ns ~ 1 ms	2.7×10^{-2} 2.4×10^{-2}	
Peak Voltage		(±) 1 kV (1 ~ 35) kV	3.0×10^{-2} 2.5×10^{-2}	
EMC receivers	40614			Network Analyzer , Pulse Generator/ SICT-CP-40614
Reference accuracy		100 kHz ~ 100 MHz	6.1×10^{-10}	
Reflection coefficient		(0 ~ 1) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.004 6 0.009 2 0.015 0.018	
SWR		(1 ~ ∞) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.009 2 0.018 0.029 0.037	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC receivers	40614			
Response to Sine-wave Voltage		10 Hz ~ 5 kHz	0.01 dB	
		5 kHz ~ 2 GHz	0.11 dB	
		(2 ~ 18) GHz	0.15 dB	
		(18 ~ 26.5) GHz	0.20 dB	
		(26.5 ~ 40) GHz	0.26 dB	
		(40 ~ 50) GHz	0.29 dB	
Variation with Repetition Frequency relative calibration		9 kHz ~ 1 GHz	0.88 dB	
Amplitude relationship absolute calibration		9 kHz ~ 1 GHz	0.88 dB	
Overall pass-band selectivity		9 kHz ~ 1 GHz	0.14 dB	
Intermediate frequency rejection ratio		9 kHz ~ 40 GHz	0.14 dB	
Image frequency rejection ratio		9 kHz ~ 40 GHz	0.14 dB	
Other spurious responses		9 kHz ~ 40 GHz	0.14 dB	
Random Noise		DC ~ 40 GHz	0.32 dB	
IF bandwidth accuracy		(-6 dB) 1 Hz ~ 20 MHz	3.0×10^{-3}	
		(-60 dB) 1 Hz ~ 20 MHz	3.3×10^{-4}	
IF bandwidth selectivity		1 Hz ~ 20 MHz	3.0×10^{-3}	
RF filters	40615			
Reject Frequency		(9 ~ 90) kHz (90 ~ 900) kHz 900 kHz ~ 900 MHz 900 MHz ~ 18 GHz (18 ~ 50) GHz	0.024 kHz 0.24 kHz 0.025 MHz 0.068 MHz 0.12 MHz	
Insertion Loss		(9 kHz ~ 8 GHz) (0 ~ 10) dB (10 ~ 20) dB (20 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (8 ~ 18) GHz (0 ~ 10) dB (10 ~ 30) dB (30 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB	0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.18 dB 0.23 dB 0.66 dB 1.7 dB 4.3 dB 0.23 dB 0.24 dB 0.25 dB 0.26 dB 0.31 dB 0.73 dB 1.7 dB 4.3 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF filters	40615	(18 ~ 50) GHz (0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (20 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB	0.48 dB 0.51 dB 0.52 dB 0.53 dB 0.54 dB 0.59 dB 0.78 dB 1.6 dB 2.8 dB 6.0 dB	Network Analyzer/ SICT-CP-40615
RF impedance meters	40616	(100 kHz ~ 18 GHz) (35 ~ 20) dBm (20 Hz ~ 18 GHz) (20 ~ -70) dBm Frequency 9 kHz ~ 0.1 MHz 0.1 MHz ~ 18 GHz Load Measurement DC 10 Hz ~ 100 MHz (100 ~ 500) MHz 500 MHz ~ 1.8 GHz (1.8 ~ 3.0) GHz (3.0 ~ 18) GHz	0.11 dB 0.12 dB 6.8 × 10 ⁻¹⁰ 6.2 × 10 ⁻¹¹ 0.02 Ω 0.06 Ω 0.15 Ω 0.21 Ω 0.41 Ω 1.1 Ω	Performance Kit/ SICT-CP-40616
RF impulse generators	40617	9 kHz ~ 1 GHz	0.33 dB	Digital Oscilloscope/ SICT-CP-40617
Line impedance stabilization networks; LISN, CDN, ISN, etc.	40618	Impedance 9 kHz ~ 1 GHz Phase 9 kHz ~ 1 GHz Insertion Loss 9 kHz ~ 30 MHz 30 MHz ~ 1 GHz Decoupling Attenuation Isolation (0 ~ 20) dB 9 kHz ~ 1 GHz (20 ~ 50) dB 9 kHz ~ 1 GHz (50 ~ 60) dB 9 kHz ~ 1 GHz (60 ~ 70) dB 9 kHz ~ 1 MHz 1 MHz ~ 1 GHz	2.0 × 10 ⁻² 1.2 ° 0.08 dB 0.09 dB 0.13 dB 0.18 dB 0.22 dB 0.33 dB 0.29 dB	Impedance/Gain-Phase Analyzer, Calibration Kit/ SICT-CP-40618

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Line impedance stabilization networks; LISN, CDN, ISN, etc.	40618			Impedance/Gain-Phase Analyzer, Calibration Kit/
Decoupling Attenuation Isolation		(70 ~ 80) dB 9 kHz ~ 1 MHz 1 MHz ~ 1 GHz	0.70 dB 0.48 dB	
		(80 ~ 90) dB 9 kHz ~ 1 MHz 1 MHz ~ 1 GHz	1.8 dB 1.0 dB	
		(90 ~ 100) dB 9 kHz ~ 1 MHz 1 MHz ~ 1 GHz	4.5 dB 2.5 dB	
(Coupling Decoupling network)				
Impedance		9 kHz ~ 1 GHz	2.0×10^{-2}	
Insertion loss		9 kHz ~ 1 GHz	0.11 dB	
Reflection coefficient		(0 ~ 1) 9 kHz ~ 1 GHz	0.004 6	
SWR		(1 ~ ∞) 9 kHz ~ 1 GHz	0.009 2	
Coaxial standard mismatches	40619			Network Analyzer, Calibration Kit/ SICT-CP-40619-1
(Coaxial standard mismatches)				
Reflection coefficient		(0 ~ 1) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 7 0.009 2 0.015 0.019 0.033	
SWR		(1 ~ ∞) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 5 0.019 0.031 0.039 0.067	
(Calibration kit)				SICT-CP-40619-2
Magnitude of reflection coefficient		(Termination) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 7 0.009 2 0.015 0.019 0.033	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial standard mismatches (Calibration kit)	40619			Network Analyzer, Calibration Kit/ SICT-CP-40619-1
Magnitude of reflection coefficient		(Short, Open circuit) 9 kHz ~ 45 MHz 45 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.020 0.016 0.026 0.044 0.056 0.066	
	SWR	(1 ~ 1.01) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 5 0.019 0.031 0.039 0.067	
		(1.01 ~ 1.05) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.011 0.020 0.032 0.040 0.070	
		(1.05 ~ 1.2) 9 kHz ~ 45 MHz 45 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.015 0.014 0.024 0.039 0.049 0.083	
		(1.2 ~ 1.5) 9 kHz ~ 45 MHz 45 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.023 0.021 0.035 0.054 0.066 0.12	
		(1.5 ~ 2) 9 kHz ~ 45 MHz 45 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.040 0.036 0.055 0.086 0.11 0.18	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial standard mismatches	40619	(2 ~ 3) 9 kHz ~ 45 MHz 45 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.096 0.076 0.12 0.20 0.24 0.35	Network Analyzer, Calibration Kit/ SICT-CP-40619-1
Phase of reflection coefficient	SWR	(3 ~ 9) 9 kHz ~ 45 MHz 45 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.80 0.70 1.1 1.8 2.2 2.8	
Mobile communication test sets	40621	(Termination) 9 kHz ~ 67 GHz	180°	
Frequency Modulation		CW : 100 kHz ~ 50 GHz 0.1 Hz ~ 5 MHz	1.2×10^{-2}	Measuring Receiver, RF Signal Generator/ SICT-CP-40621
Amplitude Modulation		CW : 100 kHz ~ 50 GHz (0.1 ~ 100) %	1.2×10^{-2}	
Phase Modulation		CW : 100 kHz ~ 50 GHz (0.1 ~ 1 000) rad	1.2×10^{-2}	
Distortion Harmonics of Modulation Rate		(0 ~ 20) %	2.4×10^{-2}	
RF Output Level (High range level)		(35 ~ 20) dBm 100 kHz ~ 8 GHz (8 ~ 10) GHz (10 ~ 12) GHz (12 ~ 18) GHz	0.10 dB 0.13 dB 0.14 dB 0.15 dB	
		(20 ~ -30) dBm 9 kHz ~ 1 GHz (1 ~ 14) GHz (14 ~ 18) GHz (18 ~ 26) GHz (26 ~ 34) GHz (34 ~ 42) GHz (42 ~ 48) GHz (48 ~ 50) GHz	0.06 dB 0.07 dB 0.08 dB 0.10 dB 0.12 dB 0.14 dB 0.15 dB 0.19 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621	(-30 ~ -60) dBm		Measuring Receiver, RF Signal Generator/ SICT-CP-40621
RF Output Level (High range level)		9 kHz ~ 1 GHz	0.06 dB	
		(1 ~ 14) GHz	0.07 dB	
		(14 ~ 18) GHz	0.09 dB	
		(18 ~ 30) GHz	0.11 dB	
		(30 ~ 38) GHz	0.13 dB	
		(38 ~ 46) GHz	0.19 dB	
		(46 ~ 50) GHz	0.21 dB	
RF Output Level (Low range Absolute TRFL)		(9 kHz ~ 4.2 GHz)		
		30 dBm ~ 0 dBm	0.14 dB	
		0 dBm ~ -40 dBm	0.14 dB	
		-40 dBm ~ -80 dBm	0.16 dB	
		-80 dBm ~ -120 dBm	0.18 dB	
		(4.2 GHz ~ 8 GHz)		
		30 dBm ~ 0 dBm	0.19 dB	
		0 dBm ~ -40 dBm	0.17 dB	
		-40 dBm ~ -80 dBm	0.19 dB	
		-80 dBm ~ -120 dBm	0.21 dB	
		(8 GHz ~ 12.4 GHz)		
		30 dBm ~ 0 dBm	0.21 dB	
		0 dBm ~ -40 dBm	0.19 dB	
		-40 dBm ~ -80 dBm	0.21 dB	
		-80 dBm ~ -120 dBm	0.23 dB	
		(12.4 GHz ~ 18 GHz)		
		30 dBm ~ 0 dBm	0.21 dB	
		0 dBm ~ -40 dBm	0.22 dB	
		-40 dBm ~ -80 dBm	0.24 dB	
		-80 dBm ~ -120 dBm	0.26 dB	
		(18 GHz ~ 26.5 GHz)		
		30 dBm ~ 0 dBm	0.28 dB	
		0 dBm ~ -40 dBm	0.29 dB	
		-40 dBm ~ -80 dBm	0.30 dB	
		-80 dBm ~ -120 dBm	0.32 dB	
Harmonic		(0 ~ -100) dBc	0.38 dB	
Frequency Output Accuracy		9 kHz ~ 50 GHz	6.1×10^{-10}	
AC Output Level		(10 Hz ~ 100 kHz)		
		(1 ~ 10) mV	2.6×10^{-3}	
		(10 ~ 100) mV	6.3×10^{-4}	
		(0.1 ~ 10) V	6.5×10^{-4}	
		(10 ~ 100) V	6.6×10^{-4}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	DC Output Level	(1 ~ 10) mV	4.1×10^{-4}	Measuring Receiver, RF Signal Generator/ SICT-CP-40621
		(10 ~ 100) mV	4.1×10^{-5}	
		(0.1 ~ 1) V	6.8×10^{-6}	
		(1 ~ 10) V	3.6×10^{-6}	
		(10 ~ 100) V	5.3×10^{-6}	
	AC Input Level	(10 Hz ~ 100 kHz)		
		(1 ~ 10) mV	7.2×10^{-3}	
		(10 ~ 100) mV	1.9×10^{-3}	
		(0.1 ~ 1) V	1.2×10^{-3}	
		(1 ~ 10) V	4.0×10^{-4}	
	DC Input Level	(10 ~ 100) V	4.3×10^{-4}	
		(1 ~ 10) mV	1.5×10^{-3}	
		(10 ~ 100) mV	1.5×10^{-4}	
		(0.1 ~ 10) V	6.4×10^{-5}	
	RF Input Level	(10 ~ 100) V	8.0×10^{-5}	
		(9 kHz ~ 18 GHz)		
		(10 ~ -70) dBm	0.14 dB	
		(18 ~ 50) GHz		
		(10 ~ -70) dBm	0.29 dB	
Reflection coefficient	SWR	(0 ~ 1)		
		9 kHz ~ 2 GHz	0.004 6	
		(2 ~ 20) GHz	0.009 2	
		(20 ~ 40) GHz	0.015	
		(40 ~ 50) GHz	0.018	
	SWR	(1 ~ ∞)		
		9 kHz ~ 2 GHz	0.009 2	
		(2 ~ 20) GHz	0.018	
		(20 ~ 40) GHz	0.029	
		(40 ~ 50) GHz	0.037	
Modulation meters	Amplitude Modulation	40622		Measuring Receiver/ SICT-CP-40622
		0 kHz	0.01 %	
		(0 ~ 400) kHz	1.2×10^{-2}	
	Frequency Modulation	0 %	1 Hz	
		(0 ~ 100) %	1.2×10^{-2}	
	Phase Modulation	0 rad	1.2 mrad	
		(0 ~ 400) rad	1.2×10^{-2}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers	40623			Power Sensor, Verification Kit/ SICT-CP-40623
Frequency Output Accuracy		5 Hz ~ 60 GHz	6.1×10^{-10}	
RF Output Level Accuracy		(20 ~ -30) dBm		
		5 Hz ~ 100 Hz	0.01 dB	
		100 Hz ~ 1 GHz	0.06 dB	
		(1 ~ 13) GHz	0.07 dB	
		(13 ~ 18) GHz	0.08 dB	
		(18 ~ 26.5) GHz	0.10 dB	
		(26.5 ~ 30) GHz	0.11 dB	
		(30 ~ 45) GHz	0.15 dB	
		(45 ~ 55) GHz	0.23 dB	
		(55 ~ 75) GHz	0.30 dB	
		(75 ~ 100) GHz	0.35 dB	
		(100 ~ 110) GHz	0.40 dB	
		(-30 ~ -60) dBm		
		9 kHz ~ 1 GHz	0.06 dB	
		(1 ~ 18) GHz	0.08 dB	
		(18 ~ 26.5) GHz	0.10 dB	
		(26.5 ~ 30) GHz	0.11 dB	
		(30 ~ 40) GHz	0.15 dB	
		(40 ~ 50) GHz	0.21 dB	
Dynamic Accuracy		(100 kHz ~ 18 GHz)		
		(0 ~ 10) dB	0.091 dB	
		(10 ~ 20) dB	0.093 dB	
		(20 ~ 40) dB	0.095 dB	
		(40 ~ 60) dB	0.12 dB	
		(60 ~ 80) dB	0.14 dB	
		(80 ~ 90) dB	0.15 dB	
		(90 ~ 100) dB	0.16 dB	
Attenuation Accuracy		(20 dB)		
		300 kHz ~ 45 MHz	0.051 dB	
		45 MHz ~ 2 GHz	0.053 dB	
		(2 ~ 7.5) GHz	0.056 dB	
		(7.5~ 26.5) GHz	0.057 dB	
		(40 dB)		
		300 kHz ~ 45 MHz	0.055 dB	
		45 MHz ~ 6 GHz	0.060 dB	
		(6 ~ 15) GHz	0.064 dB	
		(15 ~ 26.5) GHz	0.065 dB	
Phase		(±180 °)		
		300 kHz ~ 45 MHz	0.31°	
		45 MHz ~ 9.0 GHz	0.26°	
		(9.0 ~ 15.0) GHz	0.28°	
		(15.0 ~ 20.0) GHz	0.35°	
		(20.0 ~ 26.5) GHz	0.42°	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers	40623	(30 kHz ~ 2 GHz) 1.05 1.20 1.50 2.00 (2 ~ 18) GHz 1.05 1.20 1.50 2.00	0.012 0.012 0.013 0.019 0.019 0.019 0.019 0.023	Power Sensor, Verification Kit/ SICT-CP-40623
Noise figure meters	40624	10 MHz ~ 26.5 GHz 10 MHz 9 kHz ~ 1 GHz (1 ~ 20) GHz (20 ~ 26.5) GHz DC voltage Range Noise Figure	4.3×10^{-5} 6.1×10^{-10} 0.008 0.019 0.03 0.000 18 V 0.052 dB 0.12 dB 0.16 dB 0.37 dB	Noise Source/ SICT-CP-40624
Noise generators	40625	(-80 ~ -130) dBm/Hz (0 ~ 50) dB	0.10 dB 0.27 dB	Spectrum Analyzer/ SICT-CP-40625
Noise impulse simulators	40626	Peak Voltage (\pm) 0.1 kV (0.1 ~ 5) kV Rise/Fall Time 1 ns (1 ~ 2) ns (2 ~ 4) ns Pulse Width 10 ns (10 ~ 1 000) ns	4.0×10^{-2} 3.5×10^{-2} 6.0×10^{-3} 3.1×10^{-3} 1.5×10^{-3} 2.0×10^{-3} 1.5×10^{-3}	Digital Oscilloscope/ SICT-CP-40626
RF phase noise meters	40627	RF phase noise (Carrier Frequency) 100 MHz ~ 18 GHz (Offset Frequency) 10 Hz ~ 100 MHz	1.0 dB 1.0 dB	RF Signal analyzer/ SICT-CP-40627

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial noise sources	40628	(4.5 ~ 16) dB (0.01 ~ 1) GHz (1 ~ 7) GHz (7 ~ 8) GHz (8 ~ 14) GHz (14 ~ 18) GHz (12 ~ 17) dB (0.01 ~ 1) GHz (1 ~ 2) GHz (2 ~ 6) GHz (6 ~ 7) GHz (7 ~ 12) GHz (12 ~ 18) GHz (18 ~ 26.5) GHz	0.28 dB 0.27 dB 0.30 dB 0.31 dB 0.32 dB 0.31 dB 0.28 dB 0.30 dB 0.29 dB 0.40 dB 0.41 dB 0.47 dB	Coaxial noise sources, Noise figure analyzer/ SICT-CP-40628
	ENR	(0 ~1) (0.01 ~ 3) GHz (3 ~ 20) GHz (20 ~ 26.5) GHz	0.006 8 0.010 0.015	
RF power meters	40635	(0.1 ~ 500) W 10 kHz ~ 250 MHz (0.1 ~ 150) W (80 ~ 1 000) MHz (0.1 ~ 10) W (1 000 ~ 4 200) MHz 10 μW ~ 1 mW (1 ~ 100) mW 3 μW ~ 100 mW (88 ~ 100) % 50 MHz, 1 mW	2.6×10^{-2} 2.6×10^{-2} 2.7×10^{-2} 3 nW 0.01 mW 1.6×10^{-3} 0.5×10^{-3} 8 μW	Range Calibrator/ SICT-CP-40635
	High power			
	Zero Carryover			
	Power			
	Calibration Factor			
	Power Ref. Output			

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Diode power sensors	40636			Thermistor Mount , Synthesized Sweeper / SICT-CP-40636
Calibration Factor		(1 ~ 10) μ W 9 kHz ~ 100 kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 50) GHz (10 μ W ~ 10 mW) 9 kHz ~ 100 kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	7.0 \times 10 ⁻³ 1.1 \times 10 ⁻² 1.2 \times 10 ⁻² 1.9 \times 10 ⁻² 2.8 \times 10 ⁻² 3.4 \times 10 ⁻² 4.1 \times 10 ⁻² 4.0 \times 10 ⁻³ 9.0 \times 10 ⁻³ 1.1 \times 10 ⁻² 1.9 \times 10 ⁻² 2.6 \times 10 ⁻² 3.1 \times 10 ⁻² 3.7 \times 10 ⁻² 4.2 \times 10 ⁻²	
Reflection coefficient		(0 ~ 1) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 6 0.009 2 0.015 0.018 0.033	
SWR		(1 ~ ∞) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 2 0.018 0.029 0.037 0.065	
Thermocouple power sensors	40637			Thermistor Mount , Synthesized Sweeper / SICT-CP-40637
Calibration Factor		(1 ~ 10) μ W 9 kHz ~ 100 kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 50) GHz (10 μ W ~ 10 mW) 9 kHz ~ 100 kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	7.0 \times 10 ⁻³ 1.1 \times 10 ⁻² 1.2 \times 10 ⁻² 1.9 \times 10 ⁻² 2.8 \times 10 ⁻² 3.4 \times 10 ⁻² 4.1 \times 10 ⁻² 4.0 \times 10 ⁻³ 9.0 \times 10 ⁻³ 1.1 \times 10 ⁻² 1.9 \times 10 ⁻² 2.6 \times 10 ⁻² 3.1 \times 10 ⁻² 3.7 \times 10 ⁻² 4.2 \times 10 ⁻²	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Thermocouple power sensors	40637			
Reflection coefficient		(0 ~ 1) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 6 0.009 2 0.015 0.018 0.033	Thermistor Mount, Synthesized Sweeper/ SICT-CP-40637
SWR		(1 ~ ∞) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 2 0.018 0.029 0.037 0.065	
Pulse generators	40638			Digital Oscilloscope/ SICT-CP-40638
Period		100 ps ~ 1 s	5.8×10^{-9}	
Frequency		1 Hz ~ 1 GHz (1 ~ 3.35) GHz	5.8×10^{-9} 1.7×10^{-8}	
Width		100 ps ~ 1 s	1.2×10^{-3}	
Delay Time		100 ps ~ 1 s	1.2×10^{-3}	
Double Pulse		100 ps ~ 1 s	1.2×10^{-3}	
Duty Cycle		(1 ~ 99) %	0.006 2 %	
DC Level		$\pm(10 \text{ mV} \sim 100 \text{ V})$	5.8×10^{-4}	
Output Level		(100 Hz ~ 10 kHz) (10 ~ -20) dBm	0.018 dB	
Radar test sets	40639			Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
(Ship)				
RF Level		(20 ~ -20) dBm 20 Hz ~ 1 GHz (1 ~ 18) GHz (-20 ~ -60) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (-60 ~ -120) dBm 10 MHz ~ 18 GHz	0.09 dB 0.13 dB 0.06 dB 0.11 dB 0.25 dB	
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Frequency Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets	40639			
Distortion of Modulation		(0 ~ 2) %	1.2×10^{-3}	Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
Harmonics		(9 kHz ~ 18 GHz) (0 ~ -110) dB	0.25 dB	
Frequency		9 kHz ~ 18 GHz	6.2×10^{-11}	
Pulse Period		1 ns ~ 10 ms	1.2×10^{-2}	
High power		(0.1 ~ 500) W 10 kHz ~ 250 MHz	2.6×10^{-2}	
		(0.1 ~ 150) W (80 ~ 1 000) MHz	2.6×10^{-2}	
		(0.1 ~ 10) W (1 000 ~ 4 200) MHz	2.7×10^{-2}	
(flight)				
Frequency(VOR/ILS/DME)		(74.6 ~ 1 150) MHz	8.2×10^{-8}	
Amplitude Modulation(VOR/ILS)				
Localizer				
		(108.1 ~ 111.95) MHz		
		(0.1 ~ 20) %	0.62 %	
Glideslope				
		(330.95 ~ 334.70) MHz		
		(20 ~ 40) %	0.84 %	
Marker Beacon				
		(74.6 ~ 75.4) MHz		
		(40 ~ 95) %	1.4 %	
VOR				
		(108 ~ 117.95) MHz		
		(0.1 ~ 30) %	0.62 %	
고주파 태벨(VOR/ILS)				
Localizer				
		(108.1 ~ 111.95) MHz		
		(10 ~ -30) dBm	0.19 dB	
		(-30 ~ -50) dBm	0.23 dB	
		(-50 ~ -70) dBm	0.24 dB	
		(-70 ~ -110) dBm	0.34 dB	
		(-110 ~ -120) dBm	0.35 dB	
Glideslope				
		(330.95 ~ 334.70) MHz		
		(10 ~ -30) dBm	0.19 dB	
		(-30 ~ -50) dBm	0.23 dB	
		(-50 ~ -70) dBm	0.24 dB	
		(-70 ~ -110) dBm	0.34 dB	
		(-110 ~ -120) dBm	0.35 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets	40639	Localizer (108.1 ~ 111.95) MHz	0.000 5	Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
		LEFT(-0.200 ~ -0.155)	0.000 5	
		LEFT(-0.155 ~ -0.093)	0.000 5	
		CENTER(0.000)	0.000 5	
		RIGHT(0.093 ~ 0.155)	0.000 5	
		RIGHT(0.155 ~ 0.200)	0.000 5	
		Glideslope (330.95 ~ 334.70) MHz	0.000 5	
		DOWN(0.400 ~ 0.175)	0.000 5	
		DOWN(0.175 ~ 0.091)	0.000 5	
		CENTER(0.000)	0.000 5	
SDM(VOR/ILS)	40639	UP(-0.091 ~ -0.175)	0.000 5	
		UP(-0.175 ~ -0.400)	0.000 5	
		Localizer (108.1 ~ 111.95) MHz (0 ~ 40) %	0.87 %	
		Glideslope (330.95 ~ 334.70) MHz (40 ~ 80) %	1.2 %	
BEARING ANGLE(VOR)		VOR (108 ~ 117.95) MHz (0 ~ 360) °	0.04 °	
RF signal generators	40640	CW : 100 kHz ~ 50 GHz Rate : 10 Hz ~ 100 kHz 0.1 Hz ~ 5 MHz	1.2 × 10 ⁻²	Measuring Receiver / SICT-CP-40640
		CW : 100 kHz ~ 50 GHz Rate : 10 Hz ~ 100 kHz (0.1 ~ 100) %	1.2 × 10 ⁻²	
		CW : 100 kHz ~ 50 GHz Rate : 10 Hz ~ 100 kHz (0.1 ~ 1 000) rad	1.2 × 10 ⁻²	
		(0 ~ 20) %	2.4 × 10 ⁻²	
		(57 ~ 54) dBm 100 kHz ~ 500 MHz	0.14 dB	
		(54 ~ 51) dBm 100 kHz ~ 500 MHz	0.12 dB	
		500 MHz ~ 2 GHz	0.18 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF signal generators	40640			Measuring Receiver/ SICT-CP-40640
RF Output Level (High range level)		(51 ~ 35) dBm		
		100 kHz ~ 1 GHz	0.13 dB	
		(1 ~ 8) GHz	0.14 dB	
		(8 ~ 12) GHz	0.16 dB	
		(12 ~ 18) GHz	0.17 dB	
		(35 ~ 20) dBm		
		100 kHz ~ 8 GHz	0.10 dB	
		(8 ~ 10) GHz	0.12 dB	
		(10 ~ 12) GHz	0.13 dB	
		(12 ~ 18) GHz	0.15 dB	
		(20 ~ -30) dBm		
		20 Hz ~ 1 GHz	0.06 dB	
		(1 ~ 14) GHz	0.07 dB	
		(14 ~ 18) GHz	0.08 dB	
		(18 ~ 28) GHz	0.11 dB	
		(28 ~ 38) GHz	0.13 dB	
		(38 ~ 48) GHz	0.15 dB	
		(48 ~ 50) GHz	0.19 dB	
		(50 ~ 55) GHz	0.25 dB	
		(55 ~ 65) GHz	0.29 dB	
		(65 ~ 75) GHz	0.30 dB	
		(75 ~ 80) GHz	0.34 dB	
		(80 ~ 100) GHz	0.36 dB	
		(100 ~ 110) GHz	0.40 dB	
		(-30 ~ -60) dBm		
		9 kHz ~ 1 GHz	0.06 dB	
		(1 ~ 14) GHz	0.07 dB	
		(14 ~ 18) GHz	0.08 dB	
		(18 ~ 28) GHz	0.11 dB	
		(28 ~ 35) GHz	0.13 dB	
		(35 ~ 40) GHz	0.15 dB	
		(40 ~ 50) GHz	0.21 dB	
RF Output Level (Low range Absolute TRFL)		(9 kHz ~ 4.2 GHz)		
		30 dBm ~ 0 dBm	0.14 dB	
		0 dBm ~ -40 dBm	0.14 dB	
		-40 dBm ~ -80 dBm	0.16 dB	
		-80 dBm ~ -120 dBm	0.18 dB	
		(4.2 GHz ~ 8 GHz)		
		30 dBm ~ 0 dBm	0.19 dB	
		0 dBm ~ -40 dBm	0.17 dB	
		-40 dBm ~ -80 dBm	0.19 dB	
		-80 dBm ~ -120 dBm	0.21 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF signal generators	40640			Measuring Receiver/ SICT-CP-40640
RF Output Level (Low range Absolute TRFL)		(8 GHz ~ 12.4 GHz) 30 dBm ~ 0 dBm 0 dBm ~ -40 dBm -40 dBm ~ -80 dBm -80 dBm ~ -120 dBm (12.4 GHz ~ 18 GHz) 30 dBm ~ 0 dBm 0 dBm ~ -40 dBm -40 dBm ~ -80 dBm -80 dBm ~ -120 dBm (18 GHz ~ 26.5 GHz) 30 dBm ~ 0 dBm 0 dBm ~ -40 dBm -40 dBm ~ -80 dBm -80 dBm ~ -120 dBm	0.21 dB 0.19 dB 0.21 dB 0.23 dB 0.21 dB 0.22 dB 0.24 dB 0.26 dB 0.28 dB 0.29 dB 0.30 dB 0.32 dB	
Harmonic		(-10 ~ -110) dBc	0.38 dB	
Output Frequency		20 Hz ~ 60 GHz	6.1×10^{-10}	
Pulse Modulation		(100 kHz ~ 12 000 MHz) Period (1 μs ~ 1 s) ton (100 ns ~ 1000 μs) PRR (1 Hz ~ 1 MHz)	2.0×10^{-3} 2.0×10^{-3} 2.0×10^{-3}	
RF spectrum analyzers	40641			Power Sensor, Synthesized Sweeper/ SICT-CP-40641
Center frequency readout accuracy		3 Hz ~ 100 kHz 100 kHz ~ 10 MHz 10 MHz ~ 67 GHz	0.61 mHz 61 mHz 0.61 kHz	
Frequency count function accuracy		3 Hz ~ 100 kHz 100 kHz ~ 10 MHz 10 MHz ~ 67 GHz	0.61 mHz 61 mHz 0.61 kHz	
Frequency span accuracy		10 Hz ~ 10 kHz 10 kHz ~ 67 GHz	6.1×10^{-4} 6.1×10^{-5}	
Resolution bandwidth accuracy		1 Hz ~ 100 MHz	6.1×10^{-4}	
Resolution bandwidth selectivity		1 Hz ~ 100 MHz	3.2×10^{-3}	
Resolution bandwidth switching accuracy		1 Hz ~ 100 MHz	0.023 dB	
Scale switching accuracy		1 dB ~ 10 dB	0.023 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF spectrum analyzers	40641			
Scale fidelity accuracy		(0 ~ -10) dB (-10 ~ -30) dB (-30 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (-80 ~ -90) dB (-90 ~ -100) dB	0.085 dB 0.087 dB 0.089 dB 0.092 dB 0.10 dB 0.11 dB 0.13 dB 0.14 dB 0.15 dB	Power Sensor, Synthesized Sweeper/ SICT-CP-40641
Frequency response accuracy		(5 ~ 100) Hz 100 Hz ~ 100 kHz 100 kHz ~ 8 GHz (8 ~ 15) GHz (15 ~ 18) GHz (18 ~ 25) GHz (25 ~ 26.5) GHz (26.5 ~ 30) GHz (30 ~ 45) GHz (45 ~ 67) GHz	0.01 dB 0.10 dB 0.11 dB 0.12 dB 0.15 dB 0.20 dB 0.21 dB 0.24 dB 0.26 dB 0.32 dB	
Average noise accuracy		DC ~ 50 GHz	0.15 dB	
Sideband noise accuracy		(-30 ~ 30) kHz	0.15 dB	
Calibrator output frequency accuracy		DC ~ 1 GHz	6.1×10^{-9}	
Calibrator output amplitude accuracy		(20 ~ -20) dBm	0.06 dB	
RF speed guns	40642			Signal Generator/ SICT-CP-40642
Surge generators	40643			Digital Oscilloscope/ SICT-CP-40643
(Surge generator)				
Surge Voltage		(±) 5 mV 5 mV ~ 2 V (2 ~ 5) V 5 V ~ 200 kV	3.2×10^{-2} 3.0×10^{-2} 4.0×10^{-2} 3.5×10^{-2}	
Surge Current		(±) 5 A 5 A ~ 200 kA	3.3×10^{-2} 3.6×10^{-2}	
Rise/Fall Time		1 µs (1 ~ 4) µs 4 µs ~ 10 s	2.4×10^{-3} 4.6×10^{-3} 2.9×10^{-3}	
Pulse Width		1 µs 1 µs ~ 10 s	1.2×10^{-3} 2.0×10^{-3}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Surge generators Time measurement by section	40643	1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}	Digital Oscilloscope/ SICT-CP-40643
Pulse Period		1 Hz ~ 25 MHz	1.6×10^{-3}	
Phase Shifting		at 50 Hz (0 ~ 360) $^{\circ}$	1.2°	
		at 60 Hz (0 ~ 360) $^{\circ}$	1.4°	
(Partial Discharge Calibrator) Charge Pulse Voltage		(\pm) 5 mV 5 mV ~ 50 V	3.2×10^{-2} 3.0×10^{-2}	
SWR meters	40644	9 kHz ~ 18 GHz 30 kHz ~ 100 MHz 100 MHz ~ 10 GHz (10 ~ 18) GHz SWR (30 kHz ~ 30 MHz) 1.05 1.20 1.50 2.00 (30 MHz ~ 2 GHz) 1.05 1.20 1.50 2.00 (2 ~ 18) GHz 1.05 1.20 1.50 2.00	6.4×10^{-5} 0.06 dB 0.08 dB 0.09 dB 0.019 0.019 0.019 0.020 0.021 0.021 0.021 0.021 0.018 0.018 0.018 0.024	Coaxial Mismatch/ SICT-CP-40644

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations (Open, Short, Phase)	40645	(± 180 °) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (Reflection coefficient) (SWR) (Impedance)	0.49° 0.61° 0.95° 1.2° 0.004 3 0.006 0 0.009 5 0.016 0.019 0.008 6 0.012 0.019 0.032 0.038 0.64 Ω 1.0 Ω 1.6 Ω 2.0 Ω 0.71 Ω 1.1 Ω 1.8 Ω 2.2 Ω 0.84 Ω 1.3 Ω 2.1 Ω 2.7 Ω 0.99 Ω 1.6 Ω 2.5 Ω 3.1 Ω 1.1 Ω 1.8 Ω 2.9 Ω 3.6 Ω	Network Analyzer, Coaxial Mismatch/ SICT-CP-40645

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations (Impedance)	40645	(0.285 7 ~ 0.333 4) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.000 0 ~ 0.047 6, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.047 6 ~ 0.090 9, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.090 9 ~ 0.166 7, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.166 7 ~ 0.230 8, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.230 8 ~ 0.285 7, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.285 7 ~ 0.333 4, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	1.3 Ω 2.1 Ω 3.3 Ω 4.2 Ω 0.68° 1.1° 1.7° 2.2° 0.67° 1.1° 1.7° 2.1° 0.66° 1.1° 1.7° 2.1° 0.65° 1.0° 1.6° 2.1° 0.64° 1.0° 1.6° 2.0° 0.62° 1.0° 1.6° 2.0°	Network Analyzer, Coaxial Mismatch/ SICT-CP-40645

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial thermistor mounts	40646	(1 ~ 10) μ W	0.4×10^{-2}	Thermistor Mount, Synthesized Sweeper/ SICT-CP-40646
		(9 ~ 100) kHz	0.9×10^{-2}	
		100 kHz ~ 1 GHz	1.1×10^{-2}	
		(1 ~ 10) GHz	1.9×10^{-2}	
		(10 ~ 18) GHz	2.8×10^{-2}	
		(18 ~ 26.5) GHz	4.0×10^{-2}	
		(26.5 ~ 40) GHz		
		(10 μ W ~ 10 mW)	0.4×10^{-2}	
		(9 ~ 100) kHz	0.9×10^{-2}	
		100 kHz ~ 1 GHz	1.1×10^{-2}	
Reflection coefficient	SWR	(1 ~ 10) GHz	1.9×10^{-2}	
		(10 ~ 18) GHz	2.6×10^{-2}	
		(18 ~ 26.5) GHz	3.5×10^{-2}	
		(26.5 ~ 40) GHz		
		(0 ~ 1)	4.2×10^{-3}	
		20 Hz ~ 1 GHz	9.4×10^{-3}	
		(1 ~ 20) GHz	1.5×10^{-2}	
		(20 ~ 40) GHz		
RF voltmeters	40650	(1 ~ ∞)	9.7×10^{-3}	RF Millivolt Meter Calibrator/ SICT-CP-40650
		20 Hz ~ 1 GHz	2.4×10^{-2}	
		(1 ~ 20) GHz	3.8×10^{-2}	
		(20 ~ 40) GHz		
		3 V	4.2 mV	
		1 V	1.4 mV	
		300 mV	0.36 mV	
		270 mV	0.32 mV	
		240 mV	0.28 mV	
		210 mV	0.26 mV	
		180 mV	0.24 mV	
		150 mV	0.22 mV	
		120 mV	0.17 mV	
		100 mV	0.16 mV	
Vector voltmeters	40651	90 mV	0.13 mV	Signal Generator/ SICT-CP-40651
		60 mV	0.11 mV	
		30 mV	0.048 mV	
		10 mV	0.020 mV	
		3 mV	0.018 mV	
		1 mV	0.013 mV	
		3 V	4.2 mV	
		1 V	1.4 mV	
		300 mV	0.36 mV	
RF Phase		100 mV	0.16 mV	
		30 mV	0.048 mV	
		10 mV	0.046 mV	
		3 mV	0.022 mV	
		1 mV	0.024 mV	
		(0 ~ 270) $^{\circ}$	0.006 $^{\circ}$	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Field strength meters	40652			Signal Generator/
Center frequency		(9 ~ 100) kHz 0.1 MHz ~ 18 GHz	6.8×10^{-8} 6.2×10^{-9}	
Scale Fidelity		(0 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -100) dB	0.11 dB 0.12 dB 0.13 dB 0.18 dB	SICT-CP-40652
Frequency response		9 kHz ~ 500 MHz 500 MHz ~ 18 GHz	0.05 dB 0.08 dB	
AM/FM test sources	40653			Measuring Receiver/ SICT-CP-40653
Output frequency		(10 ~ 560) MHz	6.2×10^{-10}	
Dip simulators	40654			Digital Oscilloscope/ SICT-CP-40654
DC Voltage		1 V (1 ~ 5) V (5 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 500) V	1.1×10^{-5} 6.4×10^{-6} 5.4×10^{-6} 9.3×10^{-6} 8.0×10^{-6} 1.1×10^{-5}	
AC Voltage		(50 ~ 60) Hz 50 V (50 ~ 100) V (100 ~ 500) V	5.0×10^{-5} 2.5×10^{-5} 1.6×10^{-4}	
Frequency		50 Hz 60 Hz	8.4×10^{-6} 8.3×10^{-6}	
Dip DC Voltage		(0 ~ 50) V 0 % (0 ~ 120) %	0.2 V 3.4×10^{-2}	
Dip AC Voltage		(50 Hz ~ 60 Hz, 0 V ~ 400 V) 0 % (0 ~ 120) %	0.9 V 3.4×10^{-2}	
Time measurement by section		100 ns ~ 2 μ s (2 ~ 4) μ s (4 ~ 400) μ s (0.4 ~ 2) ms 2 ms ~ 5 s	1.8×10^{-3} 2.0×10^{-3} 1.6×10^{-3} 2.0×10^{-3} 1.6×10^{-3}	
Inrush Current		(5 ~ 1 000) A	3.6×10^{-2}	
Phase Shifting		at 50 Hz (0 ~ 360) $^{\circ}$ at 60 Hz (0 ~ 360) $^{\circ}$	1.2 $^{\circ}$ 1.4 $^{\circ}$	

407. Field strength & antenna

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Probes	40702	9 kHz ~ 300 MHz (1 ~ 750) V/m	0.12	RF Power Meter, Multi Meter/ SICT-CP-40702
Electric Field Probe		300 MHz ~ 400 MHz (1 ~ 300) V/m	0.12	
		400 MHz ~ 1 GHz (1 ~ 150) V/m	0.12	
		1 GHz ~ 6 GHz (1 ~ 150) V/m	0.12	
		6 GHz ~ 10 GHz (1 ~ 100) V/m	0.13	
		10 GHz ~ 18 GHz (2 ~ 100) V/m	0.13	
		18 GHz ~ 40 GHz (2 ~ 100) V/m	0.15	
Magnetic Field Probe		10 Hz ~ 1 kHz (0.39 ~ 1 000) A/m	0.07	
		1 kHz ~ 10 kHz (0.39 ~ 600) A/m	0.07	
		10 kHz ~ 30 kHz (2.65 ~ 390) mA/m (0.39 ~ 100) A/m	0.12 0.07	
		30 kHz ~ 150 kHz (2.65 ~ 390) mA/m (0.39 ~ 20) A/m	0.12 0.07	
		150 kHz ~ 200 kHz (2.65 ~ 390) mA/m (0.39 ~ 10) A/m	0.12 0.07	
		200 kHz ~ 300 MHz 2.65 mA/m ~ 1.98 A/m	0.12	
		300 MHz ~ 400 MHz 2.65 mA/m ~ 0.79 A/m	0.12	
		400 MHz ~ 1 GHz 2.65 mA/m ~ 0.39 A/m	0.12	

407. Field strength & antenna

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dipole antennas (Dipole Antenna) Antenna Factor	40703	20 MHz ~ 18 GHz (1 ~ ∞) 20 MHz ~ 1 GHz 1 GHz ~ 18 GHz	1.1 dB 0.022 0.039	Network analyzer/ SICT-CP-40703 ancillary facilities(Chungju)
Voltage Standing Wave Ratio		700 MHz ~ 18 GHz	1.4 dB	
Antenna Pattern		20 MHz ~ 18 GHz	1.2 dB	
(Biconical Antenna) Antenna Factor		(1 ~ ∞) 20 MHz ~ 1 GHz 1 GHz ~ 18 GHz	0.022 0.039	
Voltage Standing Wave Ratio		700 MHz ~ 18 GHz	1.4 dB	
Antenna Pattern		20 MHz ~ 18 GHz	1.2 dB	
(Log Periodic Antenna) Antenna Factor		(1 ~ ∞) 20 MHz ~ 1 GHz 1 GHz ~ 18 GHz	0.022 0.039	
Voltage Standing Wave Ratio		700 MHz ~ 18 GHz	1.4 dB	
Antenna Pattern		20 MHz ~ 18 GHz	1.4 dB	
Loop antennas Antenna Factor	40704	10 Hz ~ 30 MHz	1.2 dB	Signal generator1, Signal analyzer/ SICT-CP-40704
Monopole antennas Antenna Factor	40705	10 Hz ~ 30 MHz	1.4 dB	Signal generator1, Signal analyzer/ SICT-CP-40705
Horn antennas Antenna Factor	40707	200 MHz ~ 18 GHz 18 GHz ~ 40 GHz	1.1 dB 1.4 dB	Network analyzer/ SICT-CP-40707 ancillary facilities(Chungju)
Voltage Standing Wave Ratio		(1 ~ ∞) 200 MHz ~ 1 GHz 1 GHz ~ 18 GHz 18 GHz ~ 40 GHz	0.022 0.039 0.041	
Antenna Pattern		700 MHz ~ 18 GHz	1.4 dB	

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	50101	0 °C (-196 ~ -95) °C (-95 ~ -90) °C (-90 ~ 250) °C (250 ~ 550) °C (550 ~ 660) °C (660 ~ 1 100) °C (1 100 ~ 1 600) °C	0.010 °C 0.060 °C 0.030 °C 0.017 °C 0.020 °C 0.060 °C 0.7 °C 1.7 °C	SPRT, STANDARD TC/ SICT-CP-50101
Temperature indicators/recorders /controllers, temperature calibrators (Temperature indicators/recorders/controllers)	50102	With Sensor (-196 ~ 500) °C (500 ~ 660) °C (660 ~ 700) °C (700 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 400) °C (1 400 ~ 1 600) °C Without Sensor (-196 ~ 0) °C (0 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 1 300) °C (1 300 ~ 1 600) °C (temperature calibrators) Output (-196 ~ 500) °C (500 ~ 600) °C (600 ~ 800) °C (800 ~ 1 300) °C (1 300 ~ 1 600) °C Input (-196 ~ 500) °C (500 ~ 600) °C (600 ~ 800) °C (800 ~ 1 300) °C (1 300 ~ 1 600) °C	0.020 °C 0.045 °C 0.59 °C 0.60 °C 0.61 °C 2.0 °C 2.1 °C 0.010 °C 0.013 °C 0.018 °C 0.022 °C 0.025 °C 0.029 °C 0.033 °C 0.040 °C 0.044 °C 0.07 °C 0.09 °C 0.005 °C 0.006 °C 0.007 °C 0.08 °C 0.10 °C 0.03 °C 0.04 °C 0.05 °C 0.07 °C 0.09 °C	SPRT, STANDARD TC/ SICT-CP-50102
Glass thermometers: liquid-in-glass, Beckmann liquid-in-glass	50103	(-90 ~ -58) °C (-58 ~ 400) °C (400 ~ 500) °C	0.15 °C 0.04 °C 0.15 °C	SPRT/ SICT-CP-50103

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance thermometers; SPRT, IPRT, thermistors, etc.	50104	(-196 ~ 100) °C (100 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C (Fixed point) Ar TP Hg TP H ₂ O TP Ga MP In FP Sn FP Zn FP Al FP Ag FP	0.019 °C 0.020 °C 0.021 °C 0.044 °C -189.344 2 °C -38.834 4 °C 0.01 °C 29.764 6 °C 156.598 5 °C 231.928 °C 419.527 °C 660.323 °C 961.78 °C	SPRT, Fixed point/ SICT-CP-50104
Thermal expansion thermometers; bimetal, gas or liquid type	50105	bimetal (-196 ~ -70) °C (-70 ~ 100) °C (100 ~ 200) °C (200 ~ 500) °C (500 ~ 650) °C	0.6 °C 0.2 °C 0.3 °C 0.6 °C 1.2 °C	SPRT/ SICT-CP-50105
Thermocouples: noble metal, base metal, pure metal, special type, etc.	50106	Base metal (0 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C Noble metal (-196 ~ -90) °C (-90 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C (660 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 300) °C (Fixed point) H ₂ O ICE Point Sn FP Zn FP Al FP Ag FP Cu FP Co-C MP Fe MP	0.5 °C 1.7 °C 1.8 °C 0.4 °C 0.2 °C 0.3 °C 0.4 °C 1.1 °C 1.4 °C 1.8 °C 0.2 °C 0.2 °C 0.2 °C 0.3 °C 0.3 °C 1.1 °C 1.6 °C	SPRT, Fixed point, STANDARD TC/ SICT-CP-50106

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature transducers	50107	(-196 ~ 400) °C (400 ~ 500) °C (500 ~ 660) °C (660 ~ 800) °C (800 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C	0.031 °C 0.043 °C 0.072 °C 0.6 °C 0.7 °C 2.1 °C 2.2 °C	SPRT, THERMOCOUPLE, MULTIMETER SICT-CP-50107
Primary fixed-point cells and apparatus H ₂ O TP	50108	0.01 °C	0.24 mK	Triple-Point Cell SICT-CP-50108

502. non contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical pyrometers	50203	(900 ~ 1 800) °C	5 °C	Standard Lamp/ SICT-CP-50203
Standard radiation thermometers	50204	(-40 ~ -20) °C (-20 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 900) °C (900 ~ 1 200) °C (1 200 ~ 1 400) °C (1 400 ~ 1 500) °C (1 500 ~ 1 600) °C (1 600 ~ 1 800) °C (1 800 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.9 °C 0.7 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 1.0 °C 1.1 °C 1.3 °C 1.6 °C 1.7 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 2.5 °C 2.6 °C 4.2 °C 4.4 °C 4.6 °C 4.7 °C	Transfer Standard Pyrometer/ SICT-CP-50204
Thermal image apparatus	50205	(-40 ~ -20) °C (-20 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 900) °C (900 ~ 1 200) °C (1 200 ~ 1 400) °C (1 400 ~ 1 500) °C (1 500 ~ 1 600) °C (1 600 ~ 1 800) °C (1 800 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.9 °C 0.7 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 1.0 °C 1.1 °C 1.3 °C 1.6 °C 1.7 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 2.5 °C 2.6 °C 4.2 °C 4.4 °C 4.6 °C 4.7 °C	Transfer Standard Pyrometer/ SICT-CP-50205 SICT-CP-50205

502. non contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Blackbody furnaces	50206	(-40 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 500) °C (1 500 ~ 1 700) °C (1 700 ~ 1 800) °C (1 800 ~ 1 900) °C (1 900 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.6 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 0.9 °C 1.0 °C 1.1 °C 1.4 °C 1.5 °C 1.6 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 3.7 °C 3.9 °C 4.1 °C 4.3 °C	Transfer Standard Pyrometer/ SICT-CP-50206
Others; ear thermometers, etc.	50207	(30 ~ 45) °C	0.07 °C	Standard prt/ SICT-CP-50207

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dew-point hygrometers; chilled mirror, alumina thin film, etc.	50301	(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -20) °C D.P. (-20 ~ 90) °C D.P. (90 ~ 95) °C D.P.	0.60 °C D.P. 0.32 °C D.P. 0.20 °C D.P. 0.19 °C D.P. 0.13 °C D.P. 0.15 °C D.P.	Dewpoint Meter/ SICT-CP-50301
Relative humidity hygrometers; polimer thin film, hair, etc.	50302	humidity (3 ~ 60) % R.H. (60 ~ 90) % R.H. (90 ~ 98) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H.	Dewpoint Meter/ SICT-CP-50302
		Temperature (-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	0.6 °C 0.3 °C 0.5 °C 1.5 °C	
Psychrometers; Assmann ventilated, PRT type, etc.	50303	assmann ventilated (humidity) (10 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 95) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H.	Dewpoint Meter/ SICT-CP-50303
		(Temperature) (0 ~ 50) °C	0.3 °C	
		PRT type (humidity) (10 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H.	
		(Temperature) (0 ~ 80) °C (80 ~ 100) °C	0.3 °C 0.5 °C	
Temperature humidity recorders; hygrothermograph, etc.	50304	Humidity (5 ~ 70) % R.H. (70 ~ 95) % R.H.	2.1 % R.H. 2.2 % R.H.	Dewpoint Meter/ SICT-CP-50304
		Temperature (-20 ~ 80) °C	0.7 °C	

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Transducers; dew-point /relative humidity (Dew-point Transducers)	50305			Dewpoint Meter/ SICT-CP-50305
Dew point		(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -60) °C D.P. (-60 ~ -40) °C D.P. (-40 ~ -20) °C D.P. (-20 ~ 0) °C D.P. (0 ~ 50) °C D.P. (50 ~ 90) °C D.P. (90 ~ 95) °C D.P.	0.60 °C D.P. 0.33 °C D.P. 0.22 °C D.P. 0.21 °C D.P. 0.20 °C D.P. 0.15 °C D.P. 0.14 °C D.P. 0.15 °C D.P. 0.17 °C D.P.	
(Relative humidity Transducers)				
Humidity		(3 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H.	
Temperature		(-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	0.7 °C 0.3 °C 0.5 °C 1.5 °C	
Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	50306			Dewpoint Meter/ SICT-CP-50306
Dew point		(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -30) °C D.P. (-30 ~ -10) °C D.P. (-10 ~ 60) °C D.P. (60 ~ 80) °C D.P. (80 ~ 95) °C D.P.	0.60 °C D.P. 0.32 °C D.P. 0.19 °C D.P. 0.17 °C D.P. 0.16 °C D.P. 0.13 °C D.P. 0.14 °C D.P. 0.15 °C D.P.	
Humidity		(3 ~ 20) % R.H. (20 ~ 30) % R.H. (30 ~ 40) % R.H. (40 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 98) % R.H.	1.8 % R.H. 1.7 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H. 1.8 % R.H. 1.9 % R.H.	
Temperature		(-90 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C	0.4 °C 0.5 °C 0.6 °C	

504. Moisture

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cereal moisture meters Moisture	50401	(9 ~ 20) % M.C.	0.7 % M.C.	Balance/ SICT-CP-50401
Wood moisture meters Moisture	50402	(8 ~ 25) % M.C.	2.5 % M.C.	Balance/ SICT-CP-50402
Paper moisture meters Moisture	50403	(8 ~ 20) % M.C.	3.4 % M.C.	Balance/ SICT-CP-50403

601. Sound in air

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Sound calibrators	60102	(200 ~ 300) Hz	0.08 dB	Reference microphone/ SICT-CP-60102
Pistonphones		(900 ~ 1 100) Hz	0.08 dB	
Sound Pressure Level Calibrators		(28 ~ 35) Hz	0.12 dB	
Multifunction Acoustic Calibrators		(35 ~ 90) Hz	0.09 dB	
		(90 ~ 4 500) Hz	0.08 dB	
		(4 500 ~ 9 000) Hz	0.09 dB	
		(9 000~ 14 000) Hz	0.13 dB	
		(14 000~ 17 000) Hz	0.21 dB	
Microphones	60104	20 Hz	0.15 dB	Reference microphone/ SICT-CP-60104
		(20 ~ 25) Hz	0.13 dB	
		(25 ~ 31.5) Hz	0.12 dB	
		(31.5 ~ 50) Hz	0.10 dB	
		(50 ~ 63) Hz	0.09 dB	
		(63 ~ 8 000) Hz	0.08 dB	
		(8 000 ~ 10 000) Hz	0.09 dB	
		(10 000 ~ 12 500) Hz	0.10 dB	
		(12 500 ~ 16 000) Hz	0.12 dB	
		(16 000 ~ 20 000) Hz	0.16 dB	
Sound level meters	60106	20 Hz	0.5 dB	Reference microphone/ SICT-CP-60106
		(20 ~ 50) Hz	0.4 dB	
		(50 ~ 160) Hz	0.3 dB	
		(160 ~ 2 000) Hz	0.2 dB	
		(2 000 ~ 8 000) Hz	0.3 dB	
		(8 000 ~ 16 000) Hz	0.4 dB	
		(16 000 ~ 20 000) Hz	0.5 dB	

603. Vibration

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Vibration calibrators	60301	(10 ~ 5 000) Hz	1.6×10^{-2}	Reference Accelerometer/ SICT-CP-60301
Vibration transducers	60302	(0.2 ~ 0.3) Hz (0.3 ~ 20) Hz (20 ~ 2 500) Hz (2 500 ~ 5 000) Hz (5 000 ~ 10 000) Hz (10 000 ~ 15 000) Hz (15 000 ~ 20 000) Hz	1.3×10^{-2} 1.2×10^{-2} 1.0×10^{-2} 1.5×10^{-2} 1.7×10^{-2} 2.2×10^{-2} 3.2×10^{-2}	Reference Accelerometer/ SICT-CP-60302
Shock transducers sensitivity		Pulse duration : (5 ~ 30) ms (5 ~ 200) m/s ² Pulse duration : (0.1 ~ 5) ms (200 ~ 2 000) m/s ² (2 000 ~ 20 000) m/s ² (20 000 ~ 100 000) m/s ²	1.2×10^{-2} 1.0×10^{-2} 1.9×10^{-2} 3.3×10^{-2}	
Vibration measuring instruments (Vibration measuring instruments)	60303			Reference Accelerometer/ SICT-CP-60303
Acceleration		(10 ~ 20) Hz (20 ~ 1 250) Hz (1 250 ~ 5 000) Hz	1.5×10^{-2} 1.4×10^{-2} 1.5×10^{-2}	
Velocity		(10 ~ 20) Hz (20 ~ 1 000) Hz (1 000 ~ 2 500) Hz	1.5×10^{-2} 1.4×10^{-2} 1.5×10^{-2}	
Displacement		(10 ~ 160) Hz (160 ~ 315) Hz (315 ~ 630) Hz	1.3×10^{-2} 2.0×10^{-2} 5.9×10^{-2}	
(Shock measuring instruments, recorders, etc.)				
Shock acceleration		Pulse duration : (2.7 ~ 30) ms (5 ~ 10) m/s ² (10 ~ 1 000) m/s ² Pulse duration : (0.5 ~ 2.7) ms (1 000 ~ 4 000) m/s ²	2.3×10^{-2} 1.7×10^{-2} 2.5×10^{-2}	

701. Photometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Illuminance meters Illuminance	70101	(0.5 ~ 20 000) lx	1.7×10^{-2}	Illuminance Meters/ SICT-CP-70101
Luminance meters Luminance	70102	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.1×10^{-2} 1.7×10^{-2} 1.4×10^{-2} 1.6×10^{-2}	Luminance Standard Sources/ SICT-CP-70102
Total luminous flux meters Total luminous flux	70103	70 lm (70 ~ 4 650) lm	3.2×10^{-2} 1.5×10^{-2}	Total Luminous Flux Standard Lamps/ SICT-CP-70103
Luminous intensity meters Luminance	70104	(72 ~ 3 200) cd	3.7×10^{-2}	Luminous Intensity Standard Lamps, Illuminance Meters / SICT-CP-70104

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color temperature meters	70202			Color Temperature Standard Lamps/ SICT-CP-70202
Color temperature		(2 677 ~ 3 333) K	25 K	
Chromaticity		x y	0.004 0.004	
Color temperature standard lamps	70203			Spectroradiometers/ SICT-CP-70203
Color temperature		(2 677 ~ 3 333) K	27 K	
Chromaticity		x y	0.005 0.005	
Colorimeters; source color	70204			Luminance Standard Sources/ SICT-CP-70204
Luminance		1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.1 × 10 ⁻² 1.7 × 10 ⁻² 1.4 × 10 ⁻² 1.6 × 10 ⁻²	
Chromaticity		(WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.003 0.003	
Laser power meters	70207			Optical Power Meters/ SICT-CP-70207
		(405 nm) (0.75 ~ 9) mW	1.2 × 10 ⁻²	
		(660 nm) (0.7 ~ 47) mW	1.2 × 10 ⁻²	
		(785 nm) (0.7 ~ 46) mW	1.2 × 10 ⁻²	
		(1 080 nm) (1 ~ 40) W	3.3 × 10 ⁻²	
Standard LED light sources	70208			Toral Spectral Radiant Flux Meters/ SICT-CP-70208
Total luminous flux		(68.4 ~ 72.6) lm	3.8 × 10 ⁻²	
Total luminous flux standard lamps	70209			Total Luminous Flux Standard I amnes/ SICT-CP-70209
Total luminous flux		(320 ~ 10 000) lm	4.7 × 10 ⁻²	

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical detectors Relative spectral responsivity	70210	(0 ~ 1) 300 nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 320) nm (320 ~ 370) nm (370 ~ 440) nm (440 ~ 525) nm (525 ~ 630) nm (630 ~ 955) nm (955 ~ 990) nm (990 ~ 1 100) nm	8.1×10^{-2} 7.0×10^{-2} 6.1×10^{-2} 5.2×10^{-2} 3.9×10^{-2} 3.0×10^{-2} 1.4×10^{-2} 1.0×10^{-2} 1.2×10^{-2} 2.9×10^{-2} 4.0×10^{-2}	Photodiodes/ SICT-CP-70210
Pyranometers and pyrheliometers Irradiance responsivity	70211	(250 ~ 2 500) nm (1 000 ± 150) W/m ²	2.9×10^{-2}	Standard pyranometers/ SICT-CP-70211
Display color analyzers; luminance, chromaticity, white balance, etc. Luminance	70213	1 cd/m ² (1 ~ 5) cd/m ² (5 ~ 200) cd/m ²	3.8×10^{-2} 1.8×10^{-2} 1.7×10^{-2}	Luminance Meters/ SICT-CP-70213
Chromaticity		(WHITE) x y (RED) x y (GREEN) x y (BLUE) x y	0.004 4 0.006 1 0.003 6 0.003 3 0.003 5 0.004 2 0.003 5 0.003 2	
Luminous intensity standard lamps Luminous intensity	70214	(10 ~ 20 000) cd	4.0×10^{-2}	Spectroradiometers/ SICT-CP-70214
Spectral irradiance standard lamps Illuminance	70215	(792 ~ 7 105) lx (Deuterium arc lamp) 200 nm (200 ~ 205) nm (205 ~ 400) nm (Tungsten halogen lamp) 250 nm (250 ~ 270) nm (270 ~ 295) nm (295 ~ 375) nm (375 ~ 2 295) nm (2 295 ~ 2 345) nm (2 345 ~ 2 400) nm	2.8×10^{-2} 5.4×10^{-2} 5.3×10^{-2} 4.9×10^{-2} 5.5×10^{-2} 5.2×10^{-2} 5.0×10^{-2} 4.4×10^{-2} 3.8×10^{-2} 4.0×10^{-2} 4.6×10^{-2}	Spectral Irradiance Standard Lamps/ SICT-CP-70215

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Total spectral radiant flux standard lamps	70216			Total Spectral Radiant Flux Standard Lamps/ SICT-CP-70216
Total spectral radiant flux		350 nm (350 ~ 365) nm (365 ~ 380) nm (380 ~ 400) nm (400 ~ 455) nm (455 ~ 850) nm	6.7 × 10 ⁻² 6.3 × 10 ⁻² 5.8 × 10 ⁻² 4.2 × 10 ⁻² 3.9 × 10 ⁻² 3.6 × 10 ⁻²	
Luminance standard sources	70217			Luminance Standard Sources/ SICT-CP-70217
Luminance		1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.2 × 10 ⁻² 1.8 × 10 ⁻² 1.5 × 10 ⁻² 1.8 × 10 ⁻²	
Chromaticity		(WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.004 0.004	
Spectral radiance standard sources	70218			Spectral Radiance Standard Sources/ SICT-CP-70218
Spectral radiance		300 nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 315) nm (315 ~ 320) nm (320 ~ 330) nm (330 ~ 340) nm (340 ~ 425) nm (425 ~ 470) nm (470 ~ 1 050) nm (1 050 ~ 1 600) nm	2.0 × 10 ⁻¹ 1.7 × 10 ⁻¹ 1.2 × 10 ⁻¹ 8.9 × 10 ⁻² 7.4 × 10 ⁻² 4.8 × 10 ⁻² 4.1 × 10 ⁻² 3.5 × 10 ⁻² 3.0 × 10 ⁻² 2.8 × 10 ⁻² 3.0 × 10 ⁻²	
UV irradiance meters	70219			UV Meter Standard Detectors/ SICT-CP-70219
Irradiance (UV Meter)		(254 nm) 50 μW/cm ² ~ 3 mW/cm ² (365 nm) 10 μW/cm ² ~ 230 mW/cm ² (405 nm) 10 μW/cm ² ~ 230 mW/cm ²	3.7 × 10 ⁻² 3.8 × 10 ⁻² 3.8 × 10 ⁻²	

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectral irradiance meters	70220	Wavelength	(250 ~ 2 030) nm	0.3 nm Spectral Irradiance Standard Lamps/ SICT-CP-70220
		Illuminance	(813 ~ 6 879) lx	2.1×10^{-2}
		Spectral Irradiance	200 nm	4.9×10^{-2}
			(200 ~ 245) nm	4.8×10^{-2}
			(245 ~ 345) nm	4.4×10^{-2}
			(345 ~ 375) nm	3.8×10^{-2}
			(375 ~ 895) nm	3.4×10^{-2}
			(895 ~ 1 050) nm	3.0×10^{-2}
			(1 050 ~ 2 295) nm	3.4×10^{-2}
			(2 295 ~ 2 345) nm	3.7×10^{-2}
			(2 345 ~ 2 400) nm	4.3×10^{-2}
Total spectral radiant flux meters	70221	Wavelength	(350 ~ 850) nm	0.25 nm Total Spectral Radiant Flux Standard Lamps/ SICT-CP-70221
		Total spectral radiant flux	350 nm	2.0×10^{-2}
			(350 ~ 365) nm	1.8×10^{-2}
			(365 ~ 375) nm	1.7×10^{-2}
			(375 ~ 390) nm	1.6×10^{-2}
			(390 ~ 445) nm	1.5×10^{-2}
			(445 ~ 850) nm	1.4×10^{-2}
Spectral radianc meters	70222	Wavelength	(350 ~ 1 694) nm	0.25 nm Spectral Radiance Standard Sources/ SICT-CP-70222
		Spectral radianc	300 nm	2.0×10^{-1}
			(300 ~ 305) nm	1.7×10^{-1}
			(305 ~ 310) nm	1.2×10^{-1}
			(310 ~ 315) nm	8.8×10^{-2}
			(315 ~ 320) nm	7.2×10^{-2}
			(320 ~ 325) nm	5.5×10^{-2}
			(325 ~ 335) nm	4.6×10^{-2}
			(335 ~ 345) nm	3.7×10^{-2}
			(345 ~ 405) nm	3.5×10^{-2}
			(405 ~ 455) nm	3.0×10^{-2}
			(455 ~ 755) nm	2.6×10^{-2}
			(755 ~ 1 400) nm	2.7×10^{-2}
			(1 400 ~ 1 525) nm	3.0×10^{-2}
			(1 525 ~ 1 600) nm	2.8×10^{-2}

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color	70301			Color Standard Tiles/ SICT-CP-70301
Included Reflectance Std. Light Source T_{view} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.0×10^{-2}	
2. I,Gray		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.0×10^{-2}	
3. m,Gray		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.0×10^{-2}	
4. d,Gray		X	1.1×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.0×10^{-2}	
5. Red		X	1.1×10^{-2}	
		Y	1.2×10^{-2}	
		Z	1.8×10^{-2}	
6. Yellow		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.6×10^{-2}	
7. Green		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.1×10^{-2}	
8. Cyan		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.0×10^{-2}	
Included Reflectance Std. Light Source T_{view} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		L*	0.36	
		a*	0.09	
		b*	0.07	
2. I,Gray		L*	0.32	
		a*	0.08	
		b*	0.07	
3. m,Gray		L*	0.24	
		a*	0.06	
		b*	0.06	
4. d,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.27	
		b*	0.26	
6. Yellow		L*	0.34	
		a*	0.17	
		b*	0.43	
7. Green		L*	0.24	
		a*	0.13	
		b*	0.12	
8. Cyan		L*	0.24	
		a*	0.14	
		b*	0.15	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		x	1.3×10^{-3}	
		y	1.3×10^{-3}	
2. I, Gray		x	1.3×10^{-3}	
		y	1.3×10^{-3}	
3. m, Gray		x	1.4×10^{-3}	
		y	1.3×10^{-3}	
4. d, Gray		x	1.4×10^{-3}	
		y	1.3×10^{-3}	
5. Red		x	2.7×10^{-3}	
		y	1.3×10^{-3}	
6. Yellow		x	1.2×10^{-3}	
		y	1.3×10^{-3}	
7. Green		x	1.4×10^{-3}	
		y	1.0×10^{-3}	
8. Cyan		x	2.0×10^{-3}	
		y	1.7×10^{-3}	
Exclude Reflectance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		x	1.0×10^{-2}	
		y	1.0×10^{-2}	
		z	1.0×10^{-2}	
2. I, Gray		x	1.0×10^{-2}	
		y	1.0×10^{-2}	
		z	1.0×10^{-2}	
3. m, Gray		x	1.0×10^{-2}	
		y	1.0×10^{-2}	
		z	1.0×10^{-2}	
4. d, Gray		x	1.1×10^{-2}	
		y	1.0×10^{-2}	
		z	1.4×10^{-2}	
5. Red		x	1.2×10^{-2}	
		y	1.3×10^{-2}	
		z	3.1×10^{-2}	
6. Yellow		x	1.0×10^{-2}	
		y	1.0×10^{-2}	
		z	2.3×10^{-2}	
7. Green		x	1.0×10^{-2}	
		y	1.0×10^{-2}	
		z	1.2×10^{-2}	
8. Cyan		x	1.0×10^{-2}	
		y	1.0×10^{-2}	
		z	1.0×10^{-2}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Exclude Reflectance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.36	
		a*	0.09	
		b*	0.07	
2. I,Gray		L*	0.31	
		a*	0.08	
		b*	0.07	
3. m,Gray		L*	0.23	
		a*	0.06	
		b*	0.05	
4. d,Gray		L*	0.15	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.32	
		b*	0.66	
6. Yellow		L*	0.33	
		a*	0.18	
		b*	0.58	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.14	
8. Cyan		L*	0.23	
		a*	0.16	
		b*	0.16	
Exclude Reflectance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White	x		1.3×10^{-3}	
	y		1.3×10^{-3}	
2. I,Gray	x		1.3×10^{-3}	
	y		1.3×10^{-3}	
3. m,Gray	x		1.4×10^{-3}	
	y		1.3×10^{-3}	
4. d,Gray	x		1.3×10^{-3}	
	y		1.3×10^{-3}	
5. Red	x		5.3×10^{-3}	
	y		1.3×10^{-3}	
6. Yellow	x		1.3×10^{-3}	
	y		1.5×10^{-3}	
7. Green	x		1.5×10^{-3}	
	y		1.0×10^{-3}	
8. Cyan	x		2.1×10^{-3}	
	y		1.8×10^{-3}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color	70301			Color Standard Tiles/ SICT-CP-70301
Transmittance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. E-LA200		X	4.0×10^{-3}	
		Y	4.0×10^{-3}	
		Z	9.0×10^{-3}	
2. G533		X	7.0×10^{-3}	
		Y	5.0×10^{-3}	
		Z	1.3×10^{-2}	
3. B460		X	5.0×10^{-3}	
		Y	4.0×10^{-3}	
		Z	3.0×10^{-3}	
4. ND 25		X	3.0×10^{-3}	
		Y	3.0×10^{-3}	
		Z	5.0×10^{-3}	
5. ND 40		X	3.0×10^{-3}	
		Y	3.0×10^{-3}	
		Z	3.0×10^{-3}	
6. ND 50		X	3.0×10^{-3}	
		Y	3.0×10^{-3}	
		Z	3.0×10^{-3}	
7. ND 70		X	3.0×10^{-3}	
		Y	2.0×10^{-3}	
		Z	3.0×10^{-3}	
Transmittance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. E-LA200		L*	0.11	
		a*	0.09	
		b*	0.22	
2. G533		L*	0.14	
		a*	0.12	
		b*	0.19	
3. B460		L*	0.12	
		a*	0.08	
		b*	0.16	
4. ND 25		L*	0.07	
		a*	0.04	
		b*	0.06	
5. ND 40		L*	0.07	
		a*	0.04	
		b*	0.05	
6. ND 50		L*	0.08	
		a*	0.05	
		b*	0.05	
7. ND 70		L*	0.09	
		a*	0.05	
		b*	0.05	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Transmittance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. E-LA200		x	1.0 × 10 ⁻³	
		y	1.0 × 10 ⁻³	
2. G533		x	1.4 × 10 ⁻³	
		y	1.6 × 10 ⁻³	
3. B460		x	2.0 × 10 ⁻³	
		y	1.6 × 10 ⁻³	
4. ND 25		x	1.3 × 10 ⁻³	
		y	1.2 × 10 ⁻³	
5. ND 40		x	1.3 × 10 ⁻³	
		y	1.3 × 10 ⁻³	
6. ND 50		x	1.3 × 10 ⁻³	
		y	1.3 × 10 ⁻³	
7. ND 70		x	1.3 × 10 ⁻³	
		y	1.3 × 10 ⁻³	
Color standard tiles Included Reflectance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		X	0.95	
		Y	0.86	
		Z	0.99	
2. I,Gray		X	0.66	
		Y	0.60	
		Z	0.70	
3. m,Gray		X	0.29	
		Y	0.26	
		Z	0.31	
4. d,Gray		X	0.11	
		Y	0.10	
		Z	0.11	
5. Red		X	0.36	
		Y	0.22	
		Z	0.14	
6. Yellow		X	0.78	
		Y	0.68	
		Z	0.21	
7. Green		X	0.19	
		Y	0.23	
		Z	0.20	
8. Cyan		X	0.20	
		Y	0.24	
		Z	0.49	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source A(2°, 10°), C(2°, 10°), D65(2°, 10°) 1. White 2. I,Gray 3. m,Gray 4. d,Gray 5. Red 6. Yellow 7. Green 8. Cyan	70304	L* a* b* L* a* b* L* a* b* L* a* b* L* a* b* L* a* b* L* a* b*	0.37 0.09 0.07 0.32 0.08 0.07 0.24 0.06 0.06 0.17 0.04 0.04 0.25 0.27 0.26 0.34 0.17 0.43 0.24 0.13 0.12 0.24 0.14 0.15	Color Standard Tiles/ SICT-CP-70304
Included Reflectance Std. Light Source A(2°, 10°), C(2°, 10°), D65(2°, 10°) 1. White 2. I,Gray 3. m,Gray 4. d,Gray 5. Red 6. Yellow 7. Green 8. Cyan		x y x y x y x y x y x y x y	0.000 7 0.000 6 0.000 7 0.000 6 0.000 7 0.000 6 0.000 7 0.000 6 0.001 4 0.000 6 0.000 7 0.000 8 0.000 6 0.000 7 0.000 6 0.000 6	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Exclude Reflectance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		X Y Z	0.91 0.82 0.95	
2. I,Gray		X Y Z	0.62 0.56 0.66	
3. m,Gray		X Y Z	0.25 0.23 0.26	
4. d,Gray		X Y Z	0.07 0.06 0.07	
5. Red		X Y Z	0.32 0.18 0.12	
6. Yellow		X Y Z	0.74 0.64 0.19	
7. Green		X Y Z	0.15 0.20 0.16	
8. Cyan		X Y Z	0.17 0.21 0.45	
Exclude Reflectance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		L* a* b*	0.36 0.09 0.07	
2. I,Gray		L* a* b*	0.31 0.08 0.07	
3. m,Gray		L* a* b*	0.23 0.06 0.05	
4. d,Gray		L* a* b*	0.15 0.04 0.04	
5. Red		L* a* b*	0.25 0.32 0.66	
6. Yellow		L* a* b*	0.33 0.18 0.58	
7. Green		L* a* b*	0.23 0.14 0.14	
8. Cyan		L* a* b*	0.23 0.16 0.16	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Exclude Reflectance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		x y	0.000 6 0.000 6	
2. I,Gray		x y	0.000 6 0.000 6	
3. m,Gray		x y	0.000 6 0.000 6	
4. d,Gray		x y	0.000 6 0.000 6	
5. Red		x y	0.002 9 0.000 6	
6. Yellow		x y	0.000 8 0.000 9	
7. Green		x y	0.000 6 0.000 7	
8. Cyan		x y	0.000 5 0.000 5	
Absolute Spectral Reflectance White Plate (Include, Exclude Reflectance)		360 nm (360 ~ 830) nm	0.009 4 0.012	
Gloss meters	70306			Gloss Standard/ SICT-CP-70306
Gloss		20 ° 60 ° 85 °	8.9 × 10 ⁻³ 9.5 × 10 ⁻³ 8.0 × 10 ⁻³	
Gloss standard plates	70307			Gloss Meter/ SICT-CP-70307
Gloss		20° 60° 85°	9.5 × 10 ⁻³ 9.8 × 10 ⁻³ 8.3 × 10 ⁻³	
Haze meters	70308			Haze Standard Plate, Transmittance Standard Plates/ SICT-CP-70308
Haze		H-1 H-5 H-10 H-20 H-30	0.30 0.26 0.4 0.6 0.8	
Transmittance		T-30 T-50 T-70 T-90	0.50 0.50 0.50 0.50	
Lens meters	70312			Reference Lens/ SICT-CP-70312
Vertex diopter		-25 D ~ 25 D	0.03 D	
Optical densitometers	70315			Density CRM/ SICT-CP-70315
Density		1 Step ~ 10 Step 11 Step 11 Step ~ 15 Step	0.03 0.06 0.11	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Reflectance meters	70319			Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70319
Reflectance		380 nm ~ 780 nm	1.1×10^{-2}	
Refractometers	70321			Reference Refracto CRM/ SICT-CP-70321
Refracto		(1.332 99 ~ 1.505 80) nD 1.51 nD 1.62 nD	0.000 04 nD 0.000 2 nD 0.000 2 nD	
Transmittance meters	70323			Transmittance Filter/ SICT-CP-70323
		(0.1) (250 ~ 750) nm	6.1×10^{-3}	
		(0.5) (250 ~ 750) nm	3.8×10^{-3}	
		(0.9) (250 ~ 750) nm	2.2×10^{-3}	
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Wavelength		(240.7 ~ 976.1) nm (990.2 ~ 2 536.5) nm	0.4 nm 0.5 nm	
Transmittance		(Filter NO 10) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm 900 nm 1 000 nm 1 200 nm 1 400 nm 1 600 nm 1 800 nm 2 000 nm 2 200 nm 2 400 nm 2 500 nm	8.7×10^{-3} 9.2×10^{-3} 8.2×10^{-3} 6.6×10^{-3} 6.8×10^{-3} 6.5×10^{-3} 6.6×10^{-3} 6.8×10^{-3} 6.4×10^{-3} 6.7×10^{-3} 6.6×10^{-3} 8.0×10^{-3} 8.1×10^{-3} 8.4×10^{-3} 7.7×10^{-3} 8.0×10^{-3} 8.2×10^{-3} 8.1×10^{-3} 8.6×10^{-3} 9.2×10^{-3} 1.7×10^{-2}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
		(Filter NO 30, 40, 50)		
		250 nm	8.3×10^{-3}	
		300 nm	8.2×10^{-3}	
		350 nm	8.0×10^{-3}	
		400 nm	5.5×10^{-3}	
		450 nm	5.6×10^{-3}	
		500 nm	5.7×10^{-3}	
		550 nm	5.6×10^{-3}	
		600 nm	5.6×10^{-3}	
		650 nm	5.5×10^{-3}	
		700 nm	5.4×10^{-3}	
		750 nm	5.5×10^{-3}	
		(Filter NO 30)		
		900 nm	7.7×10^{-3}	
		1 000 nm	7.4×10^{-3}	
		1 200 nm	7.3×10^{-3}	
		1 400 nm	7.3×10^{-3}	
		1 600 nm	7.4×10^{-3}	
		1 800 nm	7.3×10^{-3}	
		2 000 nm	7.3×10^{-3}	
		2 200 nm	8.0×10^{-3}	
		2 400 nm	7.3×10^{-3}	
		2 500 nm	8.9×10^{-3}	
		(Filter NO 90)		
		250 nm	8.0×10^{-3}	
		300 nm	8.0×10^{-3}	
		350 nm	7.8×10^{-3}	
		400 nm	5.4×10^{-3}	
		450 nm	5.3×10^{-3}	
		500 nm	5.4×10^{-3}	
		550 nm	5.4×10^{-3}	
		600 nm	5.4×10^{-3}	
		650 nm	5.4×10^{-3}	
		700 nm	5.4×10^{-3}	
		750 nm	5.4×10^{-3}	
		900 nm	7.3×10^{-3}	
		1 000 nm	7.3×10^{-3}	
		1 200 nm	7.3×10^{-3}	
		1 400 nm	7.3×10^{-3}	
		1 600 nm	7.3×10^{-3}	
		1 800 nm	7.3×10^{-3}	
		2 000 nm	7.3×10^{-3}	
		2 200 nm	7.3×10^{-3}	
		2 400 nm	7.3×10^{-3}	
		2 500 nm	7.8×10^{-3}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Transmittance		(Filter NO 1) 440 nm 465 nm 546 nm 590 nm 635 nm	1.4×10^{-2} 9.1×10^{-3} 9.4×10^{-3} 1.1×10^{-2} 1.0×10^{-2}	
		(Filter NO 3) 440 nm 465 nm 546 nm 590 nm 635 nm	7.9×10^{-3} 5.8×10^{-3} 6.1×10^{-3} 6.3×10^{-3} 6.2×10^{-3}	
Absorbance		(Filter NO 10) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm 900 nm 1 000 nm 1 200 nm 1 400 nm 1 600 nm 1 800 nm 2 000 nm 2 200 nm 2 400 nm 2 500 nm	0.003 7 0.003 8 0.003 5 0.002 7 0.002 6 0.002 5 0.002 5 0.002 5 0.002 8 0.002 6 0.002 4 0.003 3 0.003 3 0.003 3 0.003 2 0.003 3 0.003 3 0.003 4 0.003 7 0.007 2	
		(Filter NO 30, 40, 50) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	0.003 8 0.003 7 0.003 6 0.002 5 0.002 5 0.002 6 0.002 6 0.002 6 0.002 5 0.002 5 0.002 5	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Absorbance		(Filter NO 30)		
		900 nm	0.003 5	
		1 000 nm	0.003 4	
		1 200 nm	0.003 2	
		1 400 nm	0.003 3	
		1 600 nm	0.003 4	
		1 800 nm	0.003 2	
		2 000 nm	0.003 3	
		2 200 nm	0.003 5	
		2 400 nm	0.003 3	
		2 500 nm	0.003 9	
		(Filter NO 90)		
		250 nm	0.003 6	
		300 nm	0.003 6	
		350 nm	0.003 5	
		400 nm	0.002 5	
		450 nm	0.002 5	
		500 nm	0.002 5	
		550 nm	0.002 6	
		600 nm	0.002 6	
		650 nm	0.002 6	
		700 nm	0.002 5	
		750 nm	0.002 6	
		900 nm	0.003 3	
		1 000 nm	0.003 3	
		1 200 nm	0.003 3	
		1 400 nm	0.003 3	
		1 600 nm	0.003 2	
		1 800 nm	0.003 4	
		2 000 nm	0.003 3	
		2 200 nm	0.003 3	
		2 400 nm	0.003 3	
		2 500 nm	0.003 6	
		(Filter NO 1)		
		440 nm	0.004 1	
		465 nm	0.002 8	
		546 nm	0.002 8	
		590 nm	0.003 0	
		635 nm	0.003 5	
		(Filter NO 3)		
		440 nm	0.003 2	
		465 nm	0.002 4	
		546 nm	0.002 6	
		590 nm	0.002 5	
		635 nm	0.002 7	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Spectral Reflectance		250 nm (250 ~ 2 500) nm	1.5×10^{-2} 1.5×10^{-2}	
Wavenumber		544.9 cm^{-1} 842.1 cm^{-1} 906.82 cm^{-1} 1 028.42 cm^{-1} 1 069.27 cm^{-1} 1 154.62 cm^{-1} 1 583.04 cm^{-1} 1 601.38 cm^{-1} 2 850.20 cm^{-1} 3 001.40 cm^{-1} 3 026.44 cm^{-1} 3 060.14 cm^{-1} 3 082.22 cm^{-1}	2.5 cm^{-1} 1.3 cm^{-1} 0.12 cm^{-1} 0.28 cm^{-1} 0.78 cm^{-1} 0.12 cm^{-1} 0.12 cm^{-1} 0.13 cm^{-1} 0.14 cm^{-1} 0.12 cm^{-1} 0.12 cm^{-1} 0.12 cm^{-1}	
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Wavelength		(240 ~750) nm	0.5 nm	
Transmittance		(0.1 ~ 0.3) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm (0.3 ~ 0.5) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	8.5×10^{-3} 8.1×10^{-3} 8.1×10^{-3} 5.9×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 8.3×10^{-3} 8.1×10^{-3} 8.0×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Transmittance		(0.5 ~ 0.9) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	8.2×10^{-3} 8.1×10^{-3} 8.0×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.6×10^{-3} 5.6×10^{-3} 5.7×10^{-3} 5.9×10^{-3}	
Absorbance		(0.1 ~ 0.3) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm (0.3 ~ 0.5) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm (0.5 ~ 0.9) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	0.003 7 0.003 7 0.003 8 0.002 7 0.002 7 0.002 8 0.002 8 0.002 7 0.002 7 0.002 7 0.002 8 0.003 6 0.003 6 0.003 6 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 5 0.003 6 0.003 5 0.003 5 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4	
Reflectance		(360 ~ 830) nm	1.0×10^{-2}	

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Broadband light sources Wavelength output	70402	1 310 nm, 1 550 nm	0.058 nm	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70402
Optical power output		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.070 dB	
Optical attenuators Optical Attenuation	70410	1 310 nm, 1 550 nm (-60 ~ 0) dB	0.08 dB	Optical powermeter, Optical power stabilized lasers and LDs/ SICT-CP-70410
Absolute optical power		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.072 dB	
Fiber-optic power meters Optical Linearity	70412	1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70412
Optical Attenuation		1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	
Optical loss testers Optical Attenuation	70413	1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	Optical attenuator/ SICT-CP-70413
Absolute optical power measure		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.072 dB	
Optical multimeters Linearity measure	70415	1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70415
Absolute optical power measure		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	
Optical network analyzer (Optical multimeter) Absolute optical power (광Optical spectrum analyzer) Wavelength measure Resolution measure Absolute optical power measure (Optical attenuator) Optical Attenuation Return loss (Optical time domain reflectometer) Wavelength output Optical Length measure	70416	1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416
Wavelength measure		1 310 nm 1 550 nm	0.058 nm 0.058 nm	
Resolution measure		Resolution: (0.1 ~ 1) nm 1 310 nm 1 550 nm	0.058 nm 0.058 nm	
Absolute optical power measure		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	
Optical Attenuation		1 310 nm, 1 550 nm (-60 ~ 0) dB	0.07 dB	
Return loss		1 310 nm, 1 550 nm (20 ~ 40) dB	0.8 dB	
Wavelength output		1 310 nm, 1 550 nm	0.082 nm	
Optical Length measure		1 310 nm 3.3 km Fiber 13.4 km Fiber 1 550 nm 3.3 km Fiber 13.4 km Fiber	0.081 m 0.34 m 0.080 m 0.34 m	

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical network analyzer Optical loss measure	70416	1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	0.13 dB 0.05 dB 0.05 dB 0.05 dB	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416
Optical spectrum analyzers Wavelength measure	70417	1 310 nm 1 550 nm Resolution : (0.1 ~ 1) nm 1 310 nm 1 550 nm	0.058 nm 0.058 nm 0.058 nm 0.058 nm	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator, Optical spectrum analyzer/ SICT-CP-70417
Resolution measure				
Absolute optical power measure		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	
Linearity measure		1 310 nm, 1 550 nm (-60 ~ 0) dB	0.03 dB	
Optical time domain reflectometers: OTDR Wavelength output	70418	1 310 nm, 1 550 nm	0.08 nm	Optical length fiber reference, Optical fiberloss reference, Optical spectrum analyzer/ SICT-CP-70418
Optical Length measure		1 310 nm 3.3 km Fiber 13.4 km Fiber 1 550 nm 3.3 km Fiber 13.4 km Fiber	0.081 m 0.34 m 0.080 m 0.34 m	
Optical loss measure		1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	0.13 dB 0.05 dB 0.05 dB 0.05 dB	
Return loss meters Return loss measure	70423	1 310 nm, 1 550 nm 20 dB ~ 40 dB	0.8 dB	Optical Returnloss generator SICT-CP-70423
Frequency stabilized lasers and LDs Wavelength	70429	1 310 nm 1 550 nm	4 pm 4 pm	Wavelength meter, Optical powermeter/ SICT-CP-70429
optical power		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.07 dB	

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
ASE light sources	70430	Wavelength output	1 310 nm, 1 550 nm 0.058 nm	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70430
		Optical power output	1 310 nm, 1 550 nm (-60 ~ 0) dBm 0.07 dB	
Optical power stabilized lasers and LDs	70433	Wavelength output	1 310 nm 1 550 nm 4 pm 4 pm	Wavelength meter, Optical powermeter/ SICT-CP-70433
		Optical power output	1 310 nm, 1 550 nm (-60 ~ 0) dBm 0.07 dB	

901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Breath alcohol analyzers	90101			Standard gas/ SICT-CP-90101
Dry process		(0.000 ~ 0.080) %BAC (0.080 ~ 0.190) %BAC	3.3×10^{-2} 2.1×10^{-2}	
Wet process		(0.000 ~ 0.080) %BAC (0.080 ~ 0.150) %BAC (0.150 ~ 0.400) %BAC	2.9×10^{-2} 1.6×10^{-2} 1.3×10^{-2}	
Environmental air quality monitoring instruments	90102			Standard gas/ SICT-CP-90102
Carbon monoxide		(0 ~ 700) $\mu\text{mol/mol}$ (0.07 ~ 1.5) cmol/mol (1.5 ~ 20) cmol/mol (0 ~ 850) $\mu\text{mol/mol}$	1.0×10^{-2} 2.0×10^{-2} 1.2×10^{-2} 2.2×10^{-2}	
Carbon dioxide		(0 ~ 0.50) cmol/mol (0.50 ~ 5.00) cmol/mol (5.00 ~ 19.00) cmol/mol	2.0×10^{-2} 1.5×10^{-2} 2.1×10^{-2}	
Nitrogen monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.1×10^{-2}	
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
Hydrogen sulfide		(0 ~ 45) $\mu\text{mol/mol}$	3.6×10^{-2}	
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
Isobutylene		(0 ~ 25) $\mu\text{mol/mol}$	1.0×10^{-2}	
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$ (0.05 ~ 2.0) cmol/mol	2.3×10^{-2} 2.1×10^{-2}	
Hydrogen chloride		(0 ~ 50) $\mu\text{mol/mol}$	4.8×10^{-2}	
Sulfur hexafluoride		(0 ~ 100) cmol/mol	0.1×10^{-2}	
Ozone		0.0 nmol/mol (0.0 ~ 1 000.0) nmol/mol	2.2 nmol/mol 2.5×10^{-2}	
Gas analyzers	90103			Standard gas/ SICT-CP-90103
Oxygen		(0 ~ 700) $\mu\text{mol/mol}$ (0.07 ~ 1.5) cmol/mol (1.5 ~ 20) cmol/mol	1.0×10^{-2} 2.0×10^{-2} 1.2×10^{-2}	
Carbon monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Carbon dioxide		(0 ~ 0.50) cmol/mol (0.50 ~ 5.00) cmol/mol (5.00 ~ 19.00) cmol/mol	2.0×10^{-2} 1.5×10^{-2} 2.1×10^{-2}	
Nitrogen monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.1×10^{-2}	
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
Hydrogen sulfide		(0 ~ 45) $\mu\text{mol/mol}$	3.6×10^{-2}	
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
Isobutylene		(0 ~ 25) $\mu\text{mol/mol}$	1.0×10^{-2}	
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$ (0.05 ~ 2.0) cmol/mol	2.3×10^{-2} 2.1×10^{-2}	
Hydrogen chloride		(0 ~ 50) $\mu\text{mol/mol}$	4.8×10^{-2}	
Sulfur hexafluoride		(0 ~ 100) cmol/mol	0.1×10^{-2}	
Ozone		0.0 nmol/mol (0.0 ~ 1 000.0) nmol/mol	2.2 nmol/mol 2.5×10^{-2}	

901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Exhaust gas test instruments	90104	(0 ~ 1.5) $\mu\text{mol/mol}$	2.0×10^{-2}	Standard gas/ SICT-CP-90103
		(1.5 ~ 20) cmol/mol	1.1×10^{-2}	
		(0 ~ 5.0) cmol/mol	2.1×10^{-2}	
		(0 ~ 19) cmol/mol	2.0×10^{-2}	
		(0 ~ 2 000) $\mu\text{mol/mol}$	2.0×10^{-2}	
		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Others; pH meter, Electrical conductivity meter	90104	(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	CRM/ SICT-CP-90199
		(0 ~ 500) $\mu\text{mol/mol}$	2.3×10^{-2}	
		(0.05 ~ 2.0) cmol/mol	2.1×10^{-2}	
pH meter		(4 ~ 10) pH	0.013 pH	
Electrical conductivity meter		100 $\mu\text{S/cm}$	3.1 $\mu\text{S/cm}$	
		1 413 $\mu\text{S/cm}$	9.7 $\mu\text{S/cm}$	
		12.85 mS/cm	0.073 mS/cm	
		111.3 mS/cm	0.78 mS/cm	